

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
 Data File : VX033668.D
 Acq On : 09 Jan 2023 17:12
 Operator : JC/MD
 Sample : 01079-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2023
 Supervised By :Mahesh Dadoda 01/11/2023

Quant Time: Jan 10 00:48:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jan 07 06:16:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	12475	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	86979	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	93272	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	34158	33.816	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	112.733%#
60) 4-Bromofluorobenzene	11.079	95	48335	28.753	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.833%
63) Toluene-d8	8.647	98	90488	30.314	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.033%
Target Compounds						
15) Acetone	2.398	58	249m	1.672	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
Data File : VX033668.D
Acq On : 09 Jan 2023 17:12
Operator : JC/MD
Sample : 01079-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Jan 10 00:48:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jan 07 06:16:21 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2023
Supervised By :Mahesh Dadoda 01/11/2023

