


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
Data File : VX033654.D
Acq On : 09 Jan 2023 10:06
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Sat Jan 07 06:16:21 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1634

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	1177	PASS
75	95	30	60	51.6	3186	PASS
95	95	100	100	100.0	6169	PASS
96	95	5	9	6.8	421	PASS
173	174	0.00	2	1.8	97	PASS
174	95	50	100	86.5	5337	PASS
175	174	5	9	7.9	422	PASS
176	174	95	101	96.7	5163	PASS
177	176	5	9	7.3	377	PASS