

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010924\
 Data File : VX039691.D
 Acq On : 09 Jan 2024 19:54
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 10 07:26:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010924W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 10 07:08:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100851	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	178628	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79922	47.433	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.860%
35) Dibromofluoromethane	5.385	113	57722	49.451	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.647	98	222426	50.302	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	85094	51.390	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.780%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.167	85	73068	53.913	ug/l	97
3) Chloromethane	1.295	50	77662	51.352	ug/l	97
4) Vinyl Chloride	1.374	62	76694	53.484	ug/l	100
5) Bromomethane	1.599	94	35788	52.093	ug/l	100
6) Chloroethane	1.666	64	38435	55.910	ug/l	97
7) Trichlorofluoromethane	1.874	101	108760	51.594	ug/l	100
8) Diethyl Ether	2.136	74	39736	51.955	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	66525	53.124	ug/l	97
10) Methyl Iodide	2.447	142	77198	54.713	ug/l	94
11) Tert butyl alcohol	3.002	59	114311	257.736	ug/l #	92
12) 1,1-Dichloroethene	2.313	96	65080	52.445	ug/l	94
13) Acrolein	2.240	56	75239	249.909	ug/l	99
14) Allyl chloride	2.660	41	128807	53.163	ug/l	98
15) Acrylonitrile	3.063	53	239833	263.805	ug/l	99
16) Acetone	2.386	43	227713	250.430	ug/l	93
17) Carbon Disulfide	2.508	76	180742	50.491	ug/l	100
18) Methyl Acetate	2.703	43	104903	52.414	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	229672	52.644	ug/l	99
20) Methylene Chloride	2.788	84	74694	49.467	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	67002	51.831	ug/l	99
22) Diisopropyl ether	3.764	45	244739	53.154	ug/l	89
23) Vinyl Acetate	3.721	43	1168503	273.902	ug/l	100
24) 1,1-Dichloroethane	3.605	63	134861	53.014	ug/l	100
25) 2-Butanone	4.556	43	348866	262.455	ug/l	100
26) 2,2-Dichloropropane	4.471	77	106396	53.020	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	78925	53.093	ug/l	100
28) Bromochloromethane	4.891	49	62878	53.967	ug/l	94
29) Tetrahydrofuran	5.001	42	231383	266.949	ug/l	99
30) Chloroform	5.093	83	134124	53.990	ug/l	98
31) Cyclohexane	5.465	56	120573	52.607	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	115285	54.565	ug/l	99
36) 1,1-Dichloropropene	5.690	75	97051	53.159	ug/l	99
37) Ethyl Acetate	4.715	43	131354	55.110	ug/l	99
38) Carbon Tetrachloride	5.672	117	94559	55.003	ug/l	97
39) Methylcyclohexane	7.379	83	120729	54.841	ug/l	98
40) Benzene	6.038	78	288690	54.770	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010924\
 Data File : VX039691.D
 Acq On : 09 Jan 2024 19:54
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 10 07:26:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010924W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 10 07:08:34 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	67204	55.752	ug/l	95
42) 1,2-Dichloroethane	6.086	62	109838	53.868	ug/l	99
43) Isopropyl Acetate	6.342	43	202813	55.604	ug/l	99
44) Trichloroethene	7.123	130	68372	53.990	ug/l	99
45) 1,2-Dichloropropane	7.428	63	78095	54.325	ug/l	99
46) Dibromomethane	7.580	93	53598	56.234	ug/l	94
47) Bromodichloromethane	7.818	83	99516	56.637	ug/l	98
48) Methyl methacrylate	7.690	41	99696	57.770	ug/l	95
49) 1,4-Dioxane	7.665	88	41347	1177.805	ug/l #	96
51) 4-Methyl-2-Pentanone	8.574	43	688094	283.411	ug/l	98
52) Toluene	8.714	92	172160	54.172	ug/l	94
53) t-1,3-Dichloropropene	8.976	75	108454	56.654	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	118535	57.045	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	71944	54.180	ug/l	94
56) Ethyl methacrylate	9.116	69	123820	57.864	ug/l	97
57) 1,3-Dichloropropane	9.305	76	124029	55.045	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	285376	282.989	ug/l	94
59) 2-Hexanone	9.433	43	552184	285.198	ug/l	100
60) Dibromochloromethane	9.519	129	71194	59.134	ug/l	100
61) 1,2-Dibromoethane	9.610	107	77520	56.528	ug/l	100
64) Tetrachloroethene	9.275	164	52597	50.874	ug/l	95
65) Chlorobenzene	10.080	112	179563	52.829	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	65398	54.946	ug/l	99
67) Ethyl Benzene	10.195	91	341203	54.164	ug/l	99
68) m/p-Xylenes	10.299	106	249329	107.405	ug/l	94
69) o-Xylene	10.640	106	123824	53.753	ug/l	94
70) Styrene	10.653	104	207362	56.391	ug/l	98
71) Bromoform	10.799	173	48915	56.990	ug/l #	99
73) Isopropylbenzene	10.964	105	333184	53.685	ug/l	99
74) N-amyl acetate	10.842	43	185624	56.151	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.214	83	126577	52.645	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	142964	70.755	ug/l	92
77) Bromobenzene	11.195	156	72557	52.035	ug/l	95
78) n-propylbenzene	11.305	91	399225	54.473	ug/l	97
79) 2-Chlorotoluene	11.366	91	242647	52.857	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	285231	53.816	ug/l	95
81) trans-1,4-Dichloro-2-b...	11.018	75	36933	56.095	ug/l	97
82) 4-Chlorotoluene	11.451	91	274673	52.801	ug/l	98
83) tert-Butylbenzene	11.713	119	272422	53.331	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	284147	53.782	ug/l	99
85) sec-Butylbenzene	11.890	105	350264	55.005	ug/l	99
86) p-Isopropyltoluene	12.006	119	280051	55.291	ug/l	97
87) 1,3-Dichlorobenzene	11.969	146	138265	52.515	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	136856	52.134	ug/l	99
89) n-Butylbenzene	12.335	91	262994	56.388	ug/l	96
90) Hexachloroethane	12.536	117	49621	56.199	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	135175	51.441	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	31015	55.634	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	80122	53.529	ug/l	97
94) Hexachlorobutadiene	13.725	225	33046	53.140	ug/l	98
95) Naphthalene	13.774	128	315157	52.808	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	82393	53.837	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010924\
Data File : VX039691.D
Acq On : 09 Jan 2024 19:54
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 10 07:26:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010924W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 10 07:08:34 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.550 min Scan# 732

Delta R.T. -0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:168 Resp: 100851

Ion Ratio Lower Upper

168 100

99 68.9 56.6 85.0



#2

Dichlorodifluoromethane

Concen: 53.913 ug/l

RT: 1.167 min Scan# 13

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 85 Resp: 73068

Ion Ratio Lower Upper

85 100

87 31.7 16.7 50.1





#3

Chloromethane

Concen: 51.352 ug/l

RT: 1.295 min Scan# 34

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 77662

Ion Ratio Lower Upper

50 100

52 33.5 25.5 38.3



#4

Vinyl Chloride

Concen: 53.484 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 62 Resp: 76694

Ion Ratio Lower Upper

62 100

64 31.8 25.4 38.0









#5
Bromomethane
Concen: 52.093 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 35788
Ion Ratio Lower Upper
94 100
96 93.0 74.6 111.8



#6
Chloroethane
Concen: 55.910 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 64 Resp: 38435
Ion Ratio Lower Upper
64 100
66 33.0 25.1 37.7









#7
Trichlorofluoromethane
Concen: 51.594 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 108760
Ion Ratio Lower Upper
101 100
103 64.0 51.4 77.2



#8
Diethyl Ether
Concen: 51.955 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 74 Resp: 39736
Ion Ratio Lower Upper
74 100
45 105.0 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.124 ug/l
RT: 2.325 min Scan# 203
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 66525
Ion Ratio Lower Upper
101 100
85 47.0 37.0 55.6
151 70.0 58.6 87.8



#10
Methyl Iodide
Concen: 54.713 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion:142 Resp: 77198
Ion Ratio Lower Upper
142 100
127 46.8 41.6 62.4
141 14.1 11.6 17.4





#11

Tert butyl alcohol

Concen: 257.736 ug/l

RT: 3.002 min Scan# 314

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 114311

Ion Ratio Lower Upper

59 100

57 10.4 6.0 9.0#



#12

1,1-Dichloroethene

Concen: 52.445 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion: 96 Resp: 65080


Ion Ratio Lower Upper

96 100

61 170.3 144.6 216.8

98 66.0 50.4 75.6





#13

Acrolein

Concen: 249.909 ug/l

RT: 2.240 min Scan# 189

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 56 Resp: 75239

Ion Ratio Lower Upper

56 100

55 71.0 57.1 85.7



#14

Allyl chloride

Concen: 53.163 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion: 41 Resp: 128807


Ion Ratio Lower Upper

41 100

39 67.2 55.8 83.6


76 31.0 24.3 36.5








#15
Acrylonitrile
Concen: 263.805 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 53 Resp: 239833




Ion	Ratio	Lower	Upper
53	100		
52	83.1	66.0	99.0
51	35.9	29.8	44.6




#16
Acetone
Concen: 250.430 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 43 Resp: 227713

Ion	Ratio	Lower	Upper
43	100		
58	31.2	21.9	32.9





#17

Carbon Disulfide

Concen: 50.491 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 180742

Ion Ratio Lower Upper

76 100

78 8.9 7.0 10.6



#18

Methyl Acetate

Concen: 52.414 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 43 Resp: 104903

Ion Ratio Lower Upper

43 100

74 22.3 17.8 26.8





#19

Methyl tert-butyl Ether

Concen: 52.644 ug/l

RT: 3.111 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 73 Resp: 229672

Ion Ratio Lower Upper

73 100

57 21.2 17.3 25.9



#20

Methylene Chloride

Concen: 49.467 ug/l

RT: 2.788 min Scan# 279

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Tgt Ion: 84 Resp: 74694




Ion Ratio Lower Upper


84 100

49 130.1 103.6 155.4

51 40.9 33.8 50.6

86 64.4 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 51.831 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 67002
Ion Ratio Lower Upper
96 100
61 148.0 119.2 178.8
98 63.0 48.7 73.1



#22
Diisopropyl ether
Concen: 53.154 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 45 Resp: 244739
Ion Ratio Lower Upper
45 100
43 73.6 69.7 104.5
87 25.3 20.8 31.2
59 12.3 9.0 13.4









#23
Vinyl Acetate
Concen: 273.902 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1168503
Ion Ratio Lower Upper
43 100
86 9.7 7.8 11.8



#24
1,1-Dichloroethane
Concen: 53.014 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 63 Resp: 134861
Ion Ratio Lower Upper
63 100
98 6.4 3.3 9.8
100 4.3 2.1 6.5









#25
2-Butanone
Concen: 262.455 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 348866
Ion Ratio Lower Upper
43 100
72 24.5 19.4 29.2



#26
2,2-Dichloropropane
Concen: 53.020 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 77 Resp: 106396
Ion Ratio Lower Upper
77 100
97 22.2 10.8 32.3





#27

cis-1,2-Dichloroethene

Concen: 53.093 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 78925

Ion Ratio Lower Upper

96 100

61 155.9 0.0 312.2

98 63.5 0.0 129.0



#28

Bromochloromethane

Concen: 53.967 ug/l

RT: 4.891 min Scan# 624

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion: 49 Resp: 62878


Ion Ratio Lower Upper

49 100

129 1.7 0.0 3.8

130 67.0 57.5 86.3





#29

Tetrahydrofuran

Concen: 266.949 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 231383

Ion Ratio Lower Upper

42 100

72 42.1 34.2 51.4

71 39.5 32.2 48.2



#30

Chloroform

Concen: 53.990 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 83 Resp: 134124

Ion Ratio Lower Upper

83 100


85 64.5 50.3 75.5








#31
Cyclohexane
Concen: 52.607 ug/l
RT: 5.465 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 120573
Ion Ratio Lower Upper
56 100
69 28.7 23.4 35.2
84 79.1 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 54.565 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54



Tgt Ion: 97 Resp: 115285
Ion Ratio Lower Upper
97 100
99 63.9 51.1 76.7
61 47.4 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 47.433 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 79922

Ion Ratio Lower Upper

65 100

67 52.6 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion: 114 Resp: 178628


Ion Ratio Lower Upper

114 100

63 22.9 0.0 45.0

88 16.4 0.0 36.6









#35
Dibromofluoromethane
Concen: 49.451 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 57722
Ion Ratio Lower Upper
113 100
111 105.9 82.9 124.3
192 16.9 13.3 19.9



#36
1,1-Dichloropropene
Concen: 53.159 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 75 Resp: 97051
Ion Ratio Lower Upper
75 100
110 34.4 17.0 50.9
77 31.1 24.7 37.1





#37

Ethyl Acetate

Concen: 55.110 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min


Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	43	Resp:	131354
Ion	Ratio	Lower	Upper
43	100		
61	13.4	10.6	16.0
70	10.0	8.3	12.5



#38

Carbon Tetrachloride

Concen: 55.003 ug/l




RT: 5.672 min Scan# 752


Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Tgt Ion:	117	Resp:	94559
Ion	Ratio	Lower	Upper
117	100		
119	95.7	78.2	117.2
121	28.1	24.4	36.6





#39

Methylcyclohexane

Concen: 54.841 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 120729

Ion Ratio Lower Upper

83 100

55 83.7 64.6 97.0

98 46.5 36.7 55.1



#40

Benzene

Concen: 54.770 ug/l

RT: 6.038 min Scan# 812

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 78 Resp: 288690

Ion Ratio Lower Upper

78 100

77 24.5 19.3 28.9









#41
Methacrylonitrile
Concen: 55.752 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 41 Resp: 67204




Ion	Ratio	Lower	Upper
41	100		
39	56.3	45.2	67.8
67	71.7	52.6	78.8
52	32.2	22.7	34.1




#42
1,2-Dichloroethane
Concen: 53.868 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 62 Resp: 109838

Ion	Ratio	Lower	Upper
62	100		
98	8.2	0.0	16.0





#43

Isopropyl Acetate

Concen: 55.604 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 43 Resp: 202813

Ion Ratio Lower Upper

43 100

61 18.7 15.2 22.8

87 12.4 9.8 14.6



#44

Trichloroethene

Concen: 53.990 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion:130 Resp: 68372

Ion Ratio Lower Upper

130 100

95 103.6 0.0 204.4









#45
1,2-Dichloropropane
Concen: 54.325 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 78095
Ion Ratio Lower Upper
63 100
65 30.8 24.3 36.5



#46
Dibromomethane
Concen: 56.234 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 93 Resp: 53598
Ion Ratio Lower Upper
93 100
95 80.4 66.6 100.0
174 79.2 69.4 104.0









#47
Bromodichloromethane
Concen: 56.637 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 99516
Ion Ratio Lower Upper
83 100
85 63.7 49.8 74.6
127 7.7 7.1 10.7



#48
Methyl methacrylate
Concen: 57.770 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 41 Resp: 99696
Ion Ratio Lower Upper
41 100
69 78.8 59.2 88.8
39 62.5 47.7 71.5





#49

1,4-Dioxane

Concen: 1177.805 ug/l

RT: 7.665 min Scan# 1079

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 41347

Ion Ratio Lower Upper

88 100

43 35.6 0.0 0.0#

58 71.9 60.1 90.1



#50

Toluene-d8

Concen: 50.302 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 98 Resp: 222426

Ion Ratio Lower Upper

98 100

100 64.1 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 283.411 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 688094
Ion Ratio Lower Upper
43 100
58 36.9 30.4 45.6



#52
Toluene
Concen: 54.172 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 92 Resp: 172160
Ion Ratio Lower Upper
92 100
91 176.4 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 56.654 ug/l
 RT: 8.976 min Scan# 1294
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 108454
 Ion Ratio Lower Upper
 75 100
 77 31.8 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 57.045 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54

Tgt Ion: 75 Resp: 118535
 Ion Ratio Lower Upper
 75 100
 77 31.1 25.3 37.9
 39 53.8 46.7 70.1









#55
 1,1,2-Trichloroethane
 Concen: 54.180 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 97	Resp: 71944
Ion Ratio	Lower Upper
97 100	
83 87.4	65.8 98.8
85 57.0	41.6 62.4
99 63.2	47.8 71.8



#56
 Ethyl methacrylate
 Concen: 57.864 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54

Tgt Ion: 69	Resp: 123820
Ion Ratio	Lower Upper
69 100	
41 72.0	56.8 85.2
39 44.3	32.0 48.0





#57

1,3-Dichloropropane

Concen: 55.045 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 124029

Ion Ratio Lower Upper

76 100

78 32.2 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 282.989 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 63 Resp: 285376

Ion Ratio Lower Upper

63 100

106 24.9 22.5 33.7









#59
2-Hexanone
Concen: 285.198 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 552184
Ion Ratio Lower Upper
43 100
58 52.0 26.2 78.5



#60
Dibromochloromethane
Concen: 59.134 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion:129 Resp: 71194
Ion Ratio Lower Upper
129 100
127 77.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 56.528 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 77520
Ion Ratio Lower Upper
107 100
109 93.6 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 51.390 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54


Tgt Ion: 95 Resp: 85094
Ion Ratio Lower Upper
95 100
174 65.5 0.0 135.6
176 64.0 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 163727
Ion Ratio Lower Upper
117 100
82 59.2 46.2 69.4
119 30.9 25.5 38.3




#64
Tetrachloroethene
Concen: 50.874 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54



Tgt Ion:164 Resp: 52597
Ion Ratio Lower Upper
164 100
166 131.3 102.7 154.1
129 106.1 80.4 120.6
131 103.8 76.5 114.7









#65
Chlorobenzene
Concen: 52.829 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 179563
Ion Ratio Lower Upper
112 100
114 32.8 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 54.946 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion:131 Resp: 65398
Ion Ratio Lower Upper
131 100
133 93.6 47.1 141.4
119 65.7 32.2 96.6









#67
Ethyl Benzene
Concen: 54.164 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 341203
Ion Ratio Lower Upper
91 100
106 30.6 24.9 37.3



#68
m/p-Xylenes
Concen: 107.405 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 106 Resp: 249329
Ion Ratio Lower Upper
106 100
91 216.0 165.5 248.3





#69

o-Xylene

Concen: 53.753 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 123824

Ion Ratio Lower Upper

106 100

91 227.6 108.9 326.8



#70

Styrene

Concen: 56.391 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion:104 Resp: 207362


Ion Ratio Lower Upper

104 100

78 56.6 43.6 65.4

103 55.0 43.6 65.4





#71

Bromoform

Concen: 56.990 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX039691.D





Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:173	Resp:	48915
Ion Ratio	Lower	Upper
173	100	
175	48.9	24.1 72.4
254	0.0	0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Tgt Ion:152	Resp:	74852
Ion Ratio	Lower	Upper
152	100	
115	97.9	43.3 129.9
150	175.2	0.0 345.0





#73

Isopropylbenzene

Concen: 53.685 ug/l

RT: 10.964 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 333184

Ion Ratio Lower Upper

105 100

120 25.1 12.9 38.6



#74

N-ethyl acetate

Concen: 56.151 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Tgt Ion: 43 Resp: 185624




Ion Ratio Lower Upper


43 100

70 38.3 32.9 49.3

55 22.9 19.3 28.9

61 22.6 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 52.645 ug/l

RT: 11.214 min Scan# 1661

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 126577

Ion Ratio Lower Upper

83 100

131 8.8 5.3 15.8

85 64.4 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 70.755 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 75 Resp: 142964

Ion Ratio Lower Upper

75 100

77 33.2 19.1 57.5





#77

Bromobenzene

Concen: 52.035 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 72557

Ion Ratio Lower Upper

156 100

77 188.4 88.4 265.3

158 96.4 48.0 144.0



#78

n-propylbenzene

Concen: 54.473 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 91 Resp: 399225

Ion Ratio Lower Upper

91 100

120 21.6 11.5 34.4





#79

2-Chlorotoluene

Concen: 52.857 ug/l

RT: 11.366 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 242647

Ion Ratio Lower Upper

91 100

126 31.0 16.3 48.8



#80

1,3,5-Trimethylbenzene

Concen: 53.816 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion:105 Resp: 285231

Ion Ratio Lower Upper

105 100

120 44.9 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 56.095 ug/l
RT: 11.018 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 36933
Ion Ratio Lower Upper
75 100
53 109.2 90.5 135.7
89 45.8 36.9 55.3



#82
4-Chlorotoluene
Concen: 52.801 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion: 91 Resp: 274673
Ion Ratio Lower Upper
91 100
126 27.5 14.2 42.6





#83

tert-Butylbenzene

Concen: 53.331 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 272422

Ion Ratio Lower Upper

119 100

91 65.4 31.2 93.6

134 21.9 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 53.782 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion:105 Resp: 284147

Ion Ratio Lower Upper

105 100

120 41.9 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 55.005 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 350264
 Ion Ratio Lower Upper
 105 100
 134 19.0 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 55.291 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX039691.D
 Acq: 09 Jan 2024 19:54

Tgt Ion:119 Resp: 280051
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.8 38.3
 91 28.2 13.1 39.3









#87
1,3-Dichlorobenzene
Concen: 52.515 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 138265
Ion Ratio Lower Upper
146 100
111 44.8 22.1 66.1
148 63.7 31.8 95.3



#88
1,4-Dichlorobenzene
Concen: 52.134 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX039691.D
Acq: 09 Jan 2024 19:54

Tgt Ion:146 Resp: 136856
Ion Ratio Lower Upper
146 100
111 44.1 21.3 63.7
148 63.0 31.8 95.4





#89

n-Butylbenzene

Concen: 56.388 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 262994

Ion Ratio Lower Upper

91 100

92 51.3 27.4 82.0

134 23.1 12.6 37.8



#90

Hexachloroethane

Concen: 56.199 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX039691.D




Acq: 09 Jan 2024 19:54


Tgt Ion: 117 Resp: 49621

Ion Ratio Lower Upper

117 100

201 61.7 32.9 98.6





#91

1,2-Dichlorobenzene

Concen: 51.441 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 135175

Ion Ratio Lower Upper

146 100

111 46.4 22.4 67.0

148 62.9 31.6 95.0



#92

1,2-Dibromo-3-Chloropropane

Concen: 55.634 ug/l

RT: 12.939 min Scan# 1944

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion: 75 Resp: 31015


Ion Ratio Lower Upper

75 100

155 71.9 36.1 108.5

157 91.1 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 53.529 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 80122

Ion Ratio Lower Upper

180 100

182 93.6 48.3 144.8

145 34.6 18.0 54.0



#94

Hexachlorobutadiene

Concen: 53.140 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54




Tgt Ion:225 Resp: 33046


Ion Ratio Lower Upper

225 100

223 65.3 31.1 93.2

227 65.2 32.5 97.5





#95

Naphthalene

Concen: 52.808 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 315157

Ion Ratio Lower Upper

128 100

127 12.8 10.3 15.5

129 11.0 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 53.837 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.000 min

Lab File: VX039691.D

Acq: 09 Jan 2024 19:54

Tgt Ion:180 Resp: 82393

Ion Ratio Lower Upper

180 100

182 92.7 46.9 140.6

145 35.8 17.8 53.5

