

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011019\
 Data File : VX006984.D
 Acq On : 10 Jan 2019 12:41
 Operator : JC/MD
 Sample : K1102-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-2

Quant Time: Jan 11 07:35:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	613343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	936374	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	870748	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	420242	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	322834	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
35) Dibromofluoromethane	5.50	113	289756	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
50) Toluene-d8	8.71	98	1121893	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	11.13	95	396233	51.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.56%	
Target Compounds						
16) Acetone	2.45	43	8045	2.539	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011019\
Data File : VX006984.D
Acq On : 10 Jan 2019 12:41
Operator : JC/MD
Sample : K1102-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-2

Quant Time: Jan 11 07:35:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

