

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011019\
 Data File : VX006986.D
 Acq On : 10 Jan 2019 13:34
 Operator : JC/MD
 Sample : K1102-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Jan 11 07:39:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	619901	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	950478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	880203	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	422310	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	329459	50.85	ug/l	0.00
Spiked Amount			Recovery	=	101.70%	
35) Dibromofluoromethane	5.50	113	296292	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
50) Toluene-d8	8.71	98	1127059	50.59	ug/l	0.00
Spiked Amount			Recovery	=	101.18%	
62) 4-Bromofluorobenzene	11.13	95	398547	51.31	ug/l	0.00
Spiked Amount			Recovery	=	102.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011019\
Data File : VX006986.D
Acq On : 10 Jan 2019 13:34
Operator : JC/MD
Sample : K1102-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Jan 11 07:39:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

