

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011019\
 Data File : VX006993.D
 Acq On : 10 Jan 2019 16:41
 Operator : JC/MD
 Sample : K1099-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-16

Quant Time: Jan 11 07:54:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	616900	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	942129	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	891752	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	430080	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	323967	50.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.50%	
35) Dibromofluoromethane	5.50	113	287335	47.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.24%	
50) Toluene-d8	8.71	98	1132943	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.13	95	401842	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
Target Compounds						
16) Acetone	2.45	43	30429	9.547	ug/l	98
18) Methyl Acetate	2.78	43	31099	3.369	ug/l	95
20) Methylene Chloride	2.85	84	19298	2.647	ug/l	98
25) 2-Butanone	4.70	43	14701	3.224	ug/l	93
43) Isopropyl Acetate	6.46	43	27900	2.014	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011019\
Data File : VX006993.D
Acq On : 10 Jan 2019 16:41
Operator : JC/MD
Sample : K1099-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-16

Quant Time: Jan 11 07:54:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

