

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX011020\
 Data File : VX014567.D
 Acq On : 11 Jan 2020 03:59
 Operator : JC/SP
 Sample : PB125993ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125993ZHE#13

Quant Time: Jan 11 05:22:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	483375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	753356	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	693277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	312777	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	268271	49.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.44%	
35) Dibromofluoromethane	5.49	113	226198	49.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.20%	
50) Toluene-d8	8.71	98	903291	51.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.14	95	311820	48.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.00%	


Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.20	85	254	0.295	ug/l #	42
6) Chloroethane	1.67	64	775	Below Cal	#	52
11) Tert butyl alcohol	3.01	59	2858	2.537	ug/l #	92
13) Acrolein	2.28	56	210	0.233	ug/l #	15
15) Acrylonitrile	3.14	53	647	0.238	ug/l #	50
16) Acetone	2.43	43	18282	2.622	ug/l	94
18) Methyl Acetate	2.76	43	6611	0.852	ug/l #	86
20) Methylene Chloride	2.85	84	7496	1.381	ug/l #	96
25) 2-Butanone	4.67	43	4628	1.148	ug/l	97
36) 1,1-Dichloropropene	5.65	75	32603	4.602	ug/l #	52
38) Carbon Tetrachloride	5.65	117	42063	6.452	ug/l #	18
41) Methacrylonitrile	5.03	41	959	0.234	ug/l #	40
43) Isopropyl Acetate	6.43	43	130225	10.476	ug/l	98
48) Methyl methacrylate	7.67	41	7988	1.342	ug/l #	15
49) 1,4-Dioxane	7.73	88	258	2.045	ug/l #	1
51) 4-Methyl-2-Pentanone	8.61	43	126610	16.714	ug/l #	50
54) cis-1,3-Dichloropropene	8.61	75	60391	7.053	ug/l #	57
57) 1,3-Dichloropropane	9.54	76	11074	1.225	ug/l #	42
59) 2-Hexanone	9.43	43	159950	26.685	ug/l #	23
74) N-amyl acetate	10.82	43	6201	0.640	ug/l #	60
76) 1,2,3-Trichloropropane	11.13	75	150786	23.300	ug/l #	34
80) 1,3,5-Trimethylbenzene	11.50	105	4497	0.247	ug/l	88
81) trans-1,4-Dichloro-2-buten	11.13	75	150786	61.602	ug/l #	12
88) 1,4-Dichlorobenzene	12.02	146	597	Below Cal		90
93) 1,2,4-Trichlorobenzene	13.68	180	217	Below Cal		82
96) 1,2,3-Trichlorobenzene	14.01	180	352	Below Cal	#	47


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014567.D
Acq On : 11 Jan 2020 03:59
Operator : JC/SP
Sample : PB125993ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#13

Quant Time: Jan 11 05:22:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. -0.00 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :


PB125993ZHE#13

Tot Ion:168 Resp: 483375

Ion Ratio Lower Upper

168 100

99 49.1 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

200000



150000

100000

50000

0

Time-->



#2

Dichlorodifluoromethane

Concen: 0.295 ug/l

RT: 1.20 min Scan# 52

Delta R.T. 0.01 min

Lab File: VX014567.D


Acq: 11 Jan 2020 03:59

Tgt Ion: 85 Resp: 254

Ion Ratio Lower Upper

85 100

87 0.0 16.5 49.5#



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.20

500

400


300


200

100

0

Time-->





#6

Chloroethane

Concen: Below Cal

RT: 1.67 min Scan# 128

Delta R.T. -0.05 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :





PB125993ZHE#13

Tot Ion: 64 Resp: 775

Ion Ratio Lower Upper

64 100

66 4.7 25.3 37.9#



#11

Tert butyl alcohol

Concen: 2.537 ug/l

RT: 3.01 min Scan# 348

Delta R.T. -0.02 min

Lab File: VX014567.D




Acq: 11 Jan 2020 03:59


Tgt Ion: 59 Resp: 2858

Ion Ratio Lower Upper

59 100

57 13.5 8.5 12.7#









#13
Acrolein
Concen: 0.233 ug/l
RT: 2.28 min Scan# 228
Delta R.T. -0.01 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59




Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#13


Tot Ion: 56 Resp: 210
Ion Ratio Lower Upper
56 100
55 0.0 55.0 82.6#



#15
Acrylonitrile
Concen: 0.238 ug/l
RT: 3.14 min Scan# 369
Delta R.T. 0.01 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59

Tgt Ion: 53 Resp: 647
Ion Ratio Lower Upper
53 100
52 34.3 67.0 100.4#
51 58.1 28.5 42.7#









#16
Acetone
Concen: 2.622 ug/l
RT: 2.43 min Scan# 253
Delta R.T. 0.00 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59




Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#13


Tot Ion: 43 Resp: 18282
Ion Ratio Lower Upper
43 100
58 28.4 25.4 38.0



#18
Methyl Acetate
Concen: 0.852 ug/l
RT: 2.76 min Scan# 308
Delta R.T. -0.00 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59

Tgt Ion: 43 Resp: 6611
Ion Ratio Lower Upper
43 100
74 17.4 19.4 29.2#









#20
Methvlene Chloride
Concen: 1.381 ug/l
RT: 2.85 min Scan# 322
Delta R.T. -0.00 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59




Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#13


Tot Ion: 84 Resp: 7496
Ion Ratio Lower Upper
84 100
49 118.6 95.8 143.8
51 47.3 30.2 45.4#
86 66.4 51.8 77.6



#25
2-Butanone
Concen: 1.148 ug/l
RT: 4.67 min Scan# 621
Delta R.T. 0.02 min
Lab File: VX014567.D
Acq: 11 Jan 2020 03:59

Tgt Ion: 43 Resp: 4628
Ion Ratio Lower Upper
43 100
72 24.5 20.6 31.0





#33

1,2-Dichloroethane-d4

Concen: 49.716 ug/l

RT: 6.06 min Scan# 848

Delta R.T. 0.01 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :


PB125993ZHE#13

Tot Ion: 65 Resp: 268271

Ion Ratio Lower Upper

65 100

67 53.5 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

120000

100000

80000



60000

40000

20000

0

Time--> 5.90 6.00 6.10 6.20



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 978

Delta R.T. -0.00 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59


Tgt Ion: 114 Resp: 753356

Ion Ratio Lower Upper

114 100

63 21.5 0.0 39.6

88 15.3 0.0 30.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0

6.85

400000


300000


200000

100000

0

Time--> 6.80 6.90 7.00





#35

Dibromofluoromethane

Concen: 49.601 ug/l

RT: 5.49 min Scan# 755

Delta R.T. -0.00 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13


Tot Ion:113 Resp: 226198

Ion Ratio Lower Upper




113 100

111 101.3 81.5 122.3

192 23.6 19.2 28.8



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 4.602 ug/l

RT: 5.65 min Scan# 781

Delta R.T. -0.15 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59


Tgt Ion: 75 Resp: 32603

Ion Ratio Lower Upper



75 100


110 11.9 18.6 55.8#

77 0.5 24.5 36.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): VX0





#38

Carbon Tetrachloride

Concen: 6.452 ug/l

RT: 5.65 min Scan# 782

Delta R.T. -0.12 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13


Tot Ion:117 Resp: 42063

Ion Ratio Lower Upper

117 100

119 5.0 73.8 110.8#

121 0.9 23.4 35.0#






Abundance

Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.234 ug/l

RT: 5.03 min Scan# 680

Delta R.T. 0.01 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Tgt Ion: 41 Resp: 959


Ion Ratio Lower Upper

41 100

39 86.0 43.6 65.4#

67 0.0 57.9 86.9#

52 14.8 23.0 34.6#





Abundance


Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0









#43

Isopropyl Acetate
 Concen: 10.476 ug/l
 RT: 6.43 min Scan# 910
 Delta R.T. -0.00 min
 Lab File: VX014567.D
 Acq: 11 Jan 2020 03:59

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125993ZHE#13




Tot Ion	43	Resp	130225
Ion Ratio	Lower	Upper	
43	100		
61	19.0	16.6	24.8
87	13.5	10.8	16.2




#48

Methyl methacrylate
 Concen: 1.342 ug/l
 RT: 7.67 min Scan# 1112
 Delta R.T. -0.10 min
 Lab File: VX014567.D
 Acq: 11 Jan 2020 03:59

Tgt Ion	41	Resp	7988
Ion Ratio	Lower	Upper	
41	100		
69	0.0	65.5	98.3#
39	0.0	43.9	65.9#





#49

1,4-Dioxane

Concen: 2.045 ug/l

RT: 7.73 min Scan# 1123

Delta R.T. 0.01 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13





Tot Ion: 88 Resp: 258

Ion Ratio Lower Upper

88 100

43 0.0 27.4 41.0#

58 519.8 57.1 85.7#



#50

Toluene-d8

Concen: 51.041 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. -0.00 min

Lab File: VX014567.D




Acq: 11 Jan 2020 03:59


Tgt Ion: 98 Resp: 903291

Ion Ratio Lower Upper

98 100

100 65.6 53.2 79.8





#51

4-Methyl-2-Pentanone

Concen: 16.714 ug/l

RT: 8.61 min Scan# 1266

Delta R.T. -0.03 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :





PB125993ZHE#13

Tot Ion: 43 Resp: 126610

Ion Ratio Lower Upper

43 100

58 9.6 32.4 48.6#



#54

cis-1,3-Dichloropropene

Concen: 7.053 ug/l

RT: 8.61 min Scan# 1267

Delta R.T. 0.18 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59




Tgt Ion: 75 Resp: 60391


Ion Ratio Lower Upper

75 100

77 0.8 25.0 37.4#

39 26.5 40.7 61.1#





#57

1,3-Dichloropropane

Concen: 1.225 ug/l

RT: 9.54 min Scan# 1419

Delta R.T. 0.17 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :





PB125993ZHE#13

Tot Ion: 76 Resp: 11074

Ion Ratio Lower Upper

76 100

78 0.0 26.4 39.6#



#59

2-Hexanone

Concen: 26.685 ug/l

RT: 9.43 min Scan# 1402

Delta R.T. -0.05 min

Lab File: VX014567.D




Acq: 11 Jan 2020 03:59


Tgt Ion: 43 Resp: 159950

Ion Ratio Lower Upper

43 100

58 0.0 27.9 83.7#





#62

4-Bromofluorobenzene

Concen: 47.995 ug/l

RT: 11.14 min Scan# 1681

Delta R.T. 0.01 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13


Tot Ion: 95 Resp: 311820

Ion Ratio Lower Upper

95 100

174 88.0 0.0 180.8




176 86.0 0.0 172.8



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1513

Delta R.T. -0.00 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59


Tgt Ion: 117 Resp: 693277

Ion Ratio Lower Upper

117 100

82 53.4 41.8 62.6



119 32.6 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1835

Delta R.T. -0.00 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13


Tot Ion:152 Resp: 312777

Ion Ratio Lower Upper

152 100

115 54.8 39.5 118.3




150 157.1 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.640 ug/l

RT: 10.82 min Scan# 1629

Delta R.T. -0.07 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Tgt Ion: 43 Resp: 6201


Ion Ratio Lower Upper

43 100

70 6.7 35.1 52.7#

55 15.1 19.3 28.9#

61 10.3 20.0 30.0#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0





#76

1,2,3-Trichloropropane

Concen: 23.300 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. -0.16 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :


PB125993ZHE#13

Tot Ion: 75 Resp: 150786

Ion Ratio Lower Upper




75 100

77 1.3 21.9 65.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0



#80

1,3,5-Trimethylbenzene

Concen: 0.247 ug/l

RT: 11.50 min Scan# 1741

Delta R.T. -0.00 min

Lab File: VX014567.D


Acq: 11 Jan 2020 03:59

Tgt Ion: 105 Resp: 4497

Ion Ratio Lower Upper



105 100


120 59.0 25.2 75.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 61.602 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. 0.06 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59

Instrument :

MSVOA_X

ClientSampleId :

PB125993ZHE#13




Tot Ion: 75 Resp: 150786

Ion Ratio Lower Upper


75 100

53 0.2 74.8 112.2#

89 0.1 36.3 54.5#



Time-->



#88

1,4-Dichlorobenzene

Concen: Below Cal

RT: 12.02 min Scan# 1826

Delta R.T. -0.07 min

Lab File: VX014567.D

Acq: 11 Jan 2020 03:59



Tgt Ion: 146 Resp: 597

Ion Ratio Lower Upper


146 100

111 43.0 18.1 54.1

148 71.7 32.1 96.3



Time-->



#96
 1,2,3-Trichlorobenzene
 Concen: Below Cal
 RT: 14.01 min Scan# 2153
 Delta R.T. 0.00 min
 Lab File: VX014567.D
 Acq: 11 Jan 2020 03:59

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125993ZHE#13

Tot Ion	Ratio	Lower	Upper
180	100		
182	44.9	47.9	143.6#
145	0.0	14.8	44.3#

