

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014523.D
 Acq On : 10 Jan 2020 09:23
 Operator : JC/SP
 Sample : VX0110WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0110WBL01

Quant Time: Jan 10 12:34:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	496444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	767648	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	688961	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	313975	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	275300	49.68	ug/l	0.00
Spiked Amount						
			Recovery	=	99.36%	
35) Dibromofluoromethane	5.48	113	232467	50.03	ug/l	0.00
Spiked Amount						
			Recovery	=	100.06%	
50) Toluene-d8	8.71	98	908949	50.40	ug/l	0.00
Spiked Amount						
			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.13	95	307462	46.44	ug/l	0.00
Spiked Amount						
			Recovery	=	92.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014523.D
Acq On : 10 Jan 2020 09:23
Operator : JC/SP
Sample : VX0110WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0110WBL01

Quant Time: Jan 10 12:34:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

