

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014533.D
 Acq On : 10 Jan 2020 13:39
 Operator : JC/SP
 Sample : L1080-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Jan 10 14:09:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	496190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	759477	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	672729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	306974	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	275716	49.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.56%	
35) Dibromofluoromethane	5.49	113	223050	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	
50) Toluene-d8	8.71	98	891497	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.13	95	302622	46.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014533.D
Acq On : 10 Jan 2020 13:39
Operator : JC/SP
Sample : L1080-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Jan 10 14:09:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

