

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014534.D
 Acq On : 10 Jan 2020 14:02
 Operator : JC/SP
 Sample : L1080-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01092020

Quant Time: Jan 10 15:49:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	510585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	769092	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	692294	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	309976	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	281271	49.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.70%	
35) Dibromofluoromethane	5.49	113	228628	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
50) Toluene-d8	8.71	98	922741	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	11.14	95	311219	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
Target Compounds						
16) Acetone	2.43	43	63438	18.325	ug/l	99
20) Methylene Chloride	2.85	84	8856	1.544	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014534.D
Acq On : 10 Jan 2020 14:02
Operator : JC/SP
Sample : L1080-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01092020

Quant Time: Jan 10 15:49:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

