

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014560.D
 Acq On : 11 Jan 2020 01:16
 Operator : JC/SP
 Sample : L1070-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-200059

Quant Time: Jan 11 05:30:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	477497	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	743623	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	681375	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	314640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	270281	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
35) Dibromofluoromethane	5.49	113	217640	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
50) Toluene-d8	8.71	98	887167	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
62) 4-Bromofluorobenzene	11.13	95	309860	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
Target Compounds						
16) Acetone	2.43	43	28989	6.781	ug/l	99
20) Methylene Chloride	2.85	84	11792	2.199	ug/l #	89
43) Isopropyl Acetate	6.43	43	120875	9.851	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014560.D
Acq On : 11 Jan 2020 01:16
Operator : JC/SP
Sample : L1070-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-200059

Quant Time: Jan 11 05:30:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

