

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014563.D
 Acq On : 11 Jan 2020 02:26
 Operator : JC/SP
 Sample : PB125993ZHE#09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125993ZHE#09

Quant Time: Jan 11 05:46:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	480590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	750780	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	688575	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	311592	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	270882	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
35) Dibromofluoromethane	5.49	113	223094	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.71	98	895169	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
62) 4-Bromofluorobenzene	11.13	95	308856	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
Target Compounds						
20) Methylene Chloride	2.84	84	6246	1.157	ug/l #	91
43) Isopropyl Acetate	6.44	43	132880	10.726	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014563.D
Acq On : 11 Jan 2020 02:26
Operator : JC/SP
Sample : PB125993ZHE#09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#09

Quant Time: Jan 11 05:46:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

