

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014567.D
 Acq On : 11 Jan 2020 03:59
 Operator : JC/SP
 Sample : PB125993ZHE#13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125993ZHE#13

Quant Time: Jan 11 05:54:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	483375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	753356	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	693277	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	312777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	268271	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
35) Dibromofluoromethane	5.49	113	226198	49.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.20%	
50) Toluene-d8	8.71	98	903291	51.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.08%	
62) 4-Bromofluorobenzene	11.14	95	311820	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
Target Compounds						
20) Methylene Chloride	2.85	84	7496	1.381	ug/l #	96
43) Isopropyl Acetate	6.43	43	130225	10.476	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014567.D
Acq On : 11 Jan 2020 03:59
Operator : JC/SP
Sample : PB125993ZHE#13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125993ZHE#13

Quant Time: Jan 11 05:54:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

