

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011020\
 Data File : VX014551.D
 Acq On : 10 Jan 2020 21:47
 Operator : JC/SP
 Sample : VX0110WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0110WBL02

Quant Time: Jan 11 00:32:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	489954	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	754607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	686441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	313155	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	273871	50.07	ug/l	0.00
Spiked Amount						
			Recovery	=	100.14%	
35) Dibromofluoromethane	5.49	113	224428	49.13	ug/l	0.00
Spiked Amount						
			Recovery	=	98.26%	
50) Toluene-d8	8.71	98	893206	50.39	ug/l	0.00
Spiked Amount						
			Recovery	=	100.78%	
62) 4-Bromofluorobenzene	11.13	95	304937	46.86	ug/l	0.00
Spiked Amount						
			Recovery	=	93.72%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011020\
Data File : VX014551.D
Acq On : 10 Jan 2020 21:47
Operator : JC/SP
Sample : VX0110WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0110WBL02

Quant Time: Jan 11 00:32:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.00 min

Lab File: VX014551.D

Acq: 10 Jan 2020 21:47

Instrument :

MSVOA_X

ClientSampleId :


VX0110WBL02

Tot Ion:168 Resp: 489954

Ion Ratio Lower Upper

168 100

99 49.6 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 50.072 ug/l

RT: 6.05 min Scan# 847

Delta R.T. 0.00 min

Lab File: VX014551.D


Acq: 10 Jan 2020 21:47

Tgt Ion: 65 Resp: 273871

Ion Ratio Lower Upper

65 100

67 53.3 0.0 106.2



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

120000

100000

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 979

Delta R.T. 0.01 min

Lab File: VX014551.D

Acq: 10 Jan 2020 21:47

Instrument :

MSVOA_X

ClientSampleId :

VX0110WBL02


Tot Ion:114 Resp: 754607

Ion Ratio Lower Upper

114 100

63 19.1 0.0 39.6




88 15.0 0.0 30.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 49.131 ug/l

RT: 5.49 min Scan# 755

Delta R.T. 0.00 min

Lab File: VX014551.D

Acq: 10 Jan 2020 21:47


Tgt Ion:113 Resp: 224428

Ion Ratio Lower Upper

113 100

111 103.5 81.5 122.3



192 23.7 19.2 28.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.388 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. 0.00 min

Lab File: VX014551.D

Acq: 10 Jan 2020 21:47

Instrument :

MSVOA_X

ClientSampleId :





VX0110WBL02

Tot Ion: 98 Resp: 893206

Ion Ratio Lower Upper

98 100

100 64.9 53.2 79.8



#62

4-Bromofluorobenzene

Concen: 46.858 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. 0.00 min

Lab File: VX014551.D

Acq: 10 Jan 2020 21:47




Tgt Ion: 95 Resp: 304937


Ion Ratio Lower Upper

95 100

174 89.2 0.0 180.8

176 87.7 0.0 172.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX014551.D
Acq: 10 Jan 2020 21:47

Instrument :
MSVOA_X
ClientSampleId :
VX0110WBL02

Tot Ion:117 Resp: 686441
Ion Ratio Lower Upper
117 100
82 52.4 41.8 62.6
119 32.2 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX014551.D
Acq: 10 Jan 2020 21:47

Tgt Ion:152 Resp: 313155
Ion Ratio Lower Upper
152 100
115 55.2 39.5 118.3
150 156.2 0.0 350.2

