

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
 Data File : VX026402.D
 Acq On : 10 Jan 2022 15:07
 Operator : JC/MD
 Sample : N1073-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-RW7B-IDWGW-20220106

Quant Time: Jan 10 16:09:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	18352	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	104225	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	91065	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	45598	30.899	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.000%
60) 4-Bromofluorobenzene	11.079	95	42782	28.150	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	93.833%
63) Toluene-d8	8.653	98	121232	28.979	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.600%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.337	101	36171	34.252	ug/l	98
10) 1,1-Dichloroethene	2.325	96	23997	24.325	ug/l	98
21) 1,1-Dichloroethane	3.617	63	5040	2.481	ug/l	99
22) cis-1,2-Dichloroethene	4.501	96	3688	2.973	ug/l	84
25) Chloroform	5.105	83	2412	1.146	ug/l	98
32) 1,1,1-Trichloroethane	5.391	97	3524	1.798	ug/l	96
33) Carbon Tetrachloride	5.678	117	3702	2.132	ug/l	89
37) Trichloroethene	7.129	130	529764	410.965	ug/l	100
50) 1,1,2-Trichloroethane	9.153	97	1343	1.218	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
Data File : VX026402.D
Acq On : 10 Jan 2022 15:07
Operator : JC/MD
Sample : N1073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-RW7B-IDWGW-20220106

Quant Time: Jan 10 16:09:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

