

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
 Data File : VX026406.D
 Acq On : 10 Jan 2022 16:40
 Operator : JC/MD
 Sample : N1046-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD(JAN)

Quant Time: Jan 11 04:56:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	18809	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.769	114	103490	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	88866	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	45399	30.017	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.067%
60) 4-Bromofluorobenzene	11.079	95	42679	28.777	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.933%
63) Toluene-d8	8.653	98	121894	29.858	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.533%
Target Compounds						
					Qvalue	
15) Acetone	2.386	58	95365	513.857	ug/l	96
16) Carbon Disulfide	2.520	76	6780	2.375	ug/l	98
30) 2-Butanone	4.568	43	8325	8.968	ug/l #	75
58) 4-Methyl-2-Pentanone	8.574	43	6882	4.152	ug/l	99
62) Toluene	8.720	91	22803	4.806	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
Data File : VX026406.D
Acq On : 10 Jan 2022 16:40
Operator : JC/MD
Sample : N1046-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD(JAN)

Quant Time: Jan 11 04:56:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

