

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
 Data File : VX026409.D
 Acq On : 10 Jan 2022 17:49
 Operator : JC/MD
 Sample : N1036-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 QUARTERLY-OUTFALL-002

Quant Time: Jan 11 04:57:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jan 10 12:50:55 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.904	128	19087	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	107396	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	89542	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	47646	31.044	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	103.467%
60) 4-Bromofluorobenzene	11.079	95	42318	28.318	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.400%
63) Toluene-d8	8.653	98	124560	30.281	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.933%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
Data File : VX026409.D
Acq On : 10 Jan 2022 17:49
Operator : JC/MD
Sample : N1036-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
QUARTERLY-OUTFALL-002

Quant Time: Jan 11 04:57:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jan 10 12:50:55 2022
Response via : Initial Calibration

