



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011022\
 Data File : VX026391.D
 Acq On : 10 Jan 2022 09:37
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X011022W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Jan 10 12:50:55 2022



AutoFind: Averaged scan 1634 to 1647; Bkg corrected with scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	1668	PASS
75	95	30	60	50.7	4534	PASS
95	95	100	100	100.0	8946	PASS
96	95	5	9	6.6	594	PASS
173	174	0.00	2	1.1	68	PASS
174	95	50	100	71.3	6382	PASS
175	174	5	9	7.5	478	PASS
176	174	95	101	97.4	6214	PASS
177	176	5	9	6.6	413	PASS