

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011023\
 Data File : VX033682.D
 Acq On : 10 Jan 2023 14:05
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jan 11 00:33:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	174554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	255371	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	82267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	95558	48.495	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.980%		
35) Dibromofluoromethane	5.379	113	80420	47.911	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.820%		
50) Toluene-d8	8.647	98	296709	49.456	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.920%		
62) 4-Bromofluorobenzene	11.079	95	106444	50.054	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.100%		
Target Compounds						
16) Acetone	2.404	43	1104	Below Cal	Qvalue	97
20) Methylene Chloride	2.788	84	3517	Below Cal		93
37) Ethyl Acetate	4.727	43	1034	Below Cal	#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011023\
Data File : VX033682.D
Acq On : 10 Jan 2023 14:05
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jan 11 00:33:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

