

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039700.D
 Acq On : 10 Jan 2024 13:06
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 11 03:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	169496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83876	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70317	48.981	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.385	113	55086	48.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.647	98	212798	49.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.840%
62) 4-Bromofluorobenzene	11.079	95	92269	54.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.600%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	743	0.670	ug/l	95
3) Chloromethane	1.294	50	467	0.369	ug/l	95
4) Vinyl Chloride	1.374	62	637	0.533	ug/l	96
5) Bromomethane	1.611	94	587	0.861	ug/l #	65
6) Chloroethane	1.697	64	554	0.815	ug/l	92
7) Trichlorofluoromethane	1.886	101	839	0.487	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.331	101	679	0.635	ug/l	89
10) Methyl Iodide	2.459	142	606	0.537	ug/l #	72
11) Tert butyl alcohol	2.946	59	728	1.948	ug/l #	78
12) 1,1-Dichloroethene	2.319	96	649	0.629	ug/l #	81
13) Acrolein	2.233	56	710	Below Cal		98
14) Allyl chloride	2.654	41	837	0.412	ug/l #	77
15) Acrylonitrile	3.074	53	1964	2.544	ug/l	96
16) Acetone	2.379	43	749	0.928	ug/l	90
17) Carbon Disulfide	2.514	76	7930	2.675	ug/l	99
18) Methyl Acetate	2.709	43	549	0.325	ug/l	95
20) Methylene Chloride	2.794	84	862	0.706	ug/l #	91
21) trans-1,2-Dichloroethene	3.099	96	1586	1.461	ug/l	84
23) Vinyl Acetate	3.745	43	2723	0.756	ug/l #	91
25) 2-Butanone	4.617	43	666	0.574	ug/l #	51
27) cis-1,2-Dichloroethene	4.501	96	668	0.537	ug/l	86
28) Bromochloromethane	4.897	49	385	0.399	ug/l #	90
29) Tetrahydrofuran	5.050	42	241	0.323	ug/l #	34
30) Chloroform	5.098	83	435	0.207	ug/l #	70
31) Cyclohexane	5.470	56	1062	0.561	ug/l	98
36) 1,1-Dichloropropene	5.702	75	1459	0.821	ug/l #	86
38) Carbon Tetrachloride	5.672	117	390	0.240	ug/l #	72
39) Methylcyclohexane	7.385	83	1040	0.487	ug/l #	83
40) Benzene	6.050	78	1398	0.278	ug/l #	77
42) 1,2-Dichloroethane	6.086	62	582	0.304	ug/l #	77
44) Trichloroethene	7.129	130	887	0.719	ug/l	89
46) Dibromomethane	7.592	93	483	0.518	ug/l #	85
49) 1,4-Dioxane	7.677	88	156	4.716	ug/l #	85
51) 4-Methyl-2-Pentanone	8.580	43	1098	0.458	ug/l	91
52) Toluene	8.726	92	1331	0.443	ug/l	99
53) t-1,3-Dichloropropene	8.988	75	1104	0.603	ug/l #	76

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039700.D
 Acq On : 10 Jan 2024 13:06
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 11 03:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
54) cis-1,3-Dichloropropene	8.372	75	850	0.422	ug/l #	71
58) 2-Chloroethyl Vinyl ether	8.256	63	927	0.947	ug/l #	67
59) 2-Hexanone	9.439	43	1761	0.929	ug/l	81
61) 1,2-Dibromoethane	9.604	107	488	0.368	ug/l #	75
64) Tetrachloroethene	9.275	164	857	0.756	ug/l	94
65) Chlorobenzene	10.079	112	1968	0.542	ug/l #	84
67) Ethyl Benzene	10.195	91	3070	0.456	ug/l #	85
68) m/p-Xylenes	10.305	106	2361	0.953	ug/l	85
69) o-Xylene	10.640	106	960	0.391	ug/l	70
70) Styrene	10.665	104	2009	0.501	ug/l	96
73) Isopropylbenzene	10.963	105	2566	0.384	ug/l	98
74) N-amyl acetate	10.847	43	1051	0.302	ug/l #	82
76) 1,2,3-Trichloropropane	11.079	75	49909	22.951	ug/l #	39
77) Bromobenzene	11.201	156	934	0.628	ug/l	94
78) n-propylbenzene	11.305	91	5258	0.651	ug/l	98
79) 2-Chlorotoluene	11.366	91	2855	0.572	ug/l	92
80) 1,3,5-Trimethylbenzene	11.451	105	2594	0.457	ug/l	93
81) trans-1,4-Dichloro-2-b...	11.079	75	49909	66.292	ug/l #	5
82) 4-Chlorotoluene	11.457	91	4770	0.842	ug/l	99
83) tert-Butylbenzene	11.719	119	1855	0.343	ug/l	85
84) 1,2,4-Trimethylbenzene	11.756	105	3094	0.547	ug/l	94
85) sec-Butylbenzene	11.890	105	4175	0.594	ug/l	98
86) p-Isopropyltoluene	12.006	119	3642	0.648	ug/l	92
87) 1,3-Dichlorobenzene	11.969	146	3149	1.105	ug/l	92
88) 1,4-Dichlorobenzene	11.969	146	3149	1.070	ug/l	93
89) n-Butylbenzene	12.335	91	6229	1.148	ug/l	98
90) Hexachloroethane	12.536	117	249	0.263	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	2160	0.767	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	12.951	75	272	0.435	ug/l #	58
93) 1,2,4-Trichlorobenzene	13.591	180	3706	2.152	ug/l	95
94) Hexachlorobutadiene	13.725	225	1555	2.172	ug/l	96
95) Naphthalene	13.780	128	14210	2.242	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	3544	2.039	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039700.D
 Acq On : 10 Jan 2024 13:06
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Jan 11 03:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 85700
Ion Ratio Lower Upper
168 100
99 68.1 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 0.670 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 85 Resp: 743
Ion Ratio Lower Upper
85 100
87 30.6 16.7 50.1





#3

Chloromethane

Concen: 0.369 ug/l

RT: 1.294 min Scan# 34

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 467

Ion Ratio Lower Upper

50 100

52 29.1 25.5 38.3



#4

Vinyl Chloride

Concen: 0.533 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX039700.D




Acq: 10 Jan 2024 13:06


Tgt Ion: 62 Resp: 637

Ion Ratio Lower Upper

62 100


64 29.2 25.4 38.0








#5
Bromomethane
Concen: 0.861 ug/l
RT: 1.611 min Scan# 80
Delta R.T. 0.012 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 94 Resp: 587
Ion Ratio Lower Upper
94 100
96 60.0 74.6 111.8#



#6
Chloroethane
Concen: 0.815 ug/l
RT: 1.697 min Scan# 100
Delta R.T. 0.018 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion: 64 Resp: 554
Ion Ratio Lower Upper
64 100
66 26.8 25.1 37.7








#7
 Trichlorofluoromethane
 Concen: 0.487 ug/l
 RT: 1.886 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion:101 Resp: 839
 Ion Ratio Lower Upper
 101 100
 103 63.3 51.4 77.2




#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.635 ug/l
 RT: 2.331 min Scan# 204
 Delta R.T. -0.000 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06




Tgt Ion:101 Resp: 679
 Ion Ratio Lower Upper
 101 100
 85 42.4 37.0 55.6
 151 60.8 58.6 87.8








#10
Methyl Iodide
Concen: 0.537 ug/l
RT: 2.459 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:142 Resp: 606
Ion Ratio Lower Upper
142 100
127 33.7 41.6 62.4#
141 0.0 11.6 17.4#



#11
Tert butyl alcohol
Concen: 1.948 ug/l
RT: 2.946 min Scan# 305
Delta R.T. -0.012 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 59 Resp: 728
Ion Ratio Lower Upper
59 100
57 0.0 6.0 9.0#





#12

1,1-Dichloroethene

Concen: 0.629 ug/l

RT: 2.319 min Scan# 20

Delta R.T. -0.000 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 96 Resp: 649

Ion Ratio Lower Upper

96 100

61 153.5 144.6 216.8

98 48.0 50.4 75.6#



#13

Acrolein

Concen: Below Cal

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX039700.D




Acq: 10 Jan 2024 13:06


Tgt Ion: 56 Resp: 710

Ion Ratio Lower Upper

56 100


55 70.0 57.1 85.7








#14
 Allyl chloride
 Concen: 0.412 ug/l
 RT: 2.654 min Scan# 21
 Delta R.T. -0.012 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion: 41 Resp: 837
 Ion Ratio Lower Upper
 41 100
 39 62.2 55.8 83.6
 76 0.0 24.3 36.5#



#15
 Acrylonitrile
 Concen: 2.544 ug/l
 RT: 3.074 min Scan# 326
 Delta R.T. 0.012 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06


Tgt Ion: 53 Resp: 1964
 Ion Ratio Lower Upper
 53 100
 52 78.1 66.0 99.0
 51 38.3 29.8 44.6








#16
Acetone
Concen: 0.928 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 749
Ion Ratio Lower Upper
43 100
58 32.6 21.9 32.9



#17
Carbon Disulfide
Concen: 2.675 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 76 Resp: 7930
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6









#18
Methyl Acetate
Concen: 0.325 ug/l
RT: 2.709 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 549
Ion Ratio Lower Upper
43 100
74 24.8 17.8 26.8



#20
Methylene Chloride
Concen: 0.706 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion: 84 Resp: 862
Ion Ratio Lower Upper
84 100
49 139.5 103.6 155.4
51 55.6 33.8 50.6#
86 59.9 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 1.461 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 1586
Ion Ratio Lower Upper
96 100
61 122.7 119.2 178.8
98 57.1 48.7 73.1



#23
Vinyl Acetate
Concen: 0.756 ug/l
RT: 3.745 min Scan# 436
Delta R.T. 0.024 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 43 Resp: 2723
Ion Ratio Lower Upper
43 100
86 13.3 7.8 11.8#









#25
2-Butanone
Concen: 0.574 ug/l
RT: 4.617 min Scan# 51
Delta R.T. 0.067 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 666
Ion Ratio Lower Upper
43 100
72 0.0 19.4 29.2#



#27
cis-1,2-Dichloroethene
Concen: 0.537 ug/l
RT: 4.501 min Scan# 560
Delta R.T. 0.018 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion: 96 Resp: 668
Ion Ratio Lower Upper
96 100
61 157.8 0.0 312.2
98 28.1 0.0 129.0








#28
Bromochloromethane
Concen: 0.399 ug/l
RT: 4.897 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 49 Resp: 385
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 63.1 57.5 86.3




#29
Tetrahydrofuran
Concen: 0.323 ug/l
RT: 5.050 min Scan# 650
Delta R.T. 0.049 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06



Tgt Ion: 42 Resp: 241
Ion Ratio Lower Upper
42 100
72 0.0 34.2 51.4#
71 0.0 32.2 48.2#





#30

Chloroform

Concen: 0.207 ug/l

RT: 5.098 min Scan# 61

Delta R.T. 0.006 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 83 Resp: 435

Ion Ratio Lower Upper

83 100

85 39.7 50.3 75.5#



#31

Cyclohexane

Concen: 0.561 ug/l

RT: 5.470 min Scan# 719

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion: 56 Resp: 1062


Ion Ratio Lower Upper

56 100

69 28.5 23.4 35.2

84 80.2 65.9 98.9









#33
 1,2-Dichloroethane-d4
 Concen: 48.981 ug/l
 RT: 5.952 min Scan# 798
 Delta R.T. -0.000 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 65 Resp: 70317
 Ion Ratio Lower Upper
 65 100
 67 51.8 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.763 min Scan# 931
 Delta R.T. 0.006 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06

Tgt Ion: 114 Resp: 169496
 Ion Ratio Lower Upper
 114 100
 63 21.9 0.0 45.0
 88 17.0 0.0 36.6





#35

Dibromofluoromethane

Concen: 48.252 ug/l

RT: 5.385 min Scan# 705

Delta R.T. -0.000 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:113 Resp: 55086

Ion Ratio Lower Upper

113 100

111 103.1 82.9 124.3

192 16.3 13.3 19.9



#36

1,1-Dichloropropene

Concen: 0.821 ug/l

RT: 5.702 min Scan# 757

Delta R.T. 0.012 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion: 75 Resp: 1459


Ion Ratio Lower Upper

75 100

110 27.9 17.0 50.9

77 20.6 24.7 37.1#





#38

Carbon Tetrachloride

Concen: 0.240 ug/l

RT: 5.672 min Scan# 753

Delta R.T. -0.006 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:117 Resp: 390

Ion Ratio Lower Upper

117 100

119 78.8 78.2 117.2

121 0.0 24.4 36.6#



#39

Methylcyclohexane

Concen: 0.487 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. 0.006 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion: 83 Resp: 1040


Ion Ratio Lower Upper

83 100

55 65.6 64.6 97.0

98 56.8 36.7 55.1#









#40
Benzene
Concen: 0.278 ug/l
RT: 6.050 min Scan# 81
Delta R.T. 0.012 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 78 Resp: 1398
Ion Ratio Lower Upper
78 100
77 12.9 19.3 28.9#



#42
1,2-Dichloroethane
Concen: 0.304 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion: 62 Resp: 582
Ion Ratio Lower Upper
62 100
98 0.0 0.0 16.0








#44
Trichloroethene
Concen: 0.719 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:130 Resp: 887
Ion Ratio Lower Upper
130 100
95 91.0 0.0 204.4




#46
Dibromomethane
Concen: 0.518 ug/l
RT: 7.592 min Scan# 1067
Delta R.T. 0.012 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06



Tgt Ion: 93 Resp: 483
Ion Ratio Lower Upper
93 100
95 61.7 66.6 100.0#
174 81.4 69.4 104.0





#49

1,4-Dioxane

Concen: 4.716 ug/l

RT: 7.677 min Scan# 1081

Delta R.T. 0.018 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion: 88	Resp: 156
Ion Ratio	Lower Upper
88	100
43	0.0 0.0 0.0
58	62.2 60.1 90.1



#50

Toluene-d8

Concen: 49.422 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Tgt Ion: 98	Resp: 212798
Ion Ratio	Lower Upper
98	100
100	64.5 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 0.458 ug/l

RT: 8.580 min Scan# 1229

Delta R.T. 0.006 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 1098

Ion Ratio Lower Upper

43 100

58 32.4 30.4 45.6



#52

Toluene

Concen: 0.443 ug/l

RT: 8.726 min Scan# 1253

Delta R.T. 0.012 min

Lab File: VX039700.D




Acq: 10 Jan 2024 13:06


Tgt Ion: 92 Resp: 1331

Ion Ratio Lower Upper

92 100

91 168.5 134.2 201.2









#53
 t-1,3-Dichloropropene
 Concen: 0.603 ug/l
 RT: 8.988 min Scan# 1195
 Delta R.T. 0.012 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 1104
 Ion Ratio Lower Upper
 75 100
 77 45.3 25.7 38.5#



#54
 cis-1,3-Dichloropropene
 Concen: 0.422 ug/l
 RT: 8.372 min Scan# 1195
 Delta R.T. 0.006 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06

Tgt Ion: 75 Resp: 850
 Ion Ratio Lower Upper
 75 100
 77 53.6 25.3 37.9#
 39 75.3 46.7 70.1#





#58

2-Chloroethyl Vinyl ether

Concen: 0.947 ug/l

RT: 8.256 min Scan# 1176

Delta R.T. 0.018 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 927

Ion Ratio Lower Upper

63 100

106 10.6 22.5 33.7#



#59

2-Hexanone

Concen: 0.929 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.012 min

Lab File: VX039700.D




Acq: 10 Jan 2024 13:06


Tgt Ion: 43 Resp: 1761

Ion Ratio Lower Upper

43 100


58 38.9 26.2 78.5








#61
1,2-Dibromoethane
Concen: 0.368 ug/l
RT: 9.604 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:107 Resp: 488
Ion Ratio Lower Upper
107 100
109 118.4 75.2 112.8#



#62
4-Bromofluorobenzene
Concen: 54.800 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion: 95 Resp: 92269
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.6
176 64.7 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 174291
Ion Ratio Lower Upper
117 100
82 58.4 46.2 69.4
119 32.2 25.5 38.3



#64
Tetrachloroethene
Concen: 0.756 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion:164 Resp: 857
Ion Ratio Lower Upper
164 100
166 142.5 102.7 154.1
129 100.3 80.4 120.6
131 92.7 76.5 114.7





#65

Chlorobenzene

Concen: 0.542 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 1968

Ion Ratio Lower Upper

112 100

114 40.9 25.7 38.5#



#67

Ethyl Benzene

Concen: 0.456 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. -0.000 min

Lab File: VX039700.D




Acq: 10 Jan 2024 13:06


Tgt Ion: 91 Resp: 3070

Ion Ratio Lower Upper

91 100

106 39.4 24.9 37.3#









#68
m/p-Xylenes
Concen: 0.953 ug/l
RT: 10.305 min Scan# 1511
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 2361
Ion Ratio Lower Upper
106 100
91 230.2 165.5 248.3



#69
o-Xylene
Concen: 0.391 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion:106 Resp: 960
Ion Ratio Lower Upper
106 100
91 170.7 108.9 326.8





#70

Styrene

Concen: 0.501 ug/l

RT: 10.665 min Scan# 1571

Delta R.T. 0.012 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:104	Resp:	2009
Ion	Ratio	Lower Upper
104	100	
78	49.4	43.6 65.4
103	53.8	43.6 65.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Tgt Ion:152	Resp:	83876
Ion	Ratio	Lower Upper
152	100	
115	66.3	43.3 129.9
150	156.5	0.0 345.0









#73
Isopropylbenzene
Concen: 0.384 ug/l
RT: 10.963 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 2566
Ion Ratio Lower Upper
105 100
120 26.9 12.9 38.6



#74
N-ethyl acetate
Concen: 0.302 ug/l
RT: 10.847 min Scan# 1601
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 43 Resp: 1051
Ion Ratio Lower Upper
43 100
70 29.5 32.9 49.3#
55 11.8 19.3 28.9#
61 19.2 18.8 28.2









#76
1,2,3-Trichloropropane
Concen: 22.951 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.159 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 49909
Ion Ratio Lower Upper
75 100
77 1.1 19.1 57.5#



#77
Bromobenzene
Concen: 0.628 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 156 Resp: 934
Ion Ratio Lower Upper
156 100
77 186.6 88.4 265.3
158 90.8 48.0 144.0









#78
n-propylbenzene
Concen: 0.651 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 5258
Ion Ratio Lower Upper
91 100
120 23.8 11.5 34.4



#79
2-Chlorotoluene
Concen: 0.572 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 91 Resp: 2855
Ion Ratio Lower Upper
91 100
126 27.7 16.3 48.8





#80

1,3,5-Trimethylbenzene

Concen: 0.457 ug/l

RT: 11.451 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 2594

Ion Ratio Lower Upper

105 100

120 43.7 24.1 72.3



#81

trans-1,4-Dichloro-2-butene

Concen: 66.292 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.061 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion: 75 Resp: 49909


Ion Ratio Lower Upper

75 100

53 0.0 90.5 135.7#

89 0.0 36.9 55.3#





#82

4-Chlorotoluene

Concen: 0.842 ug/l

RT: 11.457 min Scan# 1700

Delta R.T. 0.006 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 91 Resp: 4770

Ion Ratio Lower Upper

91 100

126 28.8 14.2 42.6



#83

tert-Butylbenzene

Concen: 0.343 ug/l

RT: 11.719 min Scan# 1744

Delta R.T. 0.006 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion: 119 Resp: 1855


Ion Ratio Lower Upper

119 100

91 75.0 31.2 93.6

134 28.8 11.7 35.0









#84
1,2,4-Trimethylbenzene
Concen: 0.547 ug/l
RT: 11.756 min Scan# 11750
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 3094
Ion Ratio Lower Upper
105 100
120 38.5 21.2 63.6



#85
sec-Butylbenzene
Concen: 0.594 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Tgt Ion:105 Resp: 4175
Ion Ratio Lower Upper
105 100
134 18.8 9.8 29.4








#86
p-Isopropyltoluene
Concen: 0.648 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:119 Resp: 3642
Ion Ratio Lower Upper
119 100
134 26.2 12.8 38.3
91 33.6 13.1 39.3




#87
1,3-Dichlorobenzene
Concen: 1.105 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Tgt Ion:146 Resp: 3149
Ion Ratio Lower Upper
146 100
111 39.0 22.1 66.1
148 57.2 31.8 95.3








#88
1,4-Dichlorobenzene
Concen: 1.070 ug/l
RT: 11.969 min Scan# 11
Delta R.T. -0.073 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 3149
Ion Ratio Lower Upper
146 100
111 39.0 21.3 63.7
148 57.2 31.8 95.4



#89
n-Butylbenzene
Concen: 1.148 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 91 Resp: 6229
Ion Ratio Lower Upper
91 100
92 54.1 27.4 82.0
134 23.6 12.6 37.8





#90

Hexachloroethane

Concen: 0.263 ug/l

RT: 12.536 min Scan# 117

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:117 Resp: 249

Ion Ratio Lower Upper

117 100

201 64.3 32.9 98.6



#91

1,2-Dichlorobenzene

Concen: 0.767 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06




Tgt Ion:146 Resp: 2160


Ion Ratio Lower Upper

146 100

111 49.0 22.4 67.0

148 61.3 31.6 95.0









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.435 ug/l
 RT: 12.951 min Scan# 1946
 Delta R.T. 0.006 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 272
 Ion Ratio Lower Upper
 75 100
 155 34.2 36.1 108.5#
 157 53.7 44.9 134.7



#93
 1,2,4-Trichlorobenzene
 Concen: 2.152 ug/l
 RT: 13.591 min Scan# 2051
 Delta R.T. 0.006 min
 Lab File: VX039700.D
 Acq: 10 Jan 2024 13:06


Tgt Ion: 180 Resp: 3706
 Ion Ratio Lower Upper
 180 100
 182 92.4 48.3 144.8
 145 32.5 18.0 54.0








#94
Hexachlorobutadiene
Concen: 2.172 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:225 Resp: 1555
Ion Ratio Lower Upper
225 100
223 59.7 31.1 93.2
227 61.4 32.5 97.5



#95
Naphthalene
Concen: 2.242 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion:128 Resp: 14210
Ion Ratio Lower Upper
128 100
127 12.7 10.3 15.5
129 10.1 8.8 13.2





#96

1,2,3-Trichlorobenzene

Concen: 2.039 ug/l

RT: 13.963 min Scan# 21

Delta R.T. -0.000 min


Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:180 Resp: 3544

Ion Ratio Lower Upper

180 100

182 95.0 46.9 140.6

145 36.3 17.8 53.5

