

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039708.D
 Acq On : 10 Jan 2024 16:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 11 04:01:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.647	0.557	13.9	80	0.00
3 P	Chloromethane	0.739	0.680	8.0	89	0.00
4 C	Vinyl Chloride	0.697	0.653	6.3#	91	0.00
5 T	Bromomethane	0.398	0.383	3.8	93	0.00
6 T	Chloroethane	0.396	0.390	1.5	90	0.00
7 T	Trichlorofluoromethane	1.006	0.942	6.4	92	0.00
8 T	Diethyl Ether	0.372	0.361	3.0	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.624	0.565	9.5	88	0.00
10 T	Methyl Iodide	0.658	0.649	1.4	89	0.00
11 T	Tert butyl alcohol	0.218	0.201	7.8	89	0.00
12 CM	1,1-Dichloroethene	0.602	0.570	5.3#	91	0.00
13 T	Acrolein	0.147	0.138	6.1	97	0.00
14 T	Allyl chloride	1.184	1.151	2.8	93	0.00
15 T	Acrylonitrile	0.450	0.449	0.2	94	0.00
16 T	Acetone	0.471	0.491	-4.2	98	0.00
17 T	Carbon Disulfide	1.730	1.626	6.0	92	0.00
18 T	Methyl Acetate	0.985	0.974	1.1	95	0.00
19 T	Methyl tert-butyl Ether	2.136	2.104	1.5	96	0.00
20 T	Methylene Chloride	0.713	0.688	3.5	96	0.00
21 T	trans-1,2-Dichloroethene	0.633	0.620	2.1	95	0.00
22 T	Diisopropyl ether	2.260	2.235	1.1	96	0.00
23 T	Vinyl Acetate	2.101	2.183	-3.9	97	0.00
24 P	1,1-Dichloroethane	1.243	1.231	1.0	95	0.00
25 T	2-Butanone	0.677	0.681	-0.6	95	0.00
26 T	2,2-Dichloropropane	0.998	0.958	4.0	92	0.00
27 T	cis-1,2-Dichloroethene	0.726	0.737	-1.5	96	0.00
28 T	Bromochloromethane	0.564	0.580	-2.8	98	0.00
29 T	Tetrahydrofuran	0.436	0.427	2.1	93	0.00
30 C	Chloroform	1.228	1.228	0.0	97	0.00
31 T	Cyclohexane	1.105	1.023	7.4	89	0.00
32 T	1,1,1-Trichloroethane	1.051	1.028	2.2	93	0.00
33 S	1,2-Dichloroethane-d4	0.838	0.832	0.7	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.337	0.328	2.7	100	0.00
36 T	1,1-Dichloropropene	0.524	0.479	8.6	95	0.00
37 T	Ethyl Acetate	0.699	0.656	6.2	98	0.00
38 T	Carbon Tetrachloride	0.480	0.446	7.1	92	0.00
39 T	Methylcyclohexane	0.629	0.564	10.3	90	0.00
40 TM	Benzene	1.484	1.409	5.1	96	0.00
41 T	Methacrylonitrile	0.342	0.347	-1.5	95	0.00
42 TM	1,2-Dichloroethane	0.564	0.552	2.1	98	0.00
43 T	Isopropyl Acetate	1.026	1.016	1.0	96	0.00
44 TM	Trichloroethene	0.364	0.341	6.3	94	0.00
45 C	1,2-Dichloropropane	0.398	0.380	4.5#	97	0.00
46 T	Dibromomethane	0.275	0.267	2.9	97	0.00
47 T	Bromodichloromethane	0.509	0.506	0.6	98	0.00
48 T	Methyl methacrylate	0.504	0.499	1.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039708.D
 Acq On : 10 Jan 2024 16:54
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 11 04:01:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	92	0.00
50 S	Toluene-d8	1.270	1.231	3.1	100	0.00
51 T	4-Methyl-2-Pentanone	0.707	0.678	4.1	94	0.00
52 CM	Toluene	0.887	0.858	3.3#	95	0.00
53 T	t-1,3-Dichloropropene	0.540	0.557	-3.1	98	0.00
54 T	cis-1,3-Dichloropropene	0.594	0.599	-0.8	97	0.00
55 T	1,1,2-Trichloroethane	0.377	0.366	2.9	97	0.00
56 T	Ethyl methacrylate	0.611	0.618	-1.1	98	0.00
57 T	1,3-Dichloropropane	0.640	0.627	2.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.289	0.291	-0.7	97	0.00
59 T	2-Hexanone	0.559	0.552	1.3	94	0.00
60 T	Dibromochloromethane	0.356	0.362	-1.7	98	0.00
61 T	1,2-Dibromoethane	0.391	0.389	0.5	97	0.00
62 S	4-Bromofluorobenzene	0.497	0.500	-0.6	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.325	0.299	8.0	98	0.00
65 PM	Chlorobenzene	1.041	0.986	5.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.355	0.350	1.4	96	0.00
67 C	Ethyl Benzene	1.933	1.835	5.1#	95	0.00
68 T	m/p-Xylenes	0.711	0.683	3.9	96	0.00
69 T	o-Xylene	0.704	0.672	4.5	97	0.00
70 T	Styrene	1.150	1.130	1.7	98	0.00
71 P	Bromoform	0.266	0.274	-3.0	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
73 T	Isopropylbenzene	3.981	3.578	10.1	95	0.00
74 T	N-amyl acetate	2.075	2.040	1.7	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.520	1.390	8.6	97	0.00
76 T	1,2,3-Trichloropropane	1.296	1.184	8.6	97	0.00
77 T	Bromobenzene	0.886	0.808	8.8	98	0.00
78 T	n-propylbenzene	4.818	4.438	7.9	95	0.00
79 T	2-Chlorotoluene	2.978	2.640	11.3	96	0.00
80 T	1,3,5-Trimethylbenzene	3.381	3.113	7.9	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.427	0.410	4.0	96	0.00
82 T	4-Chlorotoluene	3.378	3.057	9.5	97	0.00
83 T	tert-Butylbenzene	3.228	2.940	8.9	95	0.00
84 T	1,2,4-Trimethylbenzene	3.373	3.123	7.4	96	0.00
85 T	sec-Butylbenzene	4.190	3.837	8.4	95	0.00
86 T	p-Isopropyltoluene	3.349	3.130	6.5	95	0.00
87 T	1,3-Dichlorobenzene	1.699	1.553	8.6	98	0.00
88 T	1,4-Dichlorobenzene	1.755	1.576	10.2	97	0.00
89 T	n-Butylbenzene	3.234	3.020	6.6	94	0.00
90 T	Hexachloroethane	0.565	0.547	3.2	96	0.00
91 T	1,2-Dichlorobenzene	1.680	1.522	9.4	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.372	0.350	5.9	95	0.00
93 T	1,2,4-Trichlorobenzene	1.026	0.957	6.7	100	0.00
94 T	Hexachlorobutadiene	0.427	0.380	11.0	93	0.00
95 T	Naphthalene	3.779	3.500	7.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
Data File : VX039708.D
Acq On : 10 Jan 2024 16:54
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 11 04:01:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.036	0.958	7.5	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5