

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039700.D
 Acq On : 10 Jan 2024 13:06
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 11 03:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	85700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	169496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	174291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83876	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70317	48.981	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.960%
35) Dibromofluoromethane	5.385	113	55086	48.252	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.647	98	212798	49.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.840%
62) 4-Bromofluorobenzene	11.079	95	92269	54.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.600%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
Data File : VX039700.D
Acq On : 10 Jan 2024 13:06
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 11 03:56:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 85700
Ion Ratio Lower Upper
168 100
99 68.1 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 48.981 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 65 Resp: 70317
Ion Ratio Lower Upper
65 100
67 51.8 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:114 Resp: 169496
Ion Ratio Lower Upper
114 100
63 21.9 0.0 45.0
88 17.0 0.0 36.6



#35
Dibromofluoromethane
Concen: 48.252 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion:113 Resp: 55086
Ion Ratio Lower Upper
113 100
111 103.1 82.9 124.3
192 16.3 13.3 19.9









#50
Toluene-d8
Concen: 49.422 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 212798
Ion Ratio Lower Upper
98 100
100 64.5 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 54.800 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX039700.D
Acq: 10 Jan 2024 13:06

Tgt Ion: 95 Resp: 92269
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.6
176 64.7 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Instrument :

MSVOA_X

ClientSampleId :

IBLK





Tgt Ion:117 Resp: 174291

Ion Ratio Lower Upper

117 100

82 58.4 46.2 69.4

119 32.2 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX039700.D

Acq: 10 Jan 2024 13:06

Tgt Ion:152 Resp: 83876

Ion Ratio Lower Upper

152 100

115 66.3 43.3 129.9

150 156.5 0.0 345.0

