

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039702.D
 Acq On : 10 Jan 2024 14:10
 Operator : JC/MD
 Sample : VX0110MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0110MBL01

Quant Time: Jan 11 03:58:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	89442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	175844	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	179389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86801	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70761	47.228	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.460%
35) Dibromofluoromethane	5.379	113	55910	47.206	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	8.647	98	217474	48.685	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.360%
62) 4-Bromofluorobenzene	11.079	95	96632	55.319	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
Data File : VX039702.D
Acq On : 10 Jan 2024 14:10
Operator : JC/MD
Sample : VX0110MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0110MBL01

Quant Time: Jan 11 03:58:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10




Instrument :
MSVOA_X
ClientSampleId :
VX0110MBL01


Tgt Ion:168 Resp: 89442
Ion Ratio Lower Upper
168 100
99 67.0 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 47.228 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10

Tgt Ion: 65 Resp: 70761
Ion Ratio Lower Upper
65 100
67 53.1 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10




Instrument :
MSVOA_X
ClientSampleId :
VX0110MBL01

Tgt Ion:114 Resp: 175844
Ion Ratio Lower Upper
114 100
63 22.5 0.0 45.0
88 16.2 0.0 36.6



#35
Dibromofluoromethane
Concen: 47.206 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10

Tgt Ion:113 Resp: 55910
Ion Ratio Lower Upper
113 100
111 106.0 82.9 124.3
192 16.7 13.3 19.9









#50
Toluene-d8
Concen: 48.685 ug/l
RT: 8.647 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10




Instrument :
MSVOA_X
ClientSampleId :
VX0110MBL01


Tgt Ion: 98 Resp: 217474
Ion Ratio Lower Upper
98 100
100 64.5 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 55.319 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX039702.D
Acq: 10 Jan 2024 14:10

Tgt Ion: 95 Resp: 96632
Ion Ratio Lower Upper
95 100
174 66.5 0.0 135.6
176 64.1 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1471

Delta R.T. -0.006 min

Lab File: VX039702.D

Acq: 10 Jan 2024 14:10

Instrument :

MSVOA_X

ClientSampleId :

VX0110MBL01





Tgt Ion:117 Resp: 179389

Ion Ratio Lower Upper

117 100

82 61.6 46.2 69.4

119 31.3 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX039702.D

Acq: 10 Jan 2024 14:10

Tgt Ion:152 Resp: 86801

Ion Ratio Lower Upper

152 100

115 64.3 43.3 129.9

150 155.4 0.0 345.0

