

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
 Data File : VX039707.D
 Acq On : 10 Jan 2024 16:31
 Operator : JC/MD
 Sample : P1084-03 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWS-PROJECT

Quant Time: Jan 11 04:01:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	162930	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171113	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84495	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64432	47.426	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.860%
35) Dibromofluoromethane	5.385	113	52274	47.634	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.260%
50) Toluene-d8	8.647	98	205331	49.610	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.220%
62) 4-Bromofluorobenzene	11.079	95	89298	55.173	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.340%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011024\
Data File : VX039707.D
Acq On : 10 Jan 2024 16:31
Operator : JC/MD
Sample : P1084-03 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWS-PROJECT

Quant Time: Jan 11 04:01:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31




Instrument :
MSVOA_X
ClientSampleId :
OWS-PROJECT


Tgt Ion:168 Resp: 81102
Ion Ratio Lower Upper
168 100
99 66.1 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 47.426 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31

Tgt Ion: 65 Resp: 64432
Ion Ratio Lower Upper
65 100
67 53.8 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. 0.006 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31




Instrument :
MSVOA_X
ClientSampleId :
OWS-PROJECT


Tgt Ion:114 Resp: 162930
Ion Ratio Lower Upper
114 100
63 21.6 0.0 45.0
88 16.0 0.0 36.6



#35
Dibromofluoromethane
Concen: 47.634 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31

Tgt Ion:113 Resp: 52274
Ion Ratio Lower Upper
113 100
111 102.8 82.9 124.3
192 16.3 13.3 19.9









#50
Toluene-d8
Concen: 49.610 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31




Instrument :
MSVOA_X
ClientSampleId :
OWS-PROJECT


Tgt Ion: 98 Resp: 205331
Ion Ratio Lower Upper
98 100
100 64.6 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 55.173 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX039707.D
Acq: 10 Jan 2024 16:31

Tgt Ion: 95 Resp: 89298
Ion Ratio Lower Upper
95 100
174 67.2 0.0 135.6
176 65.6 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX039707.D

Acq: 10 Jan 2024 16:31

Instrument :

MSVOA_X

ClientSampleId :

OVS-PROJECT





Tgt Ion:117 Resp: 171113

Ion Ratio Lower Upper

117 100

82 59.4 46.2 69.4

119 32.6 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VX039707.D

Acq: 10 Jan 2024 16:31

Tgt Ion:152 Resp: 84495

Ion Ratio Lower Upper

152 100

115 64.5 43.3 129.9

150 154.7 0.0 345.0

