

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\ VX011119\
 Data File : VX007010.D
 Acq On : 11 Jan 2019 13:44
 Operator : JC/SP
 Sample : VX0111MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01

Quant Time: Jan 11 14:03:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	534169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	797907	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	729350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	380327	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	265851	47.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.24%	
35) Dibromofluoromethane	5.50	113	238841	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	
50) Toluene-d8	8.71	98	911474	48.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.13	95	333657	51.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.34%	

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.20	85	98110	22.366	ug/l 99
3) Chloromethane	1.33	50	109848	21.511	ug/l 100
4) Vinyl Chloride	1.41	62	114666	20.874	ug/l 100
5) Bromomethane	1.64	94	110143	18.767	ug/l 98
6) Chloroethane	1.72	64	73249	19.097	ug/l 95
7) Trichlorofluoromethane	1.93	101	175049	19.722	ug/l 100
8) Diethyl Ether	2.19	74	70042	19.836	ug/l 100
9) 1,1,2-Trichlorotrifluoroet	2.39	101	102950	19.374	ug/l 99
10) Methyl Iodide	2.51	142	152336	21.253	ug/l 99
11) Tert butyl alcohol	3.07	59	120482	91.490	ug/l 100
12) 1,1-Dichloroethene	2.37	96	94687	19.287	ug/l 98
13) Acrolein	2.29	56	41520	75.287	ug/l 99
14) Allyl chloride	2.73	41	163750	19.462	ug/l 98
15) Acrylonitrile	3.15	53	304004	100.561	ug/l 99
16) Acetone	2.45	43	260353	94.337	ug/l 100
17) Carbon Disulfide	2.57	76	199495	15.783	ug/l 98
18) Methyl Acetate	2.78	43	162099	20.282	ug/l 99
19) Methyl tert-butyl Ether	3.20	73	344944	20.243	ug/l 96
20) Methylene Chloride	2.86	84	111656	20.942	ug/l 93
21) trans-1,2-Dichloroethene	3.17	96	102665	18.765	ug/l 98
22) Diisopropyl ether	3.87	45	312176	19.736	ug/l 92
23) Vinyl Acetate	3.82	43	1262480	94.642	ug/l 100
24) 1,1-Dichloroethane	3.70	63	173362	19.600	ug/l 98
25) 2-Butanone	4.70	43	383108	97.016	ug/l 99
26) 2,2-Dichloropropane	4.59	77	120169	17.668	ug/l 100
27) cis-1,2-Dichloroethene	4.60	96	113624	18.829	ug/l 97
28) Bromochloromethane	5.02	49	71688	18.499	ug/l 100
29) Tetrahydrofuran	5.15	42	249860	97.362	ug/l 98
30) Chloroform	5.21	83	183183	19.781	ug/l 95
31) Cyclohexane	5.57	56	154445	19.619	ug/l 96
32) 1,1,1-Trichloroethane	5.49	97	149195	19.008	ug/l 98
36) 1,1-Dichloropropene	5.80	75	132295	19.071	ug/l 99
37) Ethyl Acetate	4.85	43	143490	19.293	ug/l 99
38) Carbon Tetrachloride	5.78	117	123939	17.757	ug/l 98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\ VX011119\
 Data File : VX007010.D
 Acq On : 11 Jan 2019 13:44
 Operator : JC/SP
 Sample : VX0111MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01

Quant Time: Jan 11 14:03:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	161213	18.229	ug/l	97
40) Benzene	6.14	78	414710	19.592	ug/l	99
41) Methacrylonitrile	5.06	41	82065	19.154	ug/l	97
42) 1,2-Dichloroethane	6.20	62	139683	19.096	ug/l	99
43) Isopropyl Acetate	6.46	43	220779	18.820	ug/l	99
44) Trichloroethene	7.21	130	120836	19.050	ug/l	96
45) 1,2-Dichloropropane	7.51	63	101274	19.031	ug/l	95
46) Dibromomethane	7.66	93	72452	19.372	ug/l	98
47) Bromodichloromethane	7.90	83	113176	17.778	ug/l	95
48) Methyl methacrylate	7.77	41	113168	18.785	ug/l	98
49) 1,4-Dioxane	7.75	88	48775	348.935	ug/l	97
51) 4-Methyl-2-Pentanone	8.65	43	762279	101.109	ug/l	99
52) Toluene	8.78	92	267461	19.659	ug/l	100
53) t-1,3-Dichloropropene	9.04	75	124307	17.610	ug/l	100
54) cis-1,3-Dichloropropene	8.44	75	141828	18.193	ug/l	95
55) 1,1,2-Trichloroethane	9.21	97	107797	19.795	ug/l	97
56) Ethyl methacrylate	9.18	69	158405	19.926	ug/l	95
57) 1,3-Dichloropropane	9.37	76	177300	19.874	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.31	63	375209	91.795	ug/l	100
59) 2-Hexanone	9.49	43	577916	100.589	ug/l	99
60) Dibromochloromethane	9.59	129	96766	18.300	ug/l	96
61) 1,2-Dibromoethane	9.67	107	117130	19.538	ug/l	100
64) Tetrachloroethene	9.34	164	128441	20.341	ug/l	96
65) Chlorobenzene	10.14	112	309011	19.425	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.22	131	100194	18.723	ug/l	99
67) Ethyl Benzene	10.25	91	513532	19.555	ug/l	97
68) m/p-Xylenes	10.36	106	407605	39.395	ug/l	100
69) o-Xylene	10.69	106	200811	19.901	ug/l	99
70) Styrene	10.71	104	319273	19.978	ug/l	98
71) Bromoform	10.86	173	67464	17.486	ug/l	99
73) Isopropylbenzene	11.02	105	523036	19.503	ug/l	99
74) N-amyl acetate	10.90	43	205712	19.052	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.27	83	161564	18.976	ug/l	99
76) 1,2,3-Trichloropropane	11.29	75	174684	22.665	ug/l	94
77) Bromobenzene	11.26	156	145825	19.734	ug/l	97
78) n-propylbenzene	11.36	91	581378	19.747	ug/l	99
79) 2-Chlorotoluene	11.42	91	343579	19.387	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	435155	19.419	ug/l	100
81) trans-1,4-Dichloro-2-buten	11.07	75	34479	16.262	ug/l	87
82) 4-Chlorotoluene	11.51	91	401599	19.572	ug/l	100
83) tert-Butylbenzene	11.77	119	422280	18.887	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	445927	19.789	ug/l	100
85) sec-Butylbenzene	11.94	105	524652	19.649	ug/l	100
86) p-Isopropyltoluene	12.06	119	466931	19.550	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	260461	19.501	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	256909	20.387	ug/l	98
89) n-Butylbenzene	12.39	91	390120	19.161	ug/l	99
90) Hexachloroethane	12.60	117	57016	16.992	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	257083	19.504	ug/l	100
92) 1,2-Dibromo-3-Chloropropan	13.00	75	28836	17.125	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011119\
Data File : VX007010.D
Acq On : 11 Jan 2019 13:44
Operator : JC/SP
Sample : VX0111MBS01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01

Quant Time: Jan 11 14:03:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	175053	19.358	ug/l	99
94) Hexachlorobutadiene	13.78	225	77717	19.367	ug/l	99
95) Naphthalene	13.83	128	532610	19.437	ug/l	99
96) 1,2,3-Trichlorobenzene	14.02	180	176759	19.677	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011119\
Data File : VX007010.D
Acq On : 11 Jan 2019 13:44
Operator : JC/SP
Sample : VX0111MBS01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
Client Sample Id :
VX0111MBS01

Quant Time: Jan 11 14:03:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion:168 Resp: 534169

Ion Ratio Lower Upper




168 100

99 47.1 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 22.366 ug/l

RT: 1.20 min Scan# 7

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 85 Resp: 98110

Ion Ratio Lower Upper



85 100


87 32.3 15.9 47.6



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0









#3
Chloromethane
Concen: 21.511 ug/l
RT: 1.33 min Scan# 29
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 50 Resp: 109848
Ion Ratio Lower Upper
50 100
52 32.0 25.4 38.2



#4
Vinyl Chloride
Concen: 20.874 ug/l
RT: 1.41 min Scan# 42
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 62 Resp: 114666
Ion Ratio Lower Upper
62 100
64 31.5 25.1 37.7





#5

Bromomethane

Concen: 18.767 ug/l

RT: 1.64 min Scan# 80

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 94 Resp: 110143

Ion Ratio Lower Upper




94 100

96 94.0 73.4 110.2



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 19.097 ug/l

RT: 1.72 min Scan# 93

Delta R.T. 0.01 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 64 Resp: 73249

Ion Ratio Lower Upper



64 100


66 30.3 26.4 39.6



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#7

Trichlorofluoromethane
Concen: 19.722 ug/l
RT: 1.93 min Scan# 127
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01




Tgt Ion:101 Resp: 175049
Ion Ratio Lower Upper
101 100
103 66.4 53.2 79.8




#8

Diethyl Ether
Concen: 19.836 ug/l
RT: 2.19 min Scan# 170
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 74 Resp: 70042
Ion Ratio Lower Upper
74 100
45 92.6 46.3 138.9









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 19.374 ug/l
 RT: 2.39 min Scan# 202
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01


Tgt Ion	Ratio	Lower	Upper
101	100		
85	43.2	35.1	52.7
151	86.8	68.9	103.3



#10
 Methyl Iodide
 Concen: 21.253 ug/l
 RT: 2.51 min Scan# 223
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion	Ratio	Lower	Upper
142	100		
127	43.4	33.9	50.9
141	14.6	11.4	17.0









#11

Tert butyl alcohol
 Concen: 91.490 ug/l
 RT: 3.07 min Scan# 314
 Delta R.T. -0.01 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01




Tgt Ion: 59 Resp: 120482
 Ion Ratio Lower Upper
 59 100
 57 10.6 8.5 12.7




#12

1,1-Dichloroethene
 Concen: 19.287 ug/l
 RT: 2.37 min Scan# 200
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion: 96 Resp: 94687
 Ion Ratio Lower Upper
 96 100
 61 155.6 121.4 182.2
 98 64.6 51.9 77.9









#13
Acrolein
Concen: 75.287 ug/l
RT: 2.29 min Scan# 187
Delta R.T. -0.01 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 56 Resp: 41520
Ion Ratio Lower Upper
56 100
55 72.3 58.2 87.4



#14
Allyl chloride
Concen: 19.462 ug/l
RT: 2.73 min Scan# 258
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 41 Resp: 163750
Ion Ratio Lower Upper
41 100
39 68.1 53.6 80.4
76 32.8 27.4 41.2





#15

Acrylonitrile

Concen: 100.561 ug/l

RT: 3.15 min Scan# 327

Delta R.T. -0.01 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01





Tgt Ion: 53 Resp: 304004

Ion Ratio Lower Upper

53 100

52 81.6 66.0 99.0

51 35.4 28.6 42.8



#16

Acetone

Concen: 94.337 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX007010.D




Acq: 11 Jan 2019 13:44


Tgt Ion: 43 Resp: 260353

Ion Ratio Lower Upper

43 100

58 31.3 25.0 37.6





#17

Carbon Disulfide

Concen: 15.783 ug/l

RT: 2.57 min Scan# 232

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 76 Resp: 199495

Ion Ratio Lower Upper




76 100

78 9.7 7.0 10.6



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#18

Methyl Acetate

Concen: 20.282 ug/l

RT: 2.78 min Scan# 266

Delta R.T. -0.01 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 43 Resp: 162099

Ion Ratio Lower Upper



43 100


74 23.9 18.6 28.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#19
Methyl tert-butyl Ether
Concen: 20.243 ug/l
RT: 3.20 min Scan# 336
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 73 Resp: 344944
Ion Ratio Lower Upper
73 100
57 22.8 16.8 25.2



#20
Methylene Chloride
Concen: 20.942 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 84 Resp: 111656
Ion Ratio Lower Upper
84 100
49 120.8 91.8 137.6
51 39.6 28.5 42.7
86 69.5 50.2 75.2





#21

trans-1,2-Dichloroethene

Concen: 18.765 ug/l

RT: 3.17 min Scan# 330

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion: 96 Resp: 102665

Ion Ratio Lower Upper

96 100

61 132.7 103.7 155.5




98 63.8 51.0 76.6



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 19.736 ug/l

RT: 3.87 min Scan# 446

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Tgt Ion: 45 Resp: 312176


Ion Ratio Lower Upper

45 100

43 56.7 51.2 76.8

87 28.3 25.8 38.6

59 11.0 9.0 13.4






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 94.642 ug/l

RT: 3.82 min Scan# 438

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 43 Resp: 1262480

Ion Ratio Lower Upper




43 100

86 11.4 9.2 13.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 86.00 (85.70 to 86.70): VX0



#24

1,1-Dichloroethane

Concen: 19.600 ug/l

RT: 3.70 min Scan# 417

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 63 Resp: 173362

Ion Ratio Lower Upper

63 100

98 7.0 3.9 11.5



100 5.1 2.4 7.2




Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 97.90 (97.60 to 98.60): VX0

Ion 99.90 (99.60 to 100.60): VX0





#25

2-Butanone

Concen: 97.016 ug/l

RT: 4.70 min Scan# 581

Delta R.T. -0.01 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :





VX0111MBS01

Tgt Ion: 43 Resp: 383108

Ion Ratio Lower Upper

43 100

72 27.5 22.4 33.6



#26

2,2-Dichloropropane

Concen: 17.668 ug/l

RT: 4.59 min Scan# 563

Delta R.T. -0.00 min

Lab File: VX007010.D




Acq: 11 Jan 2019 13:44


Tgt Ion: 77 Resp: 120169

Ion Ratio Lower Upper

77 100

97 25.1 12.4 37.4






#27




cis-1,2-Dichloroethene
 Concen: 18.829 ug/l
 RT: 4.60 min Scan# 565
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01

Tgt Ion: 96 Resp: 113624
 Ion Ratio Lower Upper
 96 100
 61 132.9 0.0 258.6
 98 67.3 0.0 132.0




Abundance Ion 95.90 (95.60 to 96.60): VX0
 Ion 60.90 (60.60 to 61.60): VX0
 Ion 97.90 (97.60 to 98.60): VX0





#28


Bromochloromethane
 Concen: 18.499 ug/l
 RT: 5.02 min Scan# 634
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion: 49 Resp: 71688
 Ion Ratio Lower Upper
 49 100
 129 3.0 0.0 5.4
 130 96.9 77.7 116.5



Abundance Ion 48.90 (48.60 to 49.60): VX0
 Ion 128.80 (128.50 to 129.50): V
 Ion 129.80 (129.50 to 130.50): V





#29

Tetrahydrofuran

Concen: 97.362 ug/l

RT: 5.15 min Scan# 656

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampled :

VX0111MBS01





Tgt Ion: 42 Resp: 249860

Ion Ratio Lower Upper

42 100

72 47.8 38.9 58.3

71 43.7 36.2 54.4



#30

Chloroform

Concen: 19.781 ug/l

RT: 5.21 min Scan# 666

Delta R.T. -0.00 min

Lab File: VX007010.D




Acq: 11 Jan 2019 13:44


Tgt Ion: 83 Resp: 183183

Ion Ratio Lower Upper

83 100

85 60.1 51.2 76.8









#31
Cyclohexane
Concen: 19.619 ug/l
RT: 5.57 min Scan# 725
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 56 Resp: 154445
Ion Ratio Lower Upper
56 100
69 33.2 24.4 36.6
84 90.7 70.0 105.0



#32
1,1,1-Trichloroethane
Concen: 19.008 ug/l
RT: 5.49 min Scan# 712
Delta R.T. -0.01 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 97 Resp: 149195
Ion Ratio Lower Upper
97 100
99 64.0 51.7 77.5
61 41.1 31.3 46.9





#33

1,2-Dichloroethane-d4

Concen: 47.622 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 65 Resp: 265851

Ion Ratio Lower Upper




65 100

67 54.5 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 114 Resp: 797907

Ion Ratio Lower Upper

114 100

63 17.8 0.0 34.8



88 14.4 0.0 28.8




Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0









#35
 Dibromofluoromethane
 Concen: 46.742 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01


Tgt Ion	Ratio	Lower	Upper
113	100		
111	104.0	81.6	122.4
192	22.2	16.7	25.1



#36
 1,1-Dichloropropene
 Concen: 19.071 ug/l
 RT: 5.80 min Scan# 762
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion	Ratio	Lower	Upper
75	100		
110	39.2	20.2	60.5
77	30.8	24.8	37.2









#37
Ethyl Acetate
Concen: 19.293 ug/l
RT: 4.85 min Scan# 607
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 43 Resp: 143490
Ion Ratio Lower Upper
43 100
61 14.4 11.9 17.9
70 11.5 9.7 14.5



#38
Carbon Tetrachloride
Concen: 17.757 ug/l
RT: 5.78 min Scan# 759
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 117 Resp: 123939
Ion Ratio Lower Upper
117 100
119 97.5 79.4 119.2
121 30.8 24.0 36.0









#39
Methylcyclohexane
Concen: 18.229 ug/l
RT: 7.46 min Scan# 1034
Delta R.T. -0.01 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01

Tgt Ion: 83 Resp: 161213
Ion Ratio Lower Upper
83 100
55 70.7 54.9 82.3
98 50.5 37.9 56.9




Abundance Ion 83.00 (82.70 to 83.70): VX0
Ion 55.00 (54.70 to 55.70): VX0
Ion 98.00 (97.70 to 98.70): VX0






#40
Benzene
Concen: 19.592 ug/l
RT: 6.14 min Scan# 818
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 78 Resp: 414710
Ion Ratio Lower Upper
78 100
77 24.7 19.2 28.8



Abundance Ion 78.00 (77.70 to 78.70): VX0
Ion 77.00 (76.70 to 77.70): VX0





#41

Methacrylonitrile

Concen: 19.154 ug/l

RT: 5.06 min Scan# 640

Delta R.T. -0.01 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01

Tgt Ion: 41 Resp: 82065


Ion Ratio Lower Upper

41 100

39 51.6 42.8 64.2

67 73.9 62.1 93.1

52 29.5 23.4 35.0






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0



#42

1,2-Dichloroethane

Concen: 19.096 ug/l

RT: 6.20 min Scan# 827

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 62 Resp: 139683

Ion Ratio Lower Upper



62 100


98 10.3 0.0 20.2



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 97.90 (97.60 to 98.60): VX0









#43

Isopropyl Acetate
 Concen: 18.820 ug/l
 RT: 6.46 min Scan# 870
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01




Tgt Ion	Ratio	Lower	Upper
43	100		
61	20.7	16.8	25.2
87	14.6	12.4	18.6




#44

Trichloroethene
 Concen: 19.050 ug/l
 RT: 7.21 min Scan# 994
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion	Ratio	Lower	Upper
130	100		
95	89.7	0.0	187.0





#45

1,2-Dichloropropane

Concen: 19.031 ug/l

RT: 7.51 min Scan# 1043

Delta R.T. -0.01 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 63 Resp: 101274

Ion Ratio Lower Upper




63 100

65 29.1 25.4 38.0



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 64.90 (64.60 to 65.60): VX0



#46

Dibromomethane

Concen: 19.372 ug/l

RT: 7.66 min Scan# 1068

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 93 Resp: 72452

Ion Ratio Lower Upper

93 100

95 83.3 66.2 99.4



174 117.5 91.0 136.4




Abundance Ion 92.80 (92.50 to 93.50): VX0

Ion 94.80 (94.50 to 95.50): VX0

Ion 173.70 (173.40 to 174.40): V





#47

Bromodichloromethane

Concen: 17.778 ug/l

RT: 7.90 min Scan# 1106

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion: 83 Resp: 113176

Ion Ratio Lower Upper

83 100

85 62.8 53.2 79.8




127 8.4 7.8 11.6



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0

Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 18.785 ug/l

RT: 7.77 min Scan# 1086

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 41 Resp: 113168

Ion Ratio Lower Upper

41 100

69 89.7 72.4 108.6



39 60.3 45.8 68.8




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 348.935 ug/l

RT: 7.75 min Scan# 1082

Delta R.T. 0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampled :

VX0111MBS01


Tgt Ion: 88 Resp: 48775

Ion Ratio Lower Upper




88 100

43 30.6 22.2 33.4

58 65.2 51.0 76.4



Abundance Ion 88.00 (87.70 to 88.70): VX0
Ion 43.00 (42.70 to 43.70): VX0
Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 48.733 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 98 Resp: 911474



Ion Ratio Lower Upper


98 100

100 63.9 52.0 78.0



Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): V





#51

4-Methyl-2-Pentanone

Concen: 101.109 ug/l

RT: 8.65 min Scan# 1229

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 43 Resp: 762279

Ion Ratio Lower Upper




43 100

58 39.2 31.8 47.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 19.659 ug/l

RT: 8.78 min Scan# 1251

Delta R.T. -0.01 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 92 Resp: 267461

Ion Ratio Lower Upper



92 100


91 172.2 137.6 206.4



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53
 t-1,3-Dichloropropene
 Concen: 17.610 ug/l
 RT: 9.04 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01


Tgt Ion: 75 Resp: 124307
 Ion Ratio Lower Upper
 75 100
 77 31.6 25.4 38.0



#54
 cis-1,3-Dichloropropene
 Concen: 18.193 ug/l
 RT: 8.44 min Scan# 1195
 Delta R.T. 0.01 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion: 75 Resp: 141828
 Ion Ratio Lower Upper
 75 100
 77 29.6 26.2 39.2
 39 42.4 36.7 55.1









#55
 1,1,2-Trichloroethane
 Concen: 19.795 ug/l
 RT: 9.21 min Scan# 1322
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampled :
 VX0111MBS01


Tgt Ion:	97	Resp:	107797
Ion	Ratio	Lower	Upper
97	100		
83	86.3	66.6	99.8
85	57.0	42.7	64.1
99	64.7	50.8	76.2



#56
 Ethyl methacrylate
 Concen: 19.926 ug/l
 RT: 9.18 min Scan# 1316
 Delta R.T. -0.01 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion:	69	Resp:	158405
Ion	Ratio	Lower	Upper
69	100		
41	65.2	48.7	73.1
39	36.5	27.0	40.6





#57

1,3-Dichloropropane

Concen: 19.874 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 76 Resp: 177300

Ion Ratio Lower Upper




76 100

78 32.0 26.0 39.0



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#58

2-Chloroethyl Vinyl ether

Concen: 91.795 ug/l

RT: 8.31 min Scan# 1174

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 63 Resp: 375209

Ion Ratio Lower Upper



63 100


106 30.7 24.6 36.8



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V









#59
2-Hexanone
Concen: 100.589 ug/l
RT: 9.49 min Scan# 1368
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 43 Resp: 577916
Ion Ratio Lower Upper
43 100
58 55.0 27.7 83.1



#60
Dibromochloromethane
Concen: 18.300 ug/l
RT: 9.59 min Scan# 1383
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 129 Resp: 96766
Ion Ratio Lower Upper
129 100
127 75.0 39.3 117.8





#61

1,2-Dibromoethane

Concen: 19.538 ug/l

RT: 9.67 min Scan# 1397

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 107 Resp: 117130

Ion Ratio Lower Upper




107 100

109 94.9 75.8 113.6



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 51.171 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 95 Resp: 333657

Ion Ratio Lower Upper

95 100

174 92.6 0.0 182.2



176 89.9 0.0 178.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion:117 Resp: 729350
Ion Ratio Lower Upper
117 100
82 48.8 41.0 61.6
119 31.7 25.4 38.0



#64
Tetrachloroethene
Concen: 20.341 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion:164 Resp: 128441
Ion Ratio Lower Upper
164 100
166 126.6 102.5 153.7
129 88.7 75.1 112.7
131 85.2 72.7 109.1









#65
Chlorobenzene
Concen: 19.425 ug/l
RT: 10.14 min Scan# 1474
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion:112 Resp: 309011
Ion Ratio Lower Upper
112 100
114 33.6 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 18.723 ug/l
RT: 10.22 min Scan# 1487
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion:131 Resp: 100194
Ion Ratio Lower Upper
131 100
133 95.7 48.0 144.2
119 66.2 32.6 97.8





#67

Ethyl Benzene

Concen: 19.555 ug/l

RT: 10.25 min Scan# 1492

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 91 Resp: 513532

Ion Ratio Lower Upper




91 100

106 31.4 26.4 39.6



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 106.00 (105.70 to 106.70): V



#68

m/p-Xylenes

Concen: 39.395 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 106 Resp: 407605

Ion Ratio Lower Upper



106 100


91 195.3 155.8 233.6



Abundance Ion 106.00 (105.70 to 106.70): V

Ion 91.00 (90.70 to 91.70): VX0









#69
o-Xylene
Concen: 19.901 ug/l
RT: 10.69 min Scan# 1565
Delta R.T. -0.01 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion:106 Resp: 200811
Ion Ratio Lower Upper
106 100
91 204.8 101.3 303.7



#70
Styrene
Concen: 19.978 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion:104 Resp: 319273
Ion Ratio Lower Upper
104 100
78 47.9 37.7 56.5
103 56.6 43.8 65.8





#71

Bromoform

Concen: 17.486 ug/l

RT: 10.86 min Scan# 1592

Delta R.T. 0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion:173 Resp: 67464

Ion Ratio Lower Upper

173 100

175 48.6 24.7 74.1




254 0.2 0.2 0.2



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion:152 Resp: 380327

Ion Ratio Lower Upper

152 100

115 64.0 39.1 117.2



150 163.8 0.0 344.8




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 19.503 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion: 105 Resp: 523036

Ion Ratio Lower Upper

105 100

120 27.9 14.1 42.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

500000

400000

300000

200000

100000



0

11.02

11.00

10.90

Time-->



#74

N-ethyl acetate

Concen: 19.052 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Tgt Ion: 43 Resp: 205712


Ion Ratio Lower Upper

43 100

70 43.4 35.8 53.8

55 25.3 19.4 29.0

61 24.0 20.1 30.1



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

250000

200000

150000

100000


50000


0

10.90

11.00

Time-->





#75

1,1,2,2-Tetrachloroethane

Concen: 18.976 ug/l

RT: 11.27 min Scan# 1659

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion: 83 Resp: 161564

Ion Ratio Lower Upper

83 100

131 11.3 5.7 17.1




85 65.0 32.2 96.6



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): VX0



#76

1,2,3-Trichloropropane

Concen: 22.665 ug/l

RT: 11.29 min Scan# 1663

Delta R.T. -0.01 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 75 Resp: 174684

Ion Ratio Lower Upper



75 100


77 33.1 18.5 55.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#77

Bromobenzene

Concen: 19.734 ug/l

RT: 11.26 min Scan# 1657

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion:156 Resp: 145825

Ion Ratio Lower Upper

156 100

77 130.4 68.0 204.0




158 96.4 48.3 144.8



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V



#78

n-propylbenzene

Concen: 19.747 ug/l

RT: 11.36 min Scan# 1674

Delta R.T. -0.00 min

Lab File: VX007010.D


Acq: 11 Jan 2019 13:44

Tgt Ion: 91 Resp: 581378

Ion Ratio Lower Upper



91 100


120 24.9 12.7 38.0



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 120.00 (119.70 to 120.70): V









#79
 2-Chlorotoluene
 Concen: 19.387 ug/l
 RT: 11.42 min Scan# 1684
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampled :
 VX0111MBS01


Tgt Ion: 91 Resp: 343579
 Ion Ratio Lower Upper
 91 100
 126 38.7 18.6 56.0



#80
 1,3,5-Trimethylbenzene
 Concen: 19.419 ug/l
 RT: 11.51 min Scan# 1698
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion: 105 Resp: 435155
 Ion Ratio Lower Upper
 105 100
 120 50.1 25.1 75.3





#81

trans-1,4-Dichloro-2-butene

Concen: 16.262 ug/l

RT: 11.07 min Scan# 1627

Delta R.T. 0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01





Tgt Ion: 75 Resp: 34479

Ion Ratio Lower Upper

75 100

53 118.8 79.7 119.5

89 47.3 38.0 57.0



#82

4-Chlorotoluene

Concen: 19.572 ug/l

RT: 11.51 min Scan# 1699

Delta R.T. -0.00 min

Lab File: VX007010.D




Acq: 11 Jan 2019 13:44


Tgt Ion: 91 Resp: 401599

Ion Ratio Lower Upper

91 100

126 32.7 16.3 48.9









#83
tert-Butylbenzene
Concen: 18.887 ug/l
RT: 11.77 min Scan# 1741
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion:119 Resp: 422280
Ion Ratio Lower Upper
119 100
91 57.3 27.7 83.1
134 24.7 11.9 35.5



#84
1,2,4-Trimethylbenzene
Concen: 19.789 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion:105 Resp: 445927
Ion Ratio Lower Upper
105 100
120 46.3 23.2 69.6





#85

sec-Butylbenzene

Concen: 19.649 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :


VX0111MBS01

Tgt Ion:105 Resp: 524652

Ion Ratio Lower Upper

105 100

134 21.1 10.7 31.9



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V

11.94

400000



300000

200000

100000

0

Time-->



#86

p-Isopropyltoluene

Concen: 19.550 ug/l

RT: 12.06 min Scan# 1789

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion:119 Resp: 466931

Ion Ratio Lower Upper

119 100

134 27.6 14.0 41.9

91 22.5 11.3 33.9



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0

12.06

400000


300000


200000

100000

0

Time-->





#87

1,3-Dichlorobenzene

Concen: 19.501 ug/l

RT: 12.02 min Scan# 1783

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion:146 Resp: 260461

Ion Ratio Lower Upper




146 100

111 36.5 18.6 55.8

148 63.6 31.8 95.3



Abundance Ion 145.90 (145.60 to 146.60): V
Ion 110.90 (110.60 to 111.60): V
Ion 147.90 (147.60 to 148.60): V



#88

1,4-Dichlorobenzene

Concen: 20.387 ug/l

RT: 12.10 min Scan# 1795

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion:146 Resp: 256909

Ion Ratio Lower Upper



146 100


111 38.0 18.1 54.1

148 64.9 31.9 95.9



Abundance Ion 145.90 (145.60 to 146.60): V
Ion 110.90 (110.60 to 111.60): V
Ion 147.90 (147.60 to 148.60): V









#89
n-Butylbenzene
Concen: 19.161 ug/l
RT: 12.39 min Scan# 1843
Delta R.T. -0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44




Instrument :
MSVOA_X
ClientSampleId :
VX0111MBS01


Tgt Ion: 91 Resp: 390120
Ion Ratio Lower Upper
91 100
92 52.5 26.8 80.4
134 28.4 14.4 43.2



#90
Hexachloroethane
Concen: 16.992 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. 0.00 min
Lab File: VX007010.D
Acq: 11 Jan 2019 13:44

Tgt Ion: 117 Resp: 57016
Ion Ratio Lower Upper
117 100
201 90.3 42.9 128.7





#91

1,2-Dichlorobenzene

Concen: 19.504 ug/l

RT: 12.39 min Scan# 1843

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampled :

VX0111MBS01


Tgt Ion:146 Resp: 257083

Ion Ratio Lower Upper

146 100

111 38.8 19.4 58.2




148 64.3 32.3 96.9



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 17.125 ug/l

RT: 13.00 min Scan# 1943

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion: 75 Resp: 28836

Ion Ratio Lower Upper

75 100

155 103.0 50.6 151.8



157 133.3 64.6 193.8




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 19.358 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111MBS01


Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.5	47.3	142.0
145	28.7	14.8	44.3



#94
 Hexachlorobutadiene
 Concen: 19.367 ug/l
 RT: 13.78 min Scan# 2071
 Delta R.T. -0.00 min
 Lab File: VX007010.D
 Acq: 11 Jan 2019 13:44

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.0	32.1	96.3
227	64.6	32.3	96.8





#95

Naphthalene

Concen: 19.437 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44

Instrument :

MSVOA_X

ClientSampleId :

VX0111MBS01


Tgt Ion:128 Resp: 532610

Ion Ratio Lower Upper




128 100

127 12.9 10.1 15.1

129 10.7 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 19.677 ug/l

RT: 14.02 min Scan# 2110

Delta R.T. -0.00 min

Lab File: VX007010.D

Acq: 11 Jan 2019 13:44


Tgt Ion:180 Resp: 176759

Ion Ratio Lower Upper

180 100

182 94.1 48.5 145.5

145 31.1 16.0 47.9



Abundance Ion 179.80 (179.50 to 180.50): V
Ion 181.80 (181.50 to 182.50): V
Ion 144.90 (144.60 to 145.60): V

