

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007012.D
 Acq On : 11 Jan 2019 14:37
 Operator : JC/SP
 Sample : K1094-03ME 10X
 Misc : 4.20g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CROSSWICKS-DRUM-2ME

Quant Time: Jan 14 09:27:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	599939	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	912806	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	862394	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	431749	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	313802	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
35) Dibromofluoromethane	5.50	113	268990	46.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.04%	
50) Toluene-d8	8.71	98	1094963	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
62) 4-Bromofluorobenzene	11.13	95	395294	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
Target Compounds						
67) Ethyl Benzene	10.25	91	280154	9.022	ug/l	99
68) m/p-Xylenes	10.36	106	618878	50.587	ug/l	100
69) o-Xylene	10.70	106	308704	25.874	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	125321	4.926	ug/l	99
84) 1,2,4-Trimethylbenzene	11.80	105	604202	23.619	ug/l	99
95) Naphthalene	13.83	128	35919	1.155	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007012.D
Acq On : 11 Jan 2019 14:37
Operator : JC/SP
Sample : K1094-03ME 10X
Misc : 4.20g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CROSSWICKS-DRUM-2ME

Quant Time: Jan 14 09:27:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

