

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007017.D
 Acq On : 11 Jan 2019 16:50
 Operator : JC/SP
 Sample : K1125-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 11 22:38:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	573744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	894404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	835208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	413883	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	305965	51.03	ug/l	0.00
Spiked Amount						
			Recovery	=	102.06%	
35) Dibromofluoromethane	5.50	113	274021	47.84	ug/l	0.00
Spiked Amount						
			Recovery	=	95.68%	
50) Toluene-d8	8.71	98	1061727	50.64	ug/l	0.00
Spiked Amount						
			Recovery	=	101.28%	
62) 4-Bromofluorobenzene	11.13	95	387062	52.96	ug/l	0.00
Spiked Amount						
			Recovery	=	105.92%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007017.D
Acq On : 11 Jan 2019 16:50
Operator : JC/SP
Sample : K1125-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 11 22:38:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

