

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007023.D
 Acq On : 11 Jan 2019 19:29
 Operator : JC/SP
 Sample : K1133-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022MW357-190109

Manual Integrations
 APPROVED

MMDadoda
 1/14/2019 10:40:26 AM

Quant Time: Jan 14 09:33:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	542781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	837728	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	789633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	388154	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.07	65	293549	51.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.50%	
35) Dibromofluoromethane	5.50	113	253555	47.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.52%	
50) Toluene-d8	8.71	98	1006218	51.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.48%	
62) 4-Bromofluorobenzene	11.13	95	363657	53.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.24%	

Target Compounds

					Qvalue
4) Vinyl Chloride	1.41	62	264473	47.382	ug/l 98
21) trans-1,2-Dichloroethene	3.16	96	9234	1.661	ug/l 95
24) 1,1-Dichloroethane	3.70	63	34017	3.785	ug/l 97
27) cis-1,2-Dichloroethene	4.60	96	191010	31.150	ug/l 93
40) Benzene	6.15	78	40548	1.825	ug/l 98
42) 1,2-Dichloroethane	6.20	62	5582	0.727	ug/l 77
44) Trichloroethene	7.21	130	16123	2.421	ug/l 100
64) Tetrachloroethene	9.33	164	2185	0.320	ug/l 89
65) Chlorobenzene	10.14	112	1053093	61.147	ug/l 99
87) 1,3-Dichlorobenzene	12.02	146	4865	0.357	ug/l 98
88) 1,4-Dichlorobenzene	12.10	146	19471m	1.648	ug/l
91) 1,2-Dichlorobenzene	12.39	146	13458	1.000	ug/l 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007023.D
Acq On : 11 Jan 2019 19:29
Operator : JC/SP
Sample : K1133-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022MW357-190109

Manual Integrations
APPROVED

MMDadoda
1/14/2019 10:40:26 AM

Quant Time: Jan 14 09:33:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

