

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007024.D
 Acq On : 11 Jan 2019 19:56
 Operator : JC/SP
 Sample : K1115-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S72-IDW01

Quant Time: Jan 11 22:58:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	553066	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	870513	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	836285	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	401542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	304820	52.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.48%	
35) Dibromofluoromethane	5.51	113	268320	48.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.26%	
50) Toluene-d8	8.71	98	1057283	51.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.62%	
62) 4-Bromofluorobenzene	11.13	95	380398	53.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.94%	
Target Compounds						
16) Acetone	2.45	43	16629	5.820	ug/l	90
43) Isopropyl Acetate	6.46	43	25622	2.002	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007024.D
Acq On : 11 Jan 2019 19:56
Operator : JC/SP
Sample : K1115-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S72-IDW01

Quant Time: Jan 11 22:58:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

