

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007025.D
 Acq On : 11 Jan 2019 20:23
 Operator : JC/SP
 Sample : K1116-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-IDW03

Quant Time: Jan 11 23:01:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	568595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	893541	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	857055	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	414039	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	310492	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	5.51	113	268199	46.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.74%	
50) Toluene-d8	8.71	98	1078527	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.13	95	388391	53.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.38%	
Target Compounds						
16) Acetone	2.45	43	18151	6.179	ug/l	97
18) Methyl Acetate	2.78	43	9390	1.104	ug/l	99
43) Isopropyl Acetate	6.46	43	25916	1.973	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007025.D
Acq On : 11 Jan 2019 20:23
Operator : JC/SP
Sample : K1116-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-IDW03

Quant Time: Jan 11 23:01:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

