

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX011119\
 Data File : VX007026.D
 Acq On : 11 Jan 2019 20:49
 Operator : JC/SP
 Sample : K1117-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-IDW22

Quant Time: Jan 11 23:04:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 08 14:06:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	566664	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	901858	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	855267	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	417279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	311737	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
35) Dibromofluoromethane	5.50	113	273609	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
50) Toluene-d8	8.71	98	1088666	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
62) 4-Bromofluorobenzene	11.13	95	391556	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
Target Compounds						
16) Acetone	2.45	43	23468	8.016	ug/l #	85
18) Methyl Acetate	2.78	43	12165	1.435	ug/l	93
43) Isopropyl Acetate	6.46	43	23877	1.801	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX011119\
Data File : VX007026.D
Acq On : 11 Jan 2019 20:49
Operator : JC/SP
Sample : K1117-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-IDW22

Quant Time: Jan 11 23:04:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 08 14:06:59 2019
Response via : Initial Calibration

