

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020568.D
 Acq On : 11 Jan 2021 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 12 02:57:25 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	177263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	271461	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	246624	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	132260	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	104241	48.96	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.92%
35) Dibromofluoromethane	5.458	113	91588	48.69	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.38%
50) Toluene-d8	8.695	98	344393	51.17	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.34%
62) 4-Bromofluorobenzene	11.122	95	120208	50.58	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.16%

Target Compounds					Qvalue
2) Dichlorodifluoromethane	1.185	85	93785	53.63	ug/l 99
3) Chloromethane	1.313	50	84108	51.98	ug/l 99
4) Vinyl Chloride	1.392	62	97379	52.48	ug/l 100
5) Bromomethane	1.630	94	57658	46.01	ug/l 99
6) Chloroethane	1.709	64	63124	57.54	ug/l 98
7) Trichlorofluoromethane	1.916	101	159065	51.77	ug/l 98
8) Diethyl Ether	2.166	74	55125	50.35	ug/l 94
9) 1,1,2-Trichlorotrifluo...	2.361	101	89957	52.18	ug/l 97
10) Methyl Iodide	2.489	142	121825	50.67	ug/l 97
11) Tert butyl alcohol	3.020	59	95891	245.10	ug/l 99
12) 1,1-Dichloroethene	2.355	96	86696	51.23	ug/l 93
13) Acrolein	2.270	56	75487	255.79	ug/l 100
14) Allyl chloride	2.709	41	116481	49.54	ug/l 96
15) Acrylonitrile	3.111	53	235228	260.19	ug/l 99
16) Acetone	2.422	43	185573	255.14	ug/l 100
17) Carbon Disulfide	2.550	76	240142	51.47	ug/l 99
18) Methyl Acetate	2.745	43	94114	49.89	ug/l 95
19) Methyl tert-butyl Ether	3.166	73	289926	52.82	ug/l 96
20) Methylene Chloride	2.831	84	99477	49.45	ug/l 95
21) trans-1,2-Dichloroethene	3.142	96	98242	51.68	ug/l 94
22) Diisopropyl ether	3.824	45	251380	52.95	ug/l 94
23) Vinyl Acetate	3.782	43	1089293	268.82	ug/l 97
24) 1,1-Dichloroethane	3.666	63	159744	50.66	ug/l 99
25) 2-Butanone	4.635	43	291862	253.63	ug/l 96
26) 2,2-Dichloropropane	4.550	77	133282	45.44	ug/l 100
27) cis-1,2-Dichloroethene	4.562	96	108336	51.07	ug/l 94
28) Bromochloromethane	4.977	49	63455	49.42	ug/l 84
29) Tetrahydrofuran	5.086	42	188961	269.59	ug/l 94
30) Chloroform	5.178	83	179655	50.95	ug/l 97
31) Cyclohexane	5.544	56	133267	53.27	ug/l 95
32) 1,1,1-Trichloroethane	5.458	97	162793	51.33	ug/l 98
36) 1,1-Dichloropropene	5.769	75	127242	51.30	ug/l 98
37) Ethyl Acetate	4.788	43	115244	51.15	ug/l 99
38) Carbon Tetrachloride	5.751	117	145127	50.14	ug/l 97
39) Methylcyclohexane	7.440	83	148656	52.48	ug/l 95
40) Benzene	6.111	78	384881	51.63	ug/l 100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020568.D
 Acq On : 11 Jan 2021 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 12 02:57:25 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	5.007	41	61861	52.46	ug/l	94
42) 1,2-Dichloroethane	6.159	62	136889	51.50	ug/l	100
43) Isopropyl Acetate	6.409	43	181443	51.95	ug/l	98
44) Trichloroethene	7.190	130	110950	50.08	ug/l	90
45) 1,2-Dichloropropane	7.488	63	94935	51.91	ug/l	100
46) Dibromomethane	7.635	93	70914	51.07	ug/l	96
47) Bromodichloromethane	7.872	83	144108	50.85	ug/l	98
48) Methyl methacrylate	7.744	41	92003	54.19	ug/l	96
49) 1,4-Dioxane	7.726	88	47319	1105.54	ug/l	94
51) 4-Methyl-2-Pentanone	8.616	43	604935	279.24	ug/l	97
52) Toluene	8.763	92	242659	51.19	ug/l	99
53) t-1,3-Dichloropropene	9.019	75	153313	53.12	ug/l	100
54) cis-1,3-Dichloropropene	8.415	75	167837	54.29	ug/l	94
55) 1,1,2-Trichloroethane	9.195	97	105054	52.45	ug/l	96
56) Ethyl methacrylate	9.159	69	146172	56.29	ug/l	93
57) 1,3-Dichloropropane	9.354	76	173451	53.29	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.287	63	356955	262.36	ug/l	95
59) 2-Hexanone	9.470	43	459466	282.92	ug/l	98
60) Dibromochloromethane	9.561	129	115476	51.04	ug/l	98
61) 1,2-Dibromoethane	9.653	107	112488	52.64	ug/l	99
64) Tetrachloroethene	9.317	164	113940	51.15	ug/l	90
65) Chlorobenzene	10.122	112	259724	50.04	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.201	131	101696	51.33	ug/l	98
67) Ethyl Benzene	10.232	91	459426	53.77	ug/l	100
68) m/p-Xylenes	10.342	106	374009	113.57	ug/l	100
69) o-Xylene	10.683	106	173341	57.13	ug/l	99
70) Styrene	10.695	104	303272	58.56	ug/l	99
71) Bromoform	10.841	173	84329	52.45	ug/l	96
73) Isopropylbenzene	11.000	105	462849	51.90	ug/l	100
74) N-amyl acetate	10.878	43	149312	47.49	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.250	83	148061	50.98	ug/l	98
76) 1,2,3-Trichloropropane	11.280	75	173067	62.12	ug/l	88
77) Bromobenzene	11.238	156	122552	51.29	ug/l	89
78) n-propylbenzene	11.341	91	517604	52.52	ug/l	99
79) 2-Chlorotoluene	11.402	91	307726	50.60	ug/l	96
80) 1,3,5-Trimethylbenzene	11.488	105	398968	53.39	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.055	75	46986	45.95	ug/l	99
82) 4-Chlorotoluene	11.494	91	372095	51.57	ug/l	100
83) tert-Butylbenzene	11.756	119	367285	51.18	ug/l	97
84) 1,2,4-Trimethylbenzene	11.792	105	399032	53.42	ug/l	100
85) sec-Butylbenzene	11.927	105	433968	51.82	ug/l	99
86) p-Isopropyltoluene	12.049	119	407454	52.39	ug/l	98
87) 1,3-Dichlorobenzene	12.006	146	217940	49.91	ug/l	97
88) 1,4-Dichlorobenzene	12.079	146	219850	49.11	ug/l	98
89) n-Butylbenzene	12.372	91	343720	51.07	ug/l	99
90) Hexachloroethane	12.579	117	72374	49.34	ug/l	99
91) 1,2-Dichlorobenzene	12.372	146	212727	49.34	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.981	75	34715	50.91	ug/l	95
93) 1,2,4-Trichlorobenzene	13.628	180	139633	48.58	ug/l	97
94) Hexachlorobutadiene	13.768	225	64214	43.87	ug/l	98
95) Naphthalene	13.817	128	461569	47.39	ug/l	100
96) 1,2,3-Trichlorobenzene	13.999	180	141718	48.49	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
Data File : VX020568.D
Acq On : 11 Jan 2021 22:32
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Jan 12 02:57:25 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020568.D
 Acq On : 11 Jan 2021 22:32
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Jan 12 02:57:25 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.629 min Scan# 745
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:168 Resp: 177263
Ion Ratio Lower Upper
168 100
99 52.0 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 53.63 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 85 Resp: 93785
Ion Ratio Lower Upper
85 100
87 33.3 16.4 49.2





#3

Chloromethane

Concen: 51.98 ug/l

RT: 1.313 min Scan# 37

Delta R.T. -0.000 min

Lab File: VX020568.D





Acq: 11 Jan 2021 22:32

Tgt Ion: 50 Resp: 84108

Ion Ratio Lower Upper

50 100

52 33.5 26.4 39.6



#4

Vinyl Chloride

Concen: 52.48 ug/l

RT: 1.392 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX020568.D




Acq: 11 Jan 2021 22:32


Tgt Ion: 62 Resp: 97379

Ion Ratio Lower Upper

62 100

64 32.4 25.7 38.5









#5
Bromomethane
Concen: 46.01 ug/l
RT: 1.630 min Scan# 89
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 94 Resp: 57658
Ion Ratio Lower Upper
94 100
96 94.4 76.4 114.6



#6
Chloroethane
Concen: 57.54 ug/l
RT: 1.709 min Scan# 102
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 64 Resp: 63124
Ion Ratio Lower Upper
64 100
66 32.6 25.4 38.0









#7
Trichlorofluoromethane
Concen: 51.77 ug/l
RT: 1.916 min Scan# 136
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:101 Resp: 159065
Ion Ratio Lower Upper
101 100
103 65.2 53.2 79.8



#8
Diethyl Ether
Concen: 50.35 ug/l
RT: 2.166 min Scan# 177
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Tgt Ion: 74 Resp: 55125
Ion Ratio Lower Upper
74 100
45 83.0 44.5 133.5








#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.18 ug/l
RT: 2.361 min Scan# 209
Delta R.T. -0.007 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:101 Resp: 89957
Ion Ratio Lower Upper
101 100
85 41.5 34.9 52.3
151 79.7 61.4 92.2




#10
Methyl Iodide
Concen: 50.67 ug/l
RT: 2.489 min Scan# 230
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32



Tgt Ion:142 Resp: 121825
Ion Ratio Lower Upper
142 100
127 43.9 33.3 49.9
141 14.5 11.4 17.0









#11
Tert butyl alcohol
Concen: 245.10 ug/l
RT: 3.020 min Scan# 317
Delta R.T. -0.018 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 59 Resp: 95891
Ion Ratio Lower Upper
59 100
57 10.0 7.8 11.8



#12
1,1-Dichloroethene
Concen: 51.23 ug/l
RT: 2.355 min Scan# 208
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 96 Resp: 86696
Ion Ratio Lower Upper
96 100
61 142.5 124.6 187.0
98 63.2 50.6 76.0









#13
Acrolein
Concen: 255.79 ug/l
RT: 2.270 min Scan# 194
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 56 Resp: 75487
Ion Ratio Lower Upper
56 100
55 70.4 56.4 84.6



#14
Allyl chloride
Concen: 49.54 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 41 Resp: 116481
Ion Ratio Lower Upper
41 100
39 67.1 51.6 77.4
76 42.0 30.5 45.7









#15
Acrylonitrile
Concen: 260.19 ug/l
RT: 3.111 min Scan# 332
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 53 Resp: 235228
Ion Ratio Lower Upper
53 100
52 82.7 65.3 97.9
51 35.7 28.2 42.2



#16
Acetone
Concen: 255.14 ug/l
RT: 2.422 min Scan# 219
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 43 Resp: 185573
Ion Ratio Lower Upper
43 100
58 35.1 27.8 41.8









#17
Carbon Disulfide
Concen: 51.47 ug/l
RT: 2.550 min Scan# 240
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 240142
Ion Ratio Lower Upper
76 100
78 9.4 7.2 10.8



#18
Methyl Acetate
Concen: 49.89 ug/l
RT: 2.745 min Scan# 272
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 43 Resp: 94114
Ion Ratio Lower Upper
43 100
74 30.2 21.9 32.9









#19

Methyl tert-butyl Ether
 Concen: 52.82 ug/l
 RT: 3.166 min Scan# 341
 Delta R.T. -0.000 min
 Lab File: VX020568.D
 Acq: 11 Jan 2021 22:32

Instrument :
 MSVOA_X
 ClientSampled :




Tgt Ion: 73 Resp: 289926
 Ion Ratio Lower Upper
 73 100
 57 19.7 17.1 25.7




#20

Methylene Chloride
 Concen: 49.45 ug/l
 RT: 2.831 min Scan# 286
 Delta R.T. -0.000 min
 Lab File: VX020568.D
 Acq: 11 Jan 2021 22:32


Tgt Ion: 84 Resp: 99477
 Ion Ratio Lower Upper
 84 100
 49 103.2 88.6 132.8
 51 33.4 27.0 40.6
 86 66.9 50.3 75.5








#21
trans-1,2-Dichloroethene
Concen: 51.68 ug/l
RT: 3.142 min Scan# 337
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Instrument :
MSVOA_X
ClientSampled :





Tgt Ion: 96 Resp: 98242
Ion Ratio Lower Upper
96 100
61 124.4 107.7 161.5
98 64.4 50.4 75.6




#22
Diisopropyl ether
Concen: 52.95 ug/l
RT: 3.824 min Scan# 449
Delta R.T. 0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32



Tgt Ion: 45 Resp: 251380
Ion Ratio Lower Upper
45 100
43 66.3 58.0 87.0
87 34.7 25.4 38.0
59 12.8 10.2 15.2









#23
 Vinyl Acetate
 Concen: 268.82 ug/l
 RT: 3.782 min Scan# 442
 Delta R.T. -0.000 min
 Lab File: VX020568.D
 Acq: 11 Jan 2021 22:32




Instrument :
 MSVOA_X
 ClientSampled :


Tgt Ion: 43 Resp: 1089293
 Ion Ratio Lower Upper
 43 100
 86 13.7 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 50.66 ug/l
 RT: 3.666 min Scan# 423
 Delta R.T. -0.000 min
 Lab File: VX020568.D
 Acq: 11 Jan 2021 22:32

Tgt Ion: 63 Resp: 159744
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.8 11.2
 100 5.0 2.5 7.5









#25
2-Butanone
Concen: 253.63 ug/l
RT: 4.635 min Scan# 582
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 291862
Ion Ratio Lower Upper
43 100
72 31.5 23.7 35.5



#26
2,2-Dichloropropane
Concen: 45.44 ug/l
RT: 4.550 min Scan# 568
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 77 Resp: 133282
Ion Ratio Lower Upper
77 100
97 23.5 11.7 35.1









#27

cis-1,2-Dichloroethene
Concen: 51.07 ug/l
RT: 4.562 min Scan# 570
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Instrument :
MSVOA_X
ClientSampled :




Tgt Ion: 96 Resp: 108336
Ion Ratio Lower Upper
96 100
61 132.1 0.0 285.6
98 65.7 0.0 129.6




#28

Bromochloromethane
Concen: 49.42 ug/l
RT: 4.977 min Scan# 638
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 49 Resp: 63455
Ion Ratio Lower Upper
49 100
129 2.6 0.0 4.0
130 100.3 68.2 102.2





#29

Tetrahydrofuran

Concen: 269.59 ug/l

RT: 5.086 min Scan# 656





Delta R.T. -0.007 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion: 42 Resp: 188961

Ion	Ratio	Lower	Upper
42	100		
72	55.8	41.7	62.5
71	52.5	38.4	57.6



#30

Chloroform

Concen: 50.95 ug/l

RT: 5.178 min Scan# 671




Delta R.T. 0.006 min


Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion: 83 Resp: 179655


Ion	Ratio	Lower	Upper
83	100		
85	62.5	52.0	78.0








#31
Cyclohexane
Concen: 53.27 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Instrument :
MSVOA_X
ClientSampled :







Tgt Ion: 56 Resp: 133267
Ion Ratio Lower Upper
56 100
69 33.1 25.3 37.9
84 94.5 71.4 107.0



#32
1,1,1-Trichloroethane
Concen: 51.33 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.007 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 97 Resp: 162793
Ion Ratio Lower Upper
97 100
99 64.4 51.5 77.3
61 40.0 34.2 51.4





#33

1,2-Dichloroethane-d4

Concen: 48.96 ug/l

RT: 6.025 min Scan# 810

Delta R.T. 0.000 min

Lab File: VX020568.D





Acq: 11 Jan 2021 22:32

Tgt Ion: 65 Resp: 104241

Ion Ratio Lower Upper

65 100

67 52.2 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.830 min Scan# 942

Delta R.T. -0.000 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32




Tgt Ion: 114 Resp: 271461


Ion Ratio Lower Upper

114 100

63 18.5 0.0 40.8

88 15.6 0.0 32.8









#35
Dibromofluoromethane
Concen: 48.69 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.007 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 91588
Ion Ratio Lower Upper
113 100
111 102.8 81.9 122.9
192 21.4 16.7 25.1



#36
1,1-Dichloropropene
Concen: 51.30 ug/l
RT: 5.769 min Scan# 768
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Tgt Ion: 75 Resp: 127242
Ion Ratio Lower Upper
75 100
110 37.6 18.1 54.4
77 31.3 25.3 37.9








#37
Ethyl Acetate
Concen: 51.15 ug/l
RT: 4.788 min Scan# 607
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 43 Resp: 115244
Ion Ratio Lower Upper
43 100
61 15.9 12.8 19.2
70 13.8 10.5 15.7




#38
Carbon Tetrachloride
Concen: 50.14 ug/l
RT: 5.751 min Scan# 765
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32



Tgt Ion: 117 Resp: 145127
Ion Ratio Lower Upper
117 100
119 98.7 76.2 114.2
121 31.6 24.8 37.2





#39

Methylcyclohexane

Concen: 52.48 ug/l

RT: 7.440 min Scan# 1042





Delta R.T. -0.000 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion: 83 Resp: 148656

Ion	Ratio	Lower	Upper
83	100		
55	70.3	58.9	88.3
98	51.4	38.1	57.1



#40

Benzene

Concen: 51.63 ug/l

RT: 6.111 min Scan# 824




Delta R.T. -0.000 min

Lab File: VX020568.D


Acq: 11 Jan 2021 22:32

Tgt Ion: 78 Resp: 384881

Ion	Ratio	Lower	Upper
78	100		
77	23.7	18.9	28.3







Instrument :
MSVOA_X
ClientSampleId :



#41
Methacrylonitrile
Concen: 52.46 ug/l
RT: 5.007 min Scan# 643
Delta R.T. 0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:	41	Resp:	61861
Ion	Ratio	Lower	Upper
41	100		
39	55.1	42.5	63.7
67	88.6	65.0	97.6
52	33.8	25.3	37.9



#42
1,2-Dichloroethane
Concen: 51.50 ug/l
RT: 6.159 min Scan# 832
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:	62	Resp:	136889
Ion	Ratio	Lower	Upper
62	100		
98	10.2	0.0	20.2









#43
Isopropyl Acetate
Concen: 51.95 ug/l
RT: 6.409 min Scan# 873
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 181443
Ion Ratio Lower Upper
43 100
61 23.3 18.2 27.4
87 17.2 12.5 18.7



#44
Trichloroethene
Concen: 50.08 ug/l
RT: 7.190 min Scan# 1001
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:130 Resp: 110950
Ion Ratio Lower Upper
130 100
95 88.8 0.0 198.0





#45

1,2-Dichloropropane

Concen: 51.91 ug/l

RT: 7.488 min Scan# 1050

Delta R.T. -0.001 min

Lab File: VX020568.D





Acq: 11 Jan 2021 22:32

Tgt Ion: 63 Resp: 94935

Ion Ratio Lower Upper

63 100

65 32.2 25.7 38.5



#46

Dibromomethane

Concen: 51.07 ug/l

RT: 7.635 min Scan# 1074

Delta R.T. -0.000 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32




Tgt Ion: 93 Resp: 70914


Ion Ratio Lower Upper

93 100

95 83.1 66.8 100.2

174 110.9 83.1 124.7









#47
Bromodichloromethane
Concen: 50.85 ug/l
RT: 7.872 min Scan# 1113
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 144108
Ion Ratio Lower Upper
83 100
85 62.9 51.4 77.0
127 9.0 6.8 10.2



#48
Methyl methacrylate
Concen: 54.19 ug/l
RT: 7.744 min Scan# 1092
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 41 Resp: 92003
Ion Ratio Lower Upper
41 100
69 99.6 76.0 114.0
39 57.3 44.6 67.0





#49

1,4-Dioxane

Concen: 1105.54 ug/l

RT: 7.726 min Scan# 1089





Delta R.T. -0.019 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion: 88 Resp: 47319

Ion	Ratio	Lower	Upper
88	100		
43	26.1	23.2	34.8
58	63.6	55.0	82.4



#50

Toluene-d8

Concen: 51.17 ug/l

RT: 8.695 min Scan# 1248




Delta R.T. -0.001 min


Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion: 98 Resp: 344393

Ion	Ratio	Lower	Upper
98	100		
100	66.4	51.9	77.9









#51
4-Methyl-2-Pentanone
Concen: 279.24 ug/l
RT: 8.616 min Scan# 1235
Delta R.T. -0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 604935
Ion Ratio Lower Upper
43 100
58 46.2 35.4 53.0



#52
Toluene
Concen: 51.19 ug/l
RT: 8.763 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 92 Resp: 242659
Ion Ratio Lower Upper
92 100
91 170.6 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 53.12 ug/l
RT: 9.019 min Scan# 1301
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 75 Resp: 153313
Ion Ratio Lower Upper
75 100
77 31.2 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 54.29 ug/l
RT: 8.415 min Scan# 1202
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 75 Resp: 167837
Ion Ratio Lower Upper
75 100
77 31.6 25.7 38.5
39 37.5 35.1 52.7









#55
1,1,2-Trichloroethane
Concen: 52.45 ug/l
RT: 9.195 min Scan# 1330
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Instrument :
MSVOA_X
ClientSampled :

Tgt Ion: 97 Resp: 105054




Ion	Ratio	Lower	Upper
97	100		
83	79.6	69.3	103.9
85	54.5	44.3	66.5
99	60.8	50.3	75.5




#56
Ethyl methacrylate
Concen: 56.29 ug/l
RT: 9.159 min Scan# 1324
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 69 Resp: 146172

Ion	Ratio	Lower	Upper
69	100		
41	53.5	46.8	70.2
39	27.0	24.7	37.1









#57
1,3-Dichloropropane
Concen: 53.29 ug/l
RT: 9.354 min Scan# 1356
Delta R.T. 0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 76 Resp: 173451
Ion Ratio Lower Upper
76 100
78 32.7 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 262.36 ug/l
RT: 8.287 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 63 Resp: 356955
Ion Ratio Lower Upper
63 100
106 32.0 23.5 35.3









#59
2-Hexanone
Concen: 282.92 ug/l
RT: 9.470 min Scan# 1375
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 43 Resp: 459466
Ion Ratio Lower Upper
43 100
58 63.7 30.9 92.8



#60
Dibromochloromethane
Concen: 51.04 ug/l
RT: 9.561 min Scan# 1390
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 129 Resp: 115476
Ion Ratio Lower Upper
129 100
127 78.4 38.4 115.2









#61
1,2-Dibromoethane
Concen: 52.64 ug/l
RT: 9.653 min Scan# 1405
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:107 Resp: 112488
Ion Ratio Lower Upper
107 100
109 94.0 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.58 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 95 Resp: 120208
Ion Ratio Lower Upper
95 100
174 84.6 0.0 154.6
176 83.1 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.098 min Scan# 1478
Delta R.T. 0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:117 Resp: 246624
Ion Ratio Lower Upper
117 100
82 52.2 47.4 71.2
119 33.1 26.6 39.8



#64
Tetrachloroethene
Concen: 51.15 ug/l
RT: 9.317 min Scan# 1350
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:164 Resp: 113940
Ion Ratio Lower Upper
164 100
166 143.5 103.4 155.2
129 97.8 70.8 106.2
131 98.2 72.5 108.7









#65
Chlorobenzene
Concen: 50.04 ug/l
RT: 10.122 min Scan# 1482
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 259724
Ion Ratio Lower Upper
112 100
114 31.7 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 51.33 ug/l
RT: 10.201 min Scan# 1495
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:131 Resp: 101696
Ion Ratio Lower Upper
131 100
133 96.8 47.7 143.1
119 62.8 32.1 96.5









#67
Ethyl Benzene
Concen: 53.77 ug/l
RT: 10.232 min Scan# 1500
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 459426
Ion Ratio Lower Upper
91 100
106 30.9 24.6 37.0



#68
m/p-Xylenes
Concen: 113.57 ug/l
RT: 10.342 min Scan# 1518
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 106 Resp: 374009
Ion Ratio Lower Upper
106 100
91 203.1 162.8 244.2









#69
o-Xylene
Concen: 57.13 ug/l
RT: 10.683 min Scan# 1574
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 173341
Ion Ratio Lower Upper
106 100
91 214.1 108.1 324.2



#70
Styrene
Concen: 58.56 ug/l
RT: 10.695 min Scan# 1576
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:104 Resp: 303272
Ion Ratio Lower Upper
104 100
78 49.6 40.0 60.0
103 54.6 43.4 65.0









#71
Bromoform
Concen: 52.45 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 84329
Ion Ratio Lower Upper
173 100
175 50.3 23.9 71.8
254 0.1 0.1 0.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:152 Resp: 132260
Ion Ratio Lower Upper
152 100
115 82.2 41.1 123.3
150 178.4 0.0 349.0





#73

Isopropylbenzene

Concen: 51.90 ug/l

RT: 11.000 min Scan# 1626

Delta R.T. -0.000 min





Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion:105 Resp: 462849

Ion Ratio Lower Upper

105	100		
120	26.6	13.2	39.6



#74

N-ethyl acetate

Concen: 47.49 ug/l

RT: 10.878 min Scan# 1606

Delta R.T. 0.000 min




Lab File: VX020568.D


Acq: 11 Jan 2021 22:32

Tgt Ion: 43 Resp: 149312

Ion Ratio Lower Upper

43	100		
70	52.5	39.0	58.6
55	29.6	21.5	32.3
61	28.9	22.1	33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 50.98 ug/l
RT: 11.250 min Scan# 1667
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 148061
Ion Ratio Lower Upper
83 100
131 10.7 5.1 15.2
85 64.8 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 62.12 ug/l
RT: 11.280 min Scan# 1672
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 75 Resp: 173067
Ion Ratio Lower Upper
75 100
77 31.1 19.1 57.5





#77

Bromobenzene

Concen: 51.29 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX020568.D

Acq: 11 Jan 2021 22:32

Tgt Ion:156 Resp: 122552

Ion Ratio Lower Upper

156 100





77 145.6 82.8 248.6

158 94.7 49.1 147.4

Instrument :

MSVOA_X

ClientSampleId :



#78

n-propylbenzene

Concen: 52.52 ug/l

RT: 11.341 min Scan# 1682

Delta R.T. -0.001 min

Lab File: VX020568.D




Acq: 11 Jan 2021 22:32


Tgt Ion: 91 Resp: 517604

Ion Ratio Lower Upper

91 100

120 23.5 11.5 34.5









#79
2-Chlorotoluene
Concen: 50.60 ug/l
RT: 11.402 min Scan# 1692
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion: 91 Resp: 307726
Ion Ratio Lower Upper
91 100
126 35.4 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 53.39 ug/l
RT: 11.488 min Scan# 1706
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 105 Resp: 398968
Ion Ratio Lower Upper
105 100
120 49.0 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 45.95 ug/l
RT: 11.055 min Scan# 1635
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 46986
Ion Ratio Lower Upper
75 100
53 92.0 74.5 111.7
89 47.6 37.9 56.9



#82
4-Chlorotoluene
Concen: 51.57 ug/l
RT: 11.494 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 91 Resp: 372095
Ion Ratio Lower Upper
91 100
126 30.0 14.9 44.5









#83
tert-Butylbenzene
Concen: 51.18 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 367285
Ion Ratio Lower Upper
119 100
91 57.8 27.9 83.7
134 24.2 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 53.42 ug/l
RT: 11.792 min Scan# 1756
Delta R.T. 0.005 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:105 Resp: 399032
Ion Ratio Lower Upper
105 100
120 45.1 22.4 67.3









#85
sec-Butylbenzene
Concen: 51.82 ug/l
RT: 11.927 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 433968
Ion Ratio Lower Upper
105 100
134 20.6 10.0 29.8



#86
p-Isopropyltoluene
Concen: 52.39 ug/l
RT: 12.049 min Scan# 1798
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:119 Resp: 407454
Ion Ratio Lower Upper
119 100
134 26.9 12.9 38.7
91 23.8 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 49.91 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:146 Resp: 217940
Ion Ratio Lower Upper
146 100
111 38.6 20.2 60.5
148 61.0 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 49.11 ug/l
RT: 12.079 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion:146 Resp: 219850
Ion Ratio Lower Upper
146 100
111 38.2 19.6 58.8
148 62.1 31.6 95.0









#89
n-Butylbenzene
Concen: 51.07 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 343720
Ion Ratio Lower Upper
91 100
92 54.3 27.4 82.0
134 27.1 13.2 39.5



#90
Hexachloroethane
Concen: 49.34 ug/l
RT: 12.579 min Scan# 1885
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 117 Resp: 72374
Ion Ratio Lower Upper
117 100
201 81.4 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 49.34 ug/l
RT: 12.372 min Scan# 1851
Delta R.T. -0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:146 Resp: 212727
Ion Ratio Lower Upper
146 100
111 41.3 20.8 62.2
148 65.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.91 ug/l
RT: 12.981 min Scan# 1951
Delta R.T. 0.000 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32

Tgt Ion: 75 Resp: 34715
Ion Ratio Lower Upper
75 100
155 95.2 45.3 135.9
157 120.6 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 48.58 ug/l
RT: 13.628 min Scan# 2057
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32




Instrument :
MSVOA_X
ClientSampled :


Tgt Ion:180 Resp: 139633
Ion Ratio Lower Upper
180 100
182 98.5 47.7 143.1
145 31.2 15.0 45.0



#94
Hexachlorobutadiene
Concen: 43.87 ug/l
RT: 13.768 min Scan# 2080
Delta R.T. 0.006 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Tgt Ion:225 Resp: 64214
Ion Ratio Lower Upper
225 100
223 65.4 31.1 93.2
227 63.6 31.6 95.0








#95
Naphthalene
Concen: 47.39 ug/l
RT: 13.817 min Scan# 2088
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32


Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 461569
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 11.1 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 48.49 ug/l
RT: 13.999 min Scan# 2118
Delta R.T. -0.001 min
Lab File: VX020568.D
Acq: 11 Jan 2021 22:32



Tgt Ion:180 Resp: 141718
Ion Ratio Lower Upper
180 100
182 98.7 47.6 142.9
145 32.9 16.1 48.3

