

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020566.D
 Acq On : 11 Jan 2021 21:46
 Operator : JC/MD
 Sample : M1071-04 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-110B

Quant Time: Jan 12 02:56:48 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	182469	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	297536	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	292030	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	123939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	111930	51.07	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.14%
35) Dibromofluoromethane	5.464	113	101224	49.10	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.20%
50) Toluene-d8	8.696	98	335615	45.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	91.00%
62) 4-Bromofluorobenzene	11.122	95	121235	46.54	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.08%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
Data File : VX020566.D
Acq On : 11 Jan 2021 21:46
Operator : JC/MD
Sample : M1071-04 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-110B

Quant Time: Jan 12 02:56:48 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

