



Chloroethane



$R = -6.125e-002 A^2 + 3.828e-001 A - 3.260e-003$
Coef of Det (r^2) = 0.997571 Curve Fit: Quadratic
Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M
Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

Acetone


Response Ratio



Response = 2.020e-001 * Amt + 1.617e-002
 Coef of Det (r^2) = 0.999828 Curve Fit: Linear
 Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M
 Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

2-Chloroethyl Vinyl ether


Response Ratio



$\text{Response} = 2.550\text{e-}001 * \text{Amt} - 2.330\text{e-}002$
 Coef of Det (r^2) = 0.999368 Curve Fit: Linear
 Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M
 Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

Tert butyl alcohol

Response Ratio



$$\text{Response} = 1.093\text{e-}001 * \text{Amt} + 4.969\text{e-}003$$


Coef of Det (r^2) = 0.999317 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M

Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

Hexachlorobutadiene

Response Ratio



$$\text{Response} = 5.645\text{e-}001 * \text{Amt} - 9.783\text{e-}003$$


Coef of Det (r^2) = 0.998603 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M

Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

Naphthalene


Response Ratio



Response = 4.007e+000 * Amt - 3.082e-001
 Coef of Det (r^2) = 0.996451 Curve Fit: Linear
 Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M
 Calibration Table Last Updated: Mon Jan 11 15:05:37 2021

N-amyl acetate

Response Ratio



$$\text{Response} = 1.220\text{e}+000 * \text{Amt} - 3.017\text{e}-002$$

Coef of Det (r^2) = 0.998550 Curve Fit: Linear

Method Name: Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\82X011121W.M

Calibration Table Last Updated: Mon Jan 11 15:05:37 2021