

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020553.D
 Acq On : 11 Jan 2021 15:27
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011121

Quant Time: Jan 12 02:51:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.493	0.509	-3.2	104	0.00
3 P	Chloromethane	0.456	0.450	1.3	102	0.00
4 C	Vinyl Chloride	0.523	0.526	-0.6#	103	0.00
5 T	Bromomethane	0.353	0.325	7.9	95	0.00
6 T	Chloroethane	0.304	0.330	-8.6	103	0.00
7 T	Trichlorofluoromethane	0.867	0.843	2.8	102	0.00
8 T	Diethyl Ether	0.309	0.297	3.9	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.486	0.492	-1.2	106	0.00
10 T	Methyl Iodide	0.678	0.642	5.3	94	0.00
11 T	Tert butyl alcohol	0.117	0.103	12.0	99	-0.02
12 CM	1,1-Dichloroethene	0.477	0.467	2.1#	103	0.00
13 T	Acrolein	0.083	0.083	0.0	106	0.00
14 T	Allyl chloride	0.663	0.646	2.6	103	0.00
15 T	Acrylonitrile	0.255	0.244	4.3	100	0.00
16 T	Acetone	0.225	0.210	6.7	107	-0.01
17 T	Carbon Disulfide	1.316	1.290	2.0	103	0.00
18 T	Methyl Acetate	0.532	0.492	7.5	101	0.00
19 T	Methyl tert-butyl Ether	1.548	1.553	-0.3	103	0.00
20 T	Methylene Chloride	0.567	0.522	7.9	101	0.00
21 T	trans-1,2-Dichloroethene	0.536	0.522	2.6	102	0.00
22 T	Diisopropyl ether	1.339	1.341	-0.1	102	0.00
23 T	Vinyl Acetate	1.143	1.170	-2.4	103	0.00
24 P	1,1-Dichloroethane	0.889	0.842	5.3	101	0.00
25 T	2-Butanone	0.325	0.318	2.2	104	0.00
26 T	2,2-Dichloropropane	0.827	0.805	2.7	105	0.00
27 T	cis-1,2-Dichloroethene	0.598	0.586	2.0	104	0.00
28 T	Bromochloromethane	0.362	0.341	5.8	101	0.00
29 T	Tetrahydrofuran	0.198	0.195	1.5	100	-0.01
30 C	Chloroform	0.995	0.934	6.1#	100	0.00
31 T	Cyclohexane	0.706	0.737	-4.4	105	0.00
32 T	1,1,1-Trichloroethane	0.895	0.864	3.5	102	0.00
33 S	1,2-Dichloroethane-d4	0.601	0.561	6.7	104	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.346	0.323	6.6	104	0.00
36 T	1,1-Dichloropropene	0.457	0.452	1.1	103	0.00
37 T	Ethyl Acetate	0.415	0.395	4.8	100	0.00
38 T	Carbon Tetrachloride	0.533	0.511	4.1	101	0.00
39 T	Methylcyclohexane	0.522	0.556	-6.5	107	0.00
40 TM	Benzene	1.373	1.355	1.3	102	0.00
41 T	Methacrylonitrile	0.217	0.216	0.5	101	0.00
42 TM	1,2-Dichloroethane	0.490	0.471	3.9	101	0.00
43 T	Isopropyl Acetate	0.643	0.640	0.5	102	0.00
44 TM	Trichloroethene	0.408	0.395	3.2	103	0.00
45 C	1,2-Dichloropropane	0.337	0.325	3.6#	101	0.00
46 T	Dibromomethane	0.256	0.244	4.7	100	0.00
47 T	Bromodichloromethane	0.522	0.501	4.0	101	0.00
48 T	Methyl methacrylate	0.313	0.321	-2.6	103	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020553.D
 Acq On : 11 Jan 2021 15:27
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011121

Quant Time: Jan 12 02:51:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	101	-0.02
50 S	Toluene-d8	1.240	1.189	4.1	104	0.00
51 T	4-Methyl-2-Pentanone	0.399	0.399	0.0	100	0.00
52 CM	Toluene	0.873	0.874	-0.1#	102	0.00
53 T	t-1,3-Dichloropropene	0.532	0.531	0.2	103	0.00
54 T	cis-1,3-Dichloropropene	0.569	0.576	-1.2	103	0.00
55 T	1,1,2-Trichloroethane	0.369	0.354	4.1	100	0.00
56 T	Ethyl methacrylate	0.478	0.502	-5.0	104	0.00
57 T	1,3-Dichloropropane	0.599	0.574	4.2	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.226	0.269	-19.0	113	0.00
59 T	2-Hexanone	0.299	0.307	-2.7	101	0.00
60 T	Dibromochloromethane	0.417	0.406	2.6	101	0.00
61 T	1,2-Dibromoethane	0.394	0.382	3.0	101	0.00
62 S	4-Bromofluorobenzene	0.438	0.435	0.7	117	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.452	0.431	4.6	102	0.00
65 PM	Chlorobenzene	1.052	1.031	2.0	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.402	0.401	0.2	110	0.00
67 C	Ethyl Benzene	1.732	1.805	-4.2#	111	0.00
68 T	m/p-Xylenes	0.668	0.696	-4.2	105	0.00
69 T	o-Xylene	0.615	0.662	-7.6	110	0.00
70 T	Styrene	1.050	1.133	-7.9	108	0.00
71 P	Bromoform	0.326	0.336	-3.1	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
73 T	Isopropylbenzene	3.371	3.403	-0.9	119	0.00
74 T	N-amyl acetate	1.072	1.077	-0.5	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.098	1.044	4.9	115	0.00
76 T	1,2,3-Trichloropropane	1.053	0.994	5.6	113	0.00
77 T	Bromobenzene	0.903	0.879	2.7	118	0.00
78 T	n-propylbenzene	3.726	3.807	-2.2	116	0.00
79 T	2-Chlorotoluene	2.299	2.252	2.0	116	0.00
80 T	1,3,5-Trimethylbenzene	2.825	2.852	-1.0	116	0.00
81 T	trans-1,4-Dichloro-2-butene	0.387	0.362	6.5	114	0.00
82 T	4-Chlorotoluene	2.728	2.697	1.1	117	0.00
83 T	tert-Butylbenzene	2.713	2.737	-0.9	116	0.00
84 T	1,2,4-Trimethylbenzene	2.824	2.919	-3.4	114	0.00
85 T	sec-Butylbenzene	3.166	3.258	-2.9	118	0.00
86 T	p-Isopropyltoluene	2.940	3.163	-7.6	120	0.00
87 T	1,3-Dichlorobenzene	1.651	1.590	3.7	116	0.00
88 T	1,4-Dichlorobenzene	1.692	1.602	5.3	117	0.00
89 T	n-Butylbenzene	2.544	2.732	-7.4	124	0.00
90 T	Hexachloroethane	0.555	0.510	8.1	113	0.00
91 T	1,2-Dichlorobenzene	1.630	1.537	5.7	116	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.258	0.249	3.5	123	0.00
93 T	1,2,4-Trichlorobenzene	1.087	1.101	-1.3	118	0.00
94 T	Hexachlorobutadiene	0.593	0.551	7.1	119	0.00
95 T	Naphthalene	3.166	3.154	0.4	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
Data File : VX020553.D
Acq On : 11 Jan 2021 15:27
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011121

Quant Time: Jan 12 02:51:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.105	1.122	-1.5	120	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6