

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
 Data File : VX020563.D
 Acq On : 11 Jan 2021 20:37
 Operator : JC/MD
 Sample : M1071-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-108

Quant Time: Jan 12 02:55:56 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	192406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	316052	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	279182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	112543	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	113316	49.03	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.06%
35) Dibromofluoromethane	5.464	113	102648	46.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.74%
50) Toluene-d8	8.695	98	375518	47.92	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.84%
62) 4-Bromofluorobenzene	11.122	95	116990	42.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.56%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011121\
Data File : VX020563.D
Acq On : 11 Jan 2021 20:37
Operator : JC/MD
Sample : M1071-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-108

Quant Time: Jan 12 02:55:56 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 5.629 min Scan# 745
Delta R.T. -0.000 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37




Instrument :
MSVOA_X
ClientSampleId :
MW-108


Tgt Ion:168 Resp: 192406
Ion Ratio Lower Upper
168 100
99 50.5 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 49.03 ug/l
RT: 6.031 min Scan# 811
Delta R.T. 0.006 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37

Tgt Ion: 65 Resp: 113316
Ion Ratio Lower Upper
65 100
67 52.2 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 6.830 min Scan# 942
Delta R.T. -0.000 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37




Instrument :
MSVOA_X
ClientSampleId :
MW-108


Tgt Ion:114 Resp: 316052
Ion Ratio Lower Upper
114 100
63 19.0 0.0 40.8
88 15.8 0.0 32.8



#35
Dibromofluoromethane
Concen: 46.87 ug/l
RT: 5.464 min Scan# 718
Delta R.T. -0.001 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37

Tgt Ion:113 Resp: 102648
Ion Ratio Lower Upper
113 100
111 102.8 81.9 122.9
192 22.3 16.7 25.1









#50
Toluene-d8
Concen: 47.92 ug/l
RT: 8.695 min Scan# 1248
Delta R.T. -0.001 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37




Instrument :
MSVOA_X
ClientSampleId :
MW-108


Tgt Ion: 98 Resp: 375518
Ion Ratio Lower Upper
98 100
100 65.8 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 42.28 ug/l
RT: 11.122 min Scan# 1646
Delta R.T. -0.000 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37

Tgt Ion: 95 Resp: 116990
Ion Ratio Lower Upper
95 100
174 85.7 0.0 154.6
176 82.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.098 min Scan# 1478
Delta R.T. 0.006 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37

Instrument :
MSVOA_X
ClientSampleId :
MW-108

Tgt Ion:117 Resp: 279182
Ion Ratio Lower Upper
117 100
82 51.4 47.4 71.2
119 31.4 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.061 min Scan# 1800
Delta R.T. -0.000 min
Lab File: VX020563.D
Acq: 11 Jan 2021 20:37

Tgt Ion:152 Resp: 112543
Ion Ratio Lower Upper
152 100
115 57.8 41.1 123.3
150 157.7 0.0 349.0

