

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
 Data File : VX026422.D
 Acq On : 11 Jan 2022 16:01
 Operator : JC/MD
 Sample : VX0111WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111WBL01

Quant Time: Jan 12 06:45:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	155264	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95165	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104751	48.296	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.600%
35) Dibromofluoromethane	5.391	113	80926	48.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.740%
50) Toluene-d8	8.653	98	300181	50.304	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	111720	52.702	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
Data File : VX026422.D
Acq On : 11 Jan 2022 16:01
Operator : JC/MD
Sample : VX0111WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0111WBL01

Quant Time: Jan 12 06:45:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

