

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
 Data File : VX026428.D
 Acq On : 11 Jan 2022 18:31
 Operator : JC/MD
 Sample : N1083-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 12 06:47:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253351	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89784	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104629	48.972	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.940%
35) Dibromofluoromethane	5.391	113	80091	48.954	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.900%
50) Toluene-d8	8.653	98	290187	49.220	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.440%
62) 4-Bromofluorobenzene	11.079	95	103500	49.418	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
Data File : VX026428.D
Acq On : 11 Jan 2022 18:31
Operator : JC/MD
Sample : N1083-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jan 12 06:47:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

