

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
 Data File : VX026429.D
 Acq On : 11 Jan 2022 18:54
 Operator : JC/MD
 Sample : N1083-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 12 06:47:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248587	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102221	49.074	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.140%
35) Dibromofluoromethane	5.391	113	78230	48.732	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.653	98	288223	49.824	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.079	95	105145	51.166	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.340%
Target Compounds						
16) Acetone	2.392	43	1591	1.498	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
Data File : VX026429.D
Acq On : 11 Jan 2022 18:54
Operator : JC/MD
Sample : N1083-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Jan 12 06:47:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

