

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011122\
 Data File : VX026421.D
 Acq On : 11 Jan 2022 15:01
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011122

Quant Time: Jan 12 06:45:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	112	0.00
2 T	Dichlorodifluoromethane	50.000	47.725	4.5	106	0.00
3 P	Chloromethane	50.000	47.264	5.5	108	0.00
4 C	Vinyl Chloride	50.000	48.680	2.6#	106	0.00
5 T	Bromomethane	50.000	50.149	-0.3	114	0.00
6 T	Chloroethane	50.000	47.355	5.3	106	0.00
7 T	Trichlorofluoromethane	50.000	47.148	5.7	101	0.00
8 T	Diethyl Ether	50.000	47.774	4.5	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.841	2.3	106	0.00
10 T	Methyl Iodide	50.000	44.350	11.3	101	0.00
11 T	Tert butyl alcohol	250.000	270.981	-8.4	115	0.00
12 CM	1,1-Dichloroethene	50.000	47.663	4.7#	103	0.00
13 T	Acrolein	250.000	225.957	9.6	106	0.00
14 T	Allyl chloride	50.000	48.635	2.7	103	0.00
15 T	Acrylonitrile	250.000	236.637	5.3	104	0.00
16 T	Acetone	250.000	220.879	11.6	96	0.00
17 T	Carbon Disulfide	50.000	47.620	4.8	107	0.00
18 T	Methyl Acetate	50.000	46.193	7.6	103	0.00
19 T	Methyl tert-butyl Ether	50.000	48.034	3.9	105	0.00
20 T	Methylene Chloride	50.000	46.075	7.8	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.804	4.4	104	0.00
22 T	Diisopropyl ether	50.000	47.991	4.0	104	0.00
23 T	Vinyl Acetate	250.000	250.828	-0.3	105	0.00
24 P	1,1-Dichloroethane	50.000	47.823	4.4	103	0.00
25 T	2-Butanone	250.000	231.927	7.2	99	0.00
26 T	2,2-Dichloropropane	50.000	51.131	-2.3	106	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.319	3.4	103	0.00
28 T	Bromochloromethane	50.000	48.042	3.9	103	0.00
29 T	Tetrahydrofuran	250.000	233.890	6.4	102	0.00
30 C	Chloroform	50.000	48.011	4.0#	104	0.00
31 T	Cyclohexane	50.000	47.490	5.0	105	0.00
32 T	1,1,1-Trichloroethane	50.000	49.499	1.0	104	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.025	8.0	100	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	111	0.00
35 S	Dibromofluoromethane	50.000	46.940	6.1	99	0.00
36 T	1,1-Dichloropropene	50.000	48.690	2.6	105	0.00
37 T	Ethyl Acetate	50.000	48.992	2.0	105	0.00
38 T	Carbon Tetrachloride	50.000	49.455	1.1	103	0.00
39 T	Methylcyclohexane	50.000	50.199	-0.4	107	0.00
40 TM	Benzene	50.000	48.629	2.7	105	0.00
41 T	Methacrylonitrile	50.000	50.364	-0.7	107	0.00
42 TM	1,2-Dichloroethane	50.000	48.389	3.2	104	0.00
43 T	Isopropyl Acetate	50.000	50.391	-0.8	105	0.00
44 TM	Trichloroethene	50.000	49.604	0.8	105	0.00
45 C	1,2-Dichloropropane	50.000	49.862	0.3#	106	0.00
46 T	Dibromomethane	50.000	49.571	0.9	106	0.00
47 T	Bromodichloromethane	50.000	50.327	-0.7	105	0.00
48 T	Methyl methacrylate	50.000	49.546	0.9	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011122\
 Data File : VX026421.D
 Acq On : 11 Jan 2022 15:01
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011122

Quant Time: Jan 12 06:45:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	966.131	3.4	101	0.00
50 S	Toluene-d8	50.000	47.925	4.2	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	243.673	2.5	103	0.00
52 CM	Toluene	50.000	49.679	0.6#	106	0.00
53 T	t-1,3-Dichloropropene	50.000	48.357	3.3	109	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.253	-6.5	107	0.00
55 T	1,1,2-Trichloroethane	50.000	48.965	2.1	105	0.00
56 T	Ethyl methacrylate	50.000	51.316	-2.6	105	0.00
57 T	1,3-Dichloropropane	50.000	50.020	-0.0	107	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	228.212	8.7	96	0.00
59 T	2-Hexanone	250.000	244.343	2.3	102	0.00
60 T	Dibromochloromethane	50.000	51.929	-3.9	106	0.00
61 T	1,2-Dibromoethane	50.000	50.126	-0.3	106	0.00
62 S	4-Bromofluorobenzene	50.000	48.325	3.3	103	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	113	0.00
64 T	Tetrachloroethene	50.000	48.196	3.6	106	0.00
65 PM	Chlorobenzene	50.000	48.519	3.0	107	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.579	-1.2	109	0.00
67 C	Ethyl Benzene	50.000	49.847	0.3#	107	0.00
68 T	m/p-Xylenes	100.000	99.586	0.4	107	0.00
69 T	o-Xylene	50.000	49.293	1.4	106	0.00
70 T	Styrene	50.000	50.536	-1.1	106	0.00
71 P	Bromoform	50.000	46.287	7.4	107	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	112	0.00
73 T	Isopropylbenzene	50.000	48.799	2.4	107	0.00
74 T	N-ethyl acetate	50.000	51.626	-3.3	106	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.087	3.8	106	0.00
76 T	1,2,3-Trichloropropane	50.000	48.302	3.4	104	0.00
77 T	Bromobenzene	50.000	47.987	4.0	105	0.00
78 T	n-propylbenzene	50.000	49.685	0.6	108	0.00
79 T	2-Chlorotoluene	50.000	48.003	4.0	107	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.887	2.2	107	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.161	7.7	107	0.00
82 T	4-Chlorotoluene	50.000	48.140	3.7	106	0.00
83 T	tert-Butylbenzene	50.000	49.987	0.0	108	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.899	2.2	106	0.00
85 T	sec-Butylbenzene	50.000	50.043	-0.1	107	0.00
86 T	p-Isopropyltoluene	50.000	50.680	-1.4	108	0.00
87 T	1,3-Dichlorobenzene	50.000	48.387	3.2	108	0.00
88 T	1,4-Dichlorobenzene	50.000	48.360	3.3	107	0.00
89 T	n-Butylbenzene	50.000	51.703	-3.4	109	0.00
90 T	Hexachloroethane	50.000	47.222	5.6	108	0.00
91 T	1,2-Dichlorobenzene	50.000	48.588	2.8	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	49.733	0.5	103	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.813	-1.6	109	0.00
94 T	Hexachlorobutadiene	50.000	51.333	-2.7	113	0.00
95 T	Naphthalene	50.000	50.552	-1.1	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
Data File : VX026421.D
Acq On : 11 Jan 2022 15:01
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011122

Quant Time: Jan 12 06:45:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.991	0.0	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6