

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011122\
 Data File : VX026433.D
 Acq On : 11 Jan 2022 20:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 12 06:48:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	49.092	1.8	98	0.00
3 P	Chloromethane	50.000	49.922	0.2	103	0.00
4 C	Vinyl Chloride	50.000	50.455	-0.9#	99	0.00
5 T	Bromomethane	50.000	49.420	1.2	101	0.00
6 T	Chloroethane	50.000	50.244	-0.5	101	0.00
7 T	Trichlorofluoromethane	50.000	48.961	2.1	95	0.00
8 T	Diethyl Ether	50.000	50.418	-0.8	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.187	1.6	96	0.00
10 T	Methyl Iodide	50.000	52.143	-4.3	109	0.00
11 T	Tert butyl alcohol	250.000	279.077	-11.6	106	0.00
12 CM	1,1-Dichloroethene	50.000	48.485	3.0#	95	0.00
13 T	Acrolein	250.000	261.722	-4.7	111	0.00
14 T	Allyl chloride	50.000	49.798	0.4	95	0.00
15 T	Acrylonitrile	250.000	251.937	-0.8	100	0.00
16 T	Acetone	250.000	208.998	16.4	81	0.00
17 T	Carbon Disulfide	50.000	47.476	5.0	96	0.00
18 T	Methyl Acetate	50.000	49.507	1.0	100	0.00
19 T	Methyl tert-butyl Ether	50.000	49.759	0.5	97	0.00
20 T	Methylene Chloride	50.000	48.520	3.0	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.057	1.9	96	0.00
22 T	Diisopropyl ether	50.000	50.131	-0.3	98	0.00
23 T	Vinyl Acetate	250.000	264.646	-5.9	100	0.00
24 P	1,1-Dichloroethane	50.000	49.881	0.2	97	0.00
25 T	2-Butanone	250.000	238.343	4.7	92	0.00
26 T	2,2-Dichloropropane	50.000	49.323	1.4	92	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.508	1.0	95	0.00
28 T	Bromochloromethane	50.000	51.222	-2.4	98	0.00
29 T	Tetrahydrofuran	250.000	258.253	-3.3	102	0.00
30 C	Chloroform	50.000	49.639	0.7#	97	0.00
31 T	Cyclohexane	50.000	48.585	2.8	97	0.00
32 T	1,1,1-Trichloroethane	50.000	50.568	-1.1	96	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.704	0.6	97	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	49.408	1.2	95	0.00
36 T	1,1-Dichloropropene	50.000	49.102	1.8	97	0.00
37 T	Ethyl Acetate	50.000	52.316	-4.6	102	0.00
38 T	Carbon Tetrachloride	50.000	49.828	0.3	95	0.00
39 T	Methylcyclohexane	50.000	50.114	-0.2	98	0.00
40 TM	Benzene	50.000	49.673	0.7	98	0.00
41 T	Methacrylonitrile	50.000	51.906	-3.8	101	0.00
42 TM	1,2-Dichloroethane	50.000	49.873	0.3	98	0.00
43 T	Isopropyl Acetate	50.000	53.117	-6.2	101	0.00
44 TM	Trichloroethene	50.000	49.523	1.0	96	0.00
45 C	1,2-Dichloropropane	50.000	51.326	-2.7#	99	0.00
46 T	Dibromomethane	50.000	51.246	-2.5	100	0.00
47 T	Bromodichloromethane	50.000	51.507	-3.0	98	0.00
48 T	Methyl methacrylate	50.000	52.691	-5.4	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX011122\
 Data File : VX026433.D
 Acq On : 11 Jan 2022 20:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 12 06:48:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	964.579	3.5	92	0.00
50 S	Toluene-d8	50.000	50.259	-0.5	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	262.167	-4.9	101	0.00
52 CM	Toluene	50.000	50.253	-0.5#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	48.322	3.4	100	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.976	-6.0	97	0.00
55 T	1,1,2-Trichloroethane	50.000	50.558	-1.1	99	0.00
56 T	Ethyl methacrylate	50.000	53.485	-7.0	100	0.00
57 T	1,3-Dichloropropane	50.000	51.126	-2.3	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	263.872	-5.5	101	0.00
59 T	2-Hexanone	250.000	254.492	-1.8	96	0.00
60 T	Dibromochloromethane	50.000	53.103	-6.2	99	0.00
61 T	1,2-Dibromoethane	50.000	50.742	-1.5	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.766	-1.5	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	101	0.00
64 T	Tetrachloroethene	50.000	47.736	4.5	94	0.00
65 PM	Chlorobenzene	50.000	49.785	0.4	98	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.874	-3.7	100	0.00
67 C	Ethyl Benzene	50.000	51.383	-2.8#	98	0.00
68 T	m/p-Xylenes	100.000	103.463	-3.5	99	0.00
69 T	o-Xylene	50.000	52.016	-4.0	100	0.00
70 T	Styrene	50.000	52.736	-5.5	99	0.00
71 P	Bromoform	50.000	47.891	4.2	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	51.179	-2.4	99	0.00
74 T	N-amyl acetate	50.000	56.686	-13.4	103	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.239	-4.5	101	0.00
76 T	1,2,3-Trichloropropane	50.000	52.492	-5.0	99	0.00
77 T	Bromobenzene	50.000	50.138	-0.3	97	0.00
78 T	n-propylbenzene	50.000	51.391	-2.8	98	0.00
79 T	2-Chlorotoluene	50.000	50.064	-0.1	98	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.655	-3.3	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.269	3.5	99	0.00
82 T	4-Chlorotoluene	50.000	50.746	-1.5	98	0.00
83 T	tert-Butylbenzene	50.000	52.193	-4.4	99	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.246	-2.5	98	0.00
85 T	sec-Butylbenzene	50.000	51.696	-3.4	97	0.00
86 T	p-Isopropyltoluene	50.000	52.225	-4.5	98	0.00
87 T	1,3-Dichlorobenzene	50.000	49.515	1.0	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.771	2.5	95	0.00
89 T	n-Butylbenzene	50.000	52.217	-4.4	97	0.00
90 T	Hexachloroethane	50.000	48.503	3.0	98	0.00
91 T	1,2-Dichlorobenzene	50.000	50.464	-0.9	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	56.646	-13.3	104	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.450	1.1	93	0.00
94 T	Hexachlorobutadiene	50.000	48.844	2.3	94	0.00
95 T	Naphthalene	50.000	52.495	-5.0	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011122\
Data File : VX026433.D
Acq On : 11 Jan 2022 20:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jan 12 06:48:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.974	-1.9	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6