

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011123\
 Data File : VX033697.D
 Acq On : 11 Jan 2023 15:59
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX011123

Quant Time: Jan 12 00:23:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	198409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	278316	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131521	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93624	44.816	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.640%
35) Dibromofluoromethane	5.385	113	83987	48.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.200%
50) Toluene-d8	8.646	98	310573	49.323	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.640%
62) 4-Bromofluorobenzene	11.079	95	115987	49.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.220%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	76285	48.653	ug/l	98
3) Chloromethane	1.294	50	89271	51.545	ug/l	100
4) Vinyl Chloride	1.373	62	88926	52.908	ug/l	98
5) Bromomethane	1.617	94	48054	49.760	ug/l	97
6) Chloroethane	1.690	64	50754	48.843	ug/l	98
7) Trichlorofluoromethane	1.885	101	124541	45.691	ug/l	99
8) Diethyl Ether	2.129	74	42518	47.640	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	83239	52.948	ug/l	95
10) Methyl Iodide	2.452	142	115289	52.313	ug/l	94
11) Tert butyl alcohol	3.026	59	46706	128.699	ug/l #	95
12) 1,1-Dichloroethene	2.318	96	78625	48.702	ug/l	96
13) Acrolein	2.245	56	84617	252.707	ug/l	99
14) Allyl chloride	2.666	41	147553	48.458	ug/l	89
15) Acrylonitrile	3.068	53	216236	226.401	ug/l	99
16) Acetone	2.398	43	221086	293.357	ug/l	94
17) Carbon Disulfide	2.513	76	208285	50.943	ug/l	99
18) Methyl Acetate	2.708	43	129532	44.429	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	262036	46.733	ug/l	100
20) Methylene Chloride	2.788	84	90649	49.900	ug/l	97
21) trans-1,2-Dichloroethene	3.093	96	86825	46.005	ug/l	98
22) Diisopropyl ether	3.757	45	276745	45.172	ug/l	95
23) Vinyl Acetate	3.720	43	1083969	235.035	ug/l	98
24) 1,1-Dichloroethane	3.611	63	153801	45.886	ug/l	100
25) 2-Butanone	4.568	43	306809	224.774	ug/l	97
26) 2,2-Dichloropropane	4.476	77	113997	47.480	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	99507	45.860	ug/l	99
28) Bromochloromethane	4.897	49	67865	44.449	ug/l	97
29) Tetrahydrofuran	5.013	42	186976	217.341	ug/l	99
30) Chloroform	5.092	83	152534	44.431	ug/l	99
31) Cyclohexane	5.470	56	146165	47.548	ug/l	99
32) 1,1,1-Trichloroethane	5.385	97	135773	46.322	ug/l	98
36) 1,1-Dichloropropene	5.690	75	118104	46.825	ug/l	99
37) Ethyl Acetate	4.714	43	116453	48.290	ug/l	99
38) Carbon Tetrachloride	5.677	117	124545	47.676	ug/l	97
39) Methylcyclohexane	7.378	83	155081	49.789	ug/l	96
40) Benzene	6.031	78	348969	47.468	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011123\
 Data File : VX033697.D
 Acq On : 11 Jan 2023 15:59
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011123

Quant Time: Jan 12 00:23:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	66697	47.986	ug/l	99
42) 1,2-Dichloroethane	6.080	62	118754	45.581	ug/l	100
43) Isopropyl Acetate	6.336	43	189083	47.487	ug/l	99
44) Trichloroethene	7.122	130	101287	49.269	ug/l	97
45) 1,2-Dichloropropane	7.427	63	87895	47.727	ug/l	95
46) Dibromomethane	7.580	93	60783	47.252	ug/l	98
47) Bromodichloromethane	7.817	83	118090	47.635	ug/l	96
48) Methyl methacrylate	7.689	41	97551	48.635	ug/l	99
49) 1,4-Dioxane	7.744	88	42457	956.941	ug/l #	99
51) 4-Methyl-2-Pentanone	8.573	43	574414	234.895	ug/l	100
52) Toluene	8.714	92	225584	47.516	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	127030	51.238	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	141568	50.961	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	90020	47.902	ug/l	97
56) Ethyl methacrylate	9.116	69	135193	51.118	ug/l	97
57) 1,3-Dichloropropane	9.305	76	147624	48.445	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	316105	242.689	ug/l	99
59) 2-Hexanone	9.427	43	430136	252.717	ug/l	100
60) Dibromochloromethane	9.518	129	100611	51.640	ug/l	99
61) 1,2-Dibromoethane	9.610	107	94549	50.413	ug/l	98
64) Tetrachloroethene	9.274	164	90429	51.263	ug/l	98
65) Chlorobenzene	10.079	112	249194	48.007	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.164	131	94088	49.253	ug/l	100
67) Ethyl Benzene	10.195	91	433232	48.610	ug/l	99
68) m/p-Xylenes	10.299	106	347085	99.567	ug/l	100
69) o-Xylene	10.640	106	169234	48.954	ug/l	100
70) Styrene	10.652	104	283084	50.350	ug/l	100
71) Bromoform	10.799	173	78127	48.806	ug/l #	100
73) Isopropylbenzene	10.963	105	438665	50.723	ug/l	100
74) N-amyl acetate	10.841	43	164006	47.488	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.213	83	125067	49.386	ug/l	100
76) 1,2,3-Trichloropropane	11.237	75	135331	56.003	ug/l	83
77) Bromobenzene	11.195	156	117064	50.135	ug/l	99
78) n-propylbenzene	11.305	91	504337	46.993	ug/l	100
79) 2-Chlorotoluene	11.365	91	294047	50.763	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	368532	46.567	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	38121	49.028	ug/l	95
82) 4-Chlorotoluene	11.451	91	338115	51.015	ug/l	99
83) tert-Butylbenzene	11.713	119	376623	47.929	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	368669	49.246	ug/l	99
85) sec-Butylbenzene	11.890	105	463335	49.298	ug/l	100
86) p-Isopropyltoluene	12.012	119	403323	50.123	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	215145	48.413	ug/l	100
88) 1,4-Dichlorobenzene	12.042	146	214468	47.375	ug/l	99
89) n-Butylbenzene	12.335	91	334257	50.776	ug/l	99
90) Hexachloroethane	12.536	117	66712	50.400	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	206314	48.111	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.944	75	25303	47.147	ug/l	98
93) 1,2,4-Trichlorobenzene	13.591	180	151601	52.482	ug/l	99
94) Hexachlorobutadiene	13.725	225	72965	51.566	ug/l	99
95) Naphthalene	13.774	128	419247	50.448	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	147451	51.014	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011123\
Data File : VX033697.D
Acq On : 11 Jan 2023 15:59
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123

Quant Time: Jan 12 00:23:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.549 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:168 Resp: 198409
Ion Ratio Lower Upper
168 100
99 47.3 39.6 59.4



#2
Dichlorodifluoromethane
Concen: 48.653 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 85 Resp: 76285
Ion Ratio Lower Upper
85 100
87 32.8 15.8 47.4









#3
Chloromethane
Concen: 51.545 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123

Tgt Ion: 50 Resp: 89271
Ion Ratio Lower Upper
50 100
52 31.4 25.2 37.8






#4
Vinyl Chloride
Concen: 52.908 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 62 Resp: 88926
Ion Ratio Lower Upper
62 100
64 31.5 26.2 39.4



Abundance Scan 87 (1.618 min): VX033693.D\data.ms (-81)



#5

Bromomethane

Concen: 49.760 ug/l

RT: 1.617 min Scan# 81

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :

ICVVX011123


Tgt Ion: 94 Resp: 48054

Ion Ratio Lower Upper




94 100

96 95.8 74.3 111.5

Abundance



Abundance Scan 99 (1.691 min): VX033693.D\data.ms (-90)



#6

Chloroethane

Concen: 48.843 ug/l

RT: 1.690 min Scan# 99

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59


Tgt Ion: 64 Resp: 50754


Ion Ratio Lower Upper

64 100

66 31.7 26.1 39.1

Abundance









#7
Trichlorofluoromethane
Concen: 45.691 ug/l
RT: 1.885 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:101 Resp: 124541
Ion Ratio Lower Upper
101 100
103 63.6 50.6 75.8



#8
Diethyl Ether
Concen: 47.640 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 74 Resp: 42518
Ion Ratio Lower Upper
74 100
45 103.1 51.5 154.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.948 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:101 Resp: 83239
Ion Ratio Lower Upper
101 100
85 41.3 36.7 55.1
151 87.0 66.5 99.7



#10
Methyl Iodide
Concen: 52.313 ug/l
RT: 2.452 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:142 Resp: 115289
Ion Ratio Lower Upper
142 100
127 41.6 37.3 55.9
141 14.5 11.5 17.3









#11
Tert butyl alcohol
Concen: 128.699 ug/l
RT: 3.026 min Scan# 317
Delta R.T. 0.006 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 59 Resp: 46706
Ion Ratio Lower Upper
59 100
57 6.1 3.7 5.5#



#12
1,1-Dichloroethene
Concen: 48.702 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 96 Resp: 78625
Ion Ratio Lower Upper
96 100
61 161.7 134.2 201.4
98 66.0 50.2 75.2









#13
Acrolein
Concen: 252.707 ug/l
RT: 2.245 min Scan# 189
Delta R.T. 0.006 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 56 Resp: 84617
Ion Ratio Lower Upper
56 100
55 71.1 56.3 84.5



#14
Allyl chloride
Concen: 48.458 ug/l
RT: 2.666 min Scan# 259
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59


Tgt Ion: 41 Resp: 147553
Ion Ratio Lower Upper
41 100
39 61.4 59.0 88.4
76 31.9 24.6 37.0








#15
Acrylonitrile
Concen: 226.401 ug/l
RT: 3.068 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123







Tgt Ion: 53 Resp: 216236
Ion Ratio Lower Upper
53 100
52 80.8 65.7 98.5
51 35.3 29.0 43.4



#16
Acetone
Concen: 293.357 ug/l
RT: 2.398 min Scan# 215
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 43 Resp: 221086
Ion Ratio Lower Upper
43 100
58 31.6 22.6 33.8









#17
Carbon Disulfide
Concen: 50.943 ug/l
RT: 2.513 min Scan# 21
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 76 Resp: 208285
Ion Ratio Lower Upper
76 100
78 9.3 7.0 10.6



#18
Methyl Acetate
Concen: 44.429 ug/l
RT: 2.708 min Scan# 266
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 43 Resp: 129532
Ion Ratio Lower Upper
43 100
74 23.7 18.2 27.4









#19
Methyl tert-butyl Ether
Concen: 46.733 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVWX011123


Tgt Ion: 73 Resp: 262036
Ion Ratio Lower Upper
73 100
57 21.6 17.4 26.0



#20
Methylene Chloride
Concen: 49.900 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 84 Resp: 90649
Ion Ratio Lower Upper
84 100
49 121.6 101.1 151.7
51 36.3 31.0 46.6
86 66.4 52.6 79.0









#21
trans-1,2-Dichloroethene
Concen: 46.005 ug/l
RT: 3.093 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 96 Resp: 86825
Ion Ratio Lower Upper
96 100
61 131.9 106.7 160.1
98 65.1 50.0 75.0



#22
Diisopropyl ether
Concen: 45.172 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 45 Resp: 276745
Ion Ratio Lower Upper
45 100
43 75.1 64.6 97.0
87 28.3 21.9 32.9
59 11.2 9.2 13.8









#23
Vinyl Acetate
Concen: 235.035 ug/l
RT: 3.720 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 43 Resp: 1083969
Ion Ratio Lower Upper
43 100
86 11.2 8.4 12.6



#24
1,1-Dichloroethane
Concen: 45.886 ug/l
RT: 3.611 min Scan# 414
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 63 Resp: 153801
Ion Ratio Lower Upper
63 100
98 7.7 3.8 11.2
100 4.9 2.4 7.2





#25

2-Butanone

Concen: 224.774 ug/l

RT: 4.568 min Scan# 571

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :





ICVVX011123

Tgt Ion: 43 Resp: 306809

Ion Ratio Lower Upper

43 100

72 25.6 19.4 29.2



#26

2,2-Dichloropropane

Concen: 47.480 ug/l

RT: 4.476 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX033697.D




Acq: 11 Jan 2023 15:59


Tgt Ion: 77 Resp: 113997

Ion Ratio Lower Upper

77 100

97 24.1 11.9 35.5









#27
cis-1,2-Dichloroethene
Concen: 45.860 ug/l
RT: 4.489 min Scan# 511
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 96 Resp: 99507
Ion Ratio Lower Upper
96 100
61 137.8 0.0 277.6
98 65.1 0.0 128.0



#28
Bromochloromethane
Concen: 44.449 ug/l
RT: 4.897 min Scan# 625
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 49 Resp: 67865
Ion Ratio Lower Upper
49 100
129 2.2 0.0 4.6
130 88.4 68.5 102.7









#29
Tetrahydrofuran
Concen: 217.341 ug/l
RT: 5.013 min Scan# 644
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVX011123


Tgt Ion: 42 Resp: 186976
Ion Ratio Lower Upper
42 100
72 44.3 34.8 52.2
71 41.3 32.2 48.2



#30
Chloroform
Concen: 44.431 ug/l
RT: 5.092 min Scan# 657
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 83 Resp: 152534
Ion Ratio Lower Upper
83 100
85 66.0 52.1 78.1









#31
Cyclohexane
Concen: 47.548 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 56 Resp: 146165
Ion Ratio Lower Upper
56 100
69 31.5 24.6 37.0
84 88.1 71.0 106.6



#32
1,1,1-Trichloroethane
Concen: 46.322 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 97 Resp: 135773
Ion Ratio Lower Upper
97 100
99 64.4 52.3 78.5
61 41.9 35.2 52.8









#33
1,2-Dichloroethane-d4
Concen: 44.816 ug/l
RT: 5.952 min Scan# 79
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 65 Resp: 93624
Ion Ratio Lower Upper
65 100
67 53.6 0.0 104.8



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.006 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 114 Resp: 278316
Ion Ratio Lower Upper
114 100
63 19.3 0.0 37.4
88 15.9 0.0 30.6









#35
Dibromofluoromethane
Concen: 48.598 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:113 Resp: 83987
Ion Ratio Lower Upper
113 100
111 104.0 83.0 124.6
192 24.5 19.0 28.4



#36
1,1-Dichloropropene
Concen: 46.825 ug/l
RT: 5.690 min Scan# 755
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 75 Resp: 118104
Ion Ratio Lower Upper
75 100
110 38.4 18.8 56.4
77 31.2 25.1 37.7





#37

Ethyl Acetate

Concen: 48.290 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.001 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :

ICVVX011123





Tgt Ion: 43 Resp: 116453

Ion Ratio Lower Upper

43 100

61 14.3 11.1 16.7

70 10.7 8.5 12.7



#38

Carbon Tetrachloride

Concen: 47.676 ug/l

RT: 5.677 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59




Tgt Ion: 117 Resp: 124545


Ion Ratio Lower Upper

117 100

119 93.4 78.0 117.0

121 30.5 24.4 36.6









#39
Methylcyclohexane
Concen: 49.789 ug/l
RT: 7.378 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 83 Resp: 155081
Ion Ratio Lower Upper
83 100
55 73.3 61.8 92.6
98 50.7 39.3 58.9



#40
Benzene
Concen: 47.468 ug/l
RT: 6.031 min Scan# 811
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 78 Resp: 348969
Ion Ratio Lower Upper
78 100
77 22.7 19.0 28.4









#41
Methacrylonitrile
Concen: 47.986 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 41 Resp: 66697
Ion Ratio Lower Upper
41 100
39 53.5 43.8 65.6
67 68.9 54.8 82.2
52 28.1 23.0 34.6



#42
1,2-Dichloroethane
Concen: 45.581 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 62 Resp: 118754
Ion Ratio Lower Upper
62 100
98 10.2 0.0 20.2









#43
Isopropyl Acetate
Concen: 47.487 ug/l
RT: 6.336 min Scan# 80
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 43 Resp: 189083
Ion Ratio Lower Upper
43 100
61 20.4 15.7 23.5
87 13.5 10.6 15.8



#44
Trichloroethene
Concen: 49.269 ug/l
RT: 7.122 min Scan# 990
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:130 Resp: 101287
Ion Ratio Lower Upper
130 100
95 94.2 0.0 182.6





#45

1,2-Dichloropropane

Concen: 47.727 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :





ICVVX011123

Tgt Ion: 63 Resp: 87895

Ion Ratio Lower Upper

63 100

65 33.0 24.2 36.2



#46

Dibromomethane

Concen: 47.252 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59




Tgt Ion: 93 Resp: 60783


Ion Ratio Lower Upper

93 100

95 84.2 67.6 101.4

174 124.6 97.4 146.0









#47
Bromodichloromethane
Concen: 47.635 ug/l
RT: 7.817 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 83 Resp: 118090
Ion Ratio Lower Upper
83 100
85 65.3 49.8 74.6
127 9.6 7.0 10.4



#48
Methyl methacrylate
Concen: 48.635 ug/l
RT: 7.689 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 41 Resp: 97551
Ion Ratio Lower Upper
41 100
69 80.2 63.1 94.7
39 53.0 43.2 64.8









#49
1,4-Dioxane
Concen: 956.941 ug/l
RT: 7.744 min Scan# 1091
Delta R.T. 0.006 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 88 Resp: 42457
Ion Ratio Lower Upper
88 100
43 22.6 0.0 0.0#
58 69.8 56.5 84.7



#50
Toluene-d8
Concen: 49.323 ug/l
RT: 8.646 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 98 Resp: 310573
Ion Ratio Lower Upper
98 100
100 67.7 53.6 80.4









#51
4-Methyl-2-Pentanone
Concen: 234.895 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVX011123


Tgt Ion: 43 Resp: 574414
Ion Ratio Lower Upper
43 100
58 39.1 31.2 46.8



#52
Toluene
Concen: 47.516 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.007 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 92 Resp: 225584
Ion Ratio Lower Upper
92 100
91 169.8 135.5 203.3





#53

t-1,3-Dichloropropene

Concen: 51.238 ug/l

RT: 8.976 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :





ICVVX011123

Tgt Ion: 75 Resp: 127030

Ion Ratio Lower Upper

75 100

77 30.2 25.5 38.3



#54

cis-1,3-Dichloropropene

Concen: 50.961 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59




Tgt Ion: 75 Resp: 141568


Ion Ratio Lower Upper

75 100

77 31.6 25.0 37.6


39 46.2 38.6 58.0








#55
1,1,2-Trichloroethane
Concen: 47.902 ug/l
RT: 9.152 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Instrument : MSVOA_X
ClientSampleId : ICVVX011123



Tgt Ion: 97 Resp: 90020




Ion	Ratio	Lower	Upper
97	100		
83	79.2	67.4	101.0
85	53.8	43.2	64.8
99	63.2	50.2	75.2




#56
Ethyl methacrylate
Concen: 51.118 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 69 Resp: 135193

Ion	Ratio	Lower	Upper
69	100		
41	67.3	55.1	82.7
39	33.2	28.5	42.7









#57
1,3-Dichloropropane
Concen: 48.445 ug/l
RT: 9.305 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 76 Resp: 147624
Ion Ratio Lower Upper
76 100
78 32.1 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 242.689 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 63 Resp: 316105
Ion Ratio Lower Upper
63 100
106 31.1 24.5 36.7





#59

2-Hexanone

Concen: 252.717 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.007 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :





ICVX011123

Tgt Ion: 43 Resp: 430136

Ion Ratio Lower Upper

43 100

58 55.1 27.4 82.0



#60

Dibromochloromethane

Concen: 51.640 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX033697.D




Acq: 11 Jan 2023 15:59


Tgt Ion:129 Resp: 100611

Ion Ratio Lower Upper

129 100

127 76.7 38.8 116.3









#61
1,2-Dibromoethane
Concen: 50.413 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVWX011123


Tgt Ion:107 Resp: 94549
Ion Ratio Lower Upper
107 100
109 94.9 74.2 111.2



#62
4-Bromofluorobenzene
Concen: 49.613 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 95 Resp: 115987
Ion Ratio Lower Upper
95 100
174 94.2 0.0 184.4
176 90.9 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:117 Resp: 250739
Ion Ratio Lower Upper
117 100
82 51.6 40.6 60.8
119 32.3 25.6 38.4



#64
Tetrachloroethene
Concen: 51.263 ug/l
RT: 9.274 min Scan# 1343
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:164 Resp: 90429
Ion Ratio Lower Upper
164 100
166 126.7 100.2 150.2
129 83.7 69.2 103.8
131 82.2 67.2 100.8









#65
Chlorobenzene
Concen: 48.007 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:112 Resp: 249194
Ion Ratio Lower Upper
112 100
114 32.9 26.2 39.4



#66
1,1,1,2-Tetrachloroethane
Concen: 49.253 ug/l
RT: 10.164 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:131 Resp: 94088
Ion Ratio Lower Upper
131 100
133 96.2 48.4 145.0
119 63.7 31.9 95.7









#67
Ethyl Benzene
Concen: 48.610 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 91 Resp: 433232
Ion Ratio Lower Upper
91 100
106 32.0 25.8 38.8



#68
m/p-Xylenes
Concen: 99.567 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 106 Resp: 347085
Ion Ratio Lower Upper
106 100
91 194.8 156.4 234.6









#69
o-Xylene
Concen: 48.954 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument : MSVOA_X
ClientSampleId : ICVVX011123


Tgt Ion:106 Resp: 169234
Ion Ratio Lower Upper
106 100
91 205.7 103.1 309.3



#70
Styrene
Concen: 50.350 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:104 Resp: 283084
Ion Ratio Lower Upper
104 100
78 46.8 37.5 56.3
103 54.1 43.3 64.9









#71
Bromoform
Concen: 48.806 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:173 Resp: 78127
Ion Ratio Lower Upper
173 100
175 48.4 24.1 72.3
254 0.1 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:152 Resp: 131521
Ion Ratio Lower Upper
152 100
115 75.6 38.7 116.1
150 172.0 0.0 348.2









#73
Isopropylbenzene
Concen: 50.723 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:105 Resp: 438665
Ion Ratio Lower Upper
105 100
120 27.3 13.6 40.8



#74
N-ethyl acetate
Concen: 47.488 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 43 Resp: 164006
Ion Ratio Lower Upper
43 100
70 43.5 34.8 52.2
55 23.9 19.4 29.0
61 24.8 19.4 29.2









#75
1,1,2,2-Tetrachloroethane
Concen: 49.386 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 83 Resp: 125067
Ion Ratio Lower Upper
83 100
131 11.3 5.5 16.7
85 64.8 32.3 96.9



#76
1,2,3-Trichloropropane
Concen: 56.003 ug/l
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 75 Resp: 135331
Ion Ratio Lower Upper
75 100
77 32.5 21.8 65.4









#77
Bromobenzene
Concen: 50.135 ug/l
RT: 11.195 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:156 Resp: 117064
Ion Ratio Lower Upper
156 100
77 136.3 68.1 204.3
158 97.9 48.0 144.2



#78
n-propylbenzene
Concen: 46.993 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 91 Resp: 504337
Ion Ratio Lower Upper
91 100
120 24.4 12.3 36.8









#79
2-Chlorotoluene
Concen: 50.763 ug/l
RT: 11.365 min Scan# 11
Delta R.T. -0.001 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 91 Resp: 294047
Ion Ratio Lower Upper
91 100
126 36.8 18.1 54.1



#80
1,3,5-Trimethylbenzene
Concen: 46.567 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 105 Resp: 368532
Ion Ratio Lower Upper
105 100
120 51.0 25.6 76.8









#81
trans-1,4-Dichloro-2-butene
Concen: 49.028 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion: 75 Resp: 38121
Ion Ratio Lower Upper
75 100
53 97.8 83.0 124.6
89 45.7 37.7 56.5



#82
4-Chlorotoluene
Concen: 51.015 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 91 Resp: 338115
Ion Ratio Lower Upper
91 100
126 32.0 15.8 47.4









#83
tert-Butylbenzene
Concen: 47.929 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:119 Resp: 376623
Ion Ratio Lower Upper
119 100
91 53.6 26.9 80.7
134 24.2 12.3 36.8



#84
1,2,4-Trimethylbenzene
Concen: 49.246 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:105 Resp: 368669
Ion Ratio Lower Upper
105 100
120 46.4 23.4 70.2









#85
 sec-Butylbenzene
 Concen: 49.298 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX033697.D
 Acq: 11 Jan 2023 15:59




Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX011123


Tgt Ion:105 Resp: 463335
 Ion Ratio Lower Upper
 105 100
 134 21.6 10.8 32.3



#86
 p-Isopropyltoluene
 Concen: 50.123 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.006 min
 Lab File: VX033697.D
 Acq: 11 Jan 2023 15:59

Tgt Ion:119 Resp: 403323
 Ion Ratio Lower Upper
 119 100
 134 27.3 13.6 40.8
 91 21.8 11.0 33.0









#87
1,3-Dichlorobenzene
Concen: 48.413 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:146 Resp: 215145
Ion Ratio Lower Upper
146 100
111 36.7 18.6 55.6
148 64.0 31.9 95.5



#88
1,4-Dichlorobenzene
Concen: 47.375 ug/l
RT: 12.042 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:146 Resp: 214468
Ion Ratio Lower Upper
146 100
111 36.4 17.8 53.4
148 64.1 32.1 96.3





#89

n-Butylbenzene

Concen: 50.776 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX033697.D

Acq: 11 Jan 2023 15:59

Instrument :

MSVOA_X

ClientSampleId :

ICVVX011123





Tgt Ion: 91 Resp: 334257

Ion Ratio Lower Upper

91 100

92 55.4 27.5 82.5

134 29.2 14.3 42.9



#90

Hexachloroethane

Concen: 50.400 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX033697.D




Acq: 11 Jan 2023 15:59


Tgt Ion: 117 Resp: 66712

Ion Ratio Lower Upper

117 100

201 90.5 45.1 135.3









#91
1,2-Dichlorobenzene
Concen: 48.111 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:146 Resp: 206314
Ion Ratio Lower Upper
146 100
111 38.6 19.3 57.8
148 64.2 32.3 96.9



#92
1,2-Dibromo-3-Chloropropane
Concen: 47.147 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion: 75 Resp: 25303
Ion Ratio Lower Upper
75 100
155 106.2 51.7 155.1
157 135.7 67.5 202.3









#93
1,2,4-Trichlorobenzene
Concen: 52.482 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.006 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59




Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123


Tgt Ion:180 Resp: 151601
Ion Ratio Lower Upper
180 100
182 94.3 47.9 143.6
145 28.0 14.1 42.2



#94
Hexachlorobutadiene
Concen: 51.566 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59


Tgt Ion:225 Resp: 72965
Ion Ratio Lower Upper
225 100
223 63.6 31.3 93.9
227 63.3 32.3 96.9








#95
Naphthalene
Concen: 50.448 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Instrument :
MSVOA_X
ClientSampleId :
ICVVX011123



Tgt Ion:128 Resp: 419247

Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.3	15.5
129	10.9	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 51.014 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX033697.D
Acq: 11 Jan 2023 15:59

Tgt Ion:180 Resp: 147451

Ion	Ratio	Lower	Upper
180	100		
182	94.4	47.6	142.9
145	29.8	15.1	45.3

