

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011123\
 Data File : VX033688.D
 Acq On : 11 Jan 2023 10:38
 Operator : JC/MD
 Sample : VX0111WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111WBL01

Quant Time: Jan 12 00:30:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 11 14:48:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	168509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	230559	50.000	ug/l	-0.01
63) Chlorobenzene-d5	10.049	117	217863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	108720	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	90073	50.767	ug/l	-0.01
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.540%
35) Dibromofluoromethane	5.379	113	71315	49.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.620%
50) Toluene-d8	8.647	98	284760	54.591	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.180%
62) 4-Bromofluorobenzene	11.079	95	99691	51.475	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011123\
Data File : VX033688.D
Acq On : 11 Jan 2023 10:38
Operator : JC/MD
Sample : VX0111WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0111WBL01

Quant Time: Jan 12 00:30:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011123W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 11 14:48:42 2023
Response via : Initial Calibration

