

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011124\
 Data File : VX039712.D
 Acq On : 11 Jan 2024 09:58
 Operator : JC/MD
 Sample : VX0111WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0111WBL01

Quant Time: Jan 12 01:49:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	93135	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	183420	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	187657	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74655	47.851	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.700%
35) Dibromofluoromethane	5.385	113	59921	48.503	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.000%
50) Toluene-d8	8.647	98	230184	49.402	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.800%
62) 4-Bromofluorobenzene	11.079	95	98054	53.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011124\
Data File : VX039712.D
Acq On : 11 Jan 2024 09:58
Operator : JC/MD
Sample : VX0111WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0111WBL01

Quant Time: Jan 12 01:49:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

