

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011124\
 Data File : VX039716.D
 Acq On : 11 Jan 2024 11:44
 Operator : JC/MD
 Sample : P1098-03 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1311

Quant Time: Jan 12 01:52:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 11 03:51:26 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	86046	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	169505	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	87746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69249	48.043	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.080%		
35) Dibromofluoromethane	5.385	113	54982	48.158	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.320%		
50) Toluene-d8	8.647	98	213021	49.471	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.940%		
62) 4-Bromofluorobenzene	11.079	95	92291	54.810	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.620%		
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	2944	3.632	ug/l	90
40) Benzene	6.050	78	2002	0.398	ug/l	97
52) Toluene	8.720	92	1325	0.441	ug/l	86
68) m/p-Xylenes	10.305	106	717	0.285	ug/l	81
89) n-Butylbenzene	12.329	91	1245	0.219	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011124\
Data File : VX039716.D
Acq On : 11 Jan 2024 11:44
Operator : JC/MD
Sample : P1098-03 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1311

Quant Time: Jan 12 01:52:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011024W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 11 03:51:26 2024
Response via : Initial Calibration

