

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011221\
 Data File : VX020575.D
 Acq On : 12 Jan 2021 13:27
 Operator : JC/MD
 Sample : M1071-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-108

Quant Time: Jan 13 00:25:39 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	188233	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.830	114	297904	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	259882	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	121652	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	115371	51.03	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.06%	
35) Dibromofluoromethane	5.464	113	103533	50.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.32%	
50) Toluene-d8	8.695	98	344479	46.64	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 93.28%	
62) 4-Bromofluorobenzene	11.122	95	119207	45.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 91.40%	
Target Compounds						
						Qvalue
25) 2-Butanone	4.629	43	1792	1.47	ug/l	96
30) Chloroform	5.172	83	5077	1.36	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011221\
Data File : VX020575.D
Acq On : 12 Jan 2021 13:27
Operator : JC/MD
Sample : M1071-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-108

Quant Time: Jan 13 00:25:39 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

