

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011221\
 Data File : VX020583.D
 Acq On : 12 Jan 2021 16:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 13 00:27:59 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 11 15:05:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	195344	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.824	114	305763	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.092	117	291577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	135270	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.025	65	101553	43.28	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.56%
35) Dibromofluoromethane	5.458	113	99291	46.87	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.74%
50) Toluene-d8	8.689	98	369805	48.78	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.56%
62) 4-Bromofluorobenzene	11.122	95	134054	50.07	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.14%
Target Compounds						
					Qvalue	
3) Chloromethane	1.313	50	1041	0.58	ug/l #	88
4) Vinyl Chloride	1.392	62	1856	0.91	ug/l	98
27) cis-1,2-Dichloroethene	4.556	96	23817	10.19	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011221\
Data File : VX020583.D
Acq On : 12 Jan 2021 16:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 13 00:27:59 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X011121W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 11 15:05:37 2021
Response via : Initial Calibration

