

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011222\
 Data File : VX026440.D
 Acq On : 12 Jan 2022 17:46
 Operator : JC/MD
 Sample : N1122-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-PDB-472-493-011222DL

Quant Time: Jan 13 05:54:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150574	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103873	49.383	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.760%
35) Dibromofluoromethane	5.391	113	79971	49.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	8.647	98	285002	48.973	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.079	95	102946	49.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.600%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.325	96	6242	3.969	ug/l	90
16) Acetone	2.386	43	5210	4.857	ug/l	91
21) trans-1,2-Dichloroethene	3.099	96	523	0.307	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	82842	42.638	ug/l	91
42) 1,2-Dichloroethane	6.099	62	980	0.355	ug/l	81
44) Trichloroethene	7.129	130	32502	16.255	ug/l	98
52) Toluene	8.720	92	64392	14.901	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011222\
Data File : VX026440.D
Acq On : 12 Jan 2022 17:46
Operator : JC/MD
Sample : N1122-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-PDB-472-493-011222DL

Quant Time: Jan 13 05:54:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

