

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011222\
 Data File : VX026441.D
 Acq On : 12 Jan 2022 18:09
 Operator : JC/MD
 Sample : N1122-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-011222

Quant Time: Jan 13 05:54:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 11 14:45:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141437	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	200549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98488	49.848	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.700%
35) Dibromofluoromethane	5.391	113	75014	49.167	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.340%
50) Toluene-d8	8.653	98	269423	49.004	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.000%
62) 4-Bromofluorobenzene	11.079	95	97278	49.808	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.620%
Target Compounds						
16) Acetone	2.386	43	1189	1.180	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011222\
Data File : VX026441.D
Acq On : 12 Jan 2022 18:09
Operator : JC/MD
Sample : N1122-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-011222

Quant Time: Jan 13 05:54:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X011122W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 11 14:45:13 2022
Response via : Initial Calibration

