

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011223\
 Data File : VX033703.D
 Acq On : 12 Jan 2023 11:10
 Operator : JC/MD
 Sample : VX0112WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0112WBL02

Quant Time: Jan 12 23:59:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jan 07 06:16:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	13373	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	95681	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	100763	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	35528	32.811	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	109.367%
60) 4-Bromofluorobenzene	11.079	95	51771	28.507	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	95.033%
63) Toluene-d8	8.647	98	97769	30.318	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.067%
Target Compounds						
18) Methylene Chloride	2.788	84	4620	3.229	ug/l	Qvalue # 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX011223\
Data File : VX033703.D
Acq On : 12 Jan 2023 11:10
Operator : JC/MD
Sample : VX0112WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0112WBL02

Quant Time: Jan 12 23:59:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jan 07 06:16:21 2023
Response via : Initial Calibration

